

REF	530-1	6×3mL	EXP	2025-10-23	LOT	T28E6F2	更新日期：	2024-05
1 / 1								

批号：T28E6F2					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CK-MB mass)肌酸激酶-MB质量					
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.98	10.1%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.6	12%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Roche Cobas integra Roche \ Latex agglutination (LA)	mg/L	0.112	8.9%	<5	<100
(hs-CRP)超敏C反应蛋白					
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	1.1	9.1%	<5	<100
(HS-CTnT)超敏肌钙蛋白T					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.012	10%	<5	<100
(Myo / Mb)肌红蛋白					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	57.4	10%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	54.8	12%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	84.9	10%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	9.98	10%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	56.2	12%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.62	12%	<5	<100
(TnT)肌钙蛋白T					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.014	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。