



REF	500-3	6×3mL	EXP	2025-05-01	LOT	2FB6E4C	更新日期：	2022-10
1 / 3								

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(TT3)总三碘甲状腺原氨酸	(FT3)游离T3	(TT4)总T4
(FT4)游离T4	(TSH)促甲状腺素	(Cortisol)皮质醇
(FSH)促卵泡刺激素	(LH)促黄体生成素	(PROG)孕酮
(PRL)催/ 泌乳素	(TES)睾酮	(E2)雌二醇
(C-P)C-肽	(FA)叶酸	(INS)胰岛素
(VB12)维生素B12	(hCG)人绒毛膜促性腺激素	(25-OH-Vit D)25羟基维生素D

REF	500-3	6×3mL	EXP	2025-05-01	LOT	2FB6E4C	更新日期：	2022-10
								2 / 3

批号：2FB6E4C			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
(25-OH-Vit D)25羟基维生素D			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	34.6	26.3-42.9
(Cortisol)皮质醇			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	251	191-311
(C-P)C-肽			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.03	6.1-9.96
(E2)雌二醇			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3230	2450-4000
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1280	971-1580
(FA)叶酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.97	2.26-3.68
(FSH)促卵泡刺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	37.4	28.4-46.4
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	42	31.9-52.1
(FT3)游离T3			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	26.1	19.8-32.4
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	19.8	15.1-24.5
(FT4)游离T4			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	40.5	30.8-50.2
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	32.4	24.6-40.2
(hCG)人绒毛膜促性腺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	508	386-630
(INS)胰岛素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	99.2	75.4-123
(LH)促黄体生成素			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	48	36.5-59.5
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	18.6	14.1-23.1
(PRL)催/ 泌乳素			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	67.1	51-83.2
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	45.3	34.4-56.2
(PROG)孕酮			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.9	17.4-28.4
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.7	20.3-33.1
(TES)睾酮			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.79	5.92-9.66
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12	9.13-14.9
(TSH)促甲状腺素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	16	12.2-19.8
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	22.9	17.4-28.4
(TT3)总三碘甲状腺原氨酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.99	2.27-3.71
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.01	3.05-4.97



REF	500-3	6×3mL	EXP	2025-05-01	LOT	2FB6E4C	更新日期：	2022-10
3 / 3								

(TT4)总T4			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	135	103-167
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	nmol/L	162	123-201
(VB12)维生素B12			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	729	554-904

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。