



REF	500-3	6×3mL	EXP	2025-10-20	LOT	2FB6E52	更新日期：	2022-12
								1 / 3

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(25-OH-Vit D)25羟基维生素D	(Cortisol)皮质醇	(C-P)C-肽
(E2)雌二醇	(FA)叶酸	(FSH)促卵泡刺激素
(FT3)游离T3	(FT4)游离T4	(hCG)人绒毛膜促性腺激素
(INS)胰岛素	(LH)促黄体生成素	(PRL)催/ 泌乳素
(PROG)孕酮	(TES)睾酮	(TSH)促甲状腺素
(TT3)总三碘甲状原氨酸	(TT4)总T4	(VB12)维生素B12

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF	500-3	6×3mL	EXP	2025-10-20	LOT	2FB6E52	更新日期：	2022-12
								2 / 3

批号：2FB6E52			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
(25-OH-Vit D)25羟基维生素D			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35	26.6-43.4
(Cortisol)皮质醇			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	254	193-315
(C-P)C-肽			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.11	6.16-10.1
(E2)雌二醇			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1290	978-1600
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3260	2470-4040
(FA)叶酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3	2.28-3.72
(FSH)促卵泡刺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	37.8	28.7-46.9
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	42.4	32.2-52.6
(FT3)游离T3			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	20	15.2-24.8
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	26.4	20.1-32.7
(FT4)游离T4			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	32.8	24.9-40.7
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	40.9	31.1-50.7
(hCG)人绒毛膜促性腺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	514	391-637
(INS)胰岛素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	100	76-124
(LH)促黄体生成素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	18.8	14.3-23.3
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	48.5	36.9-60.1
(PRL)催/ 泌乳素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	45.8	34.8-56.8
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	67.7	51.4-84
(PROG)孕酮			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.1	17.6-28.6
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.9	20.4-33.4
(TES)睾酮			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.86	5.97-9.75
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.1	9.2-15
(TSH)促甲状腺素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	16.1	12.2-20
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	23.1	17.6-28.6
(TT3)总三碘甲状腺氨酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.02	2.3-3.74

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。



REF	500-3	6×3mL	EXP	2025-10-20	LOT	2FB6E52	更新日期：	2022-12
3 / 3								

Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.06	3.09-5.03
(TT4)总T4			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	136	103-169
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	nmol/L	164	125-203
(VB12)维生素B12			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	736	559-913

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。