

REF	530-3	6×3mL	EXP	2025-05-01	LOT	129350C	更新日期：	2022-09
1 / 1								

批号：129350C			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(CK-MB mass)肌酸激酶-MB质量			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	66.3	50.1-82.5
Perkinelmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	ng/mL	77.6	54.3-101
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	40.8	31-50.6
(CK-MB)肌酸激酶-MB (活性)			
Roche cobas 4000/6000/8000 Roche \ Enzyme immunoassay (EIA)	U/L	48.5	36.6-60.4
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	111	84.2-138
(CRP)C反应蛋白			
Perkinelmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	mg/L	11.2	7.84-14.6
(D-D(DDU)) D-二聚体(DDU)			
Sysmex CS 5100 sysmex \ Nephelometry	mg/L	3.67	2.77-4.57
Roche ACCU-CHEK series Roche \ Nephelometry	mg/L	1.58	1.2-1.96
(D-D(FEU)) D-二聚体(FEU)			
Perkinelmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	mg/L	21	14.7-27.3
(FDP)纤维蛋白(原)降解产物			
Sysmex CS 5100 sysmex \ Nephelometry	mg/L	11.3	8.44-14.2
(hs-CRP)超敏C反应蛋白			
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	8.06	6.13-9.99
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	9.57	7.24-11.9
(Myo / Mb)肌红蛋白			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/L	409	311-507
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	329	251-407
Perkinelmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	μg/L	134	93.9-174
(NT-proBNP)N-末端脑钠肽前体			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	1400	1060-1730
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	165	125-205
(TnI)肌钙蛋白I			
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	67.4	51.2-83.6
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.1	8.4-13.8

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。