

REF	530-1	6×3mL	EXP	2023-05-27	LOT	T328E61	更新日期：	2021-08
								1 / 1

		批号：T328E61	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(CK-MB mass)CK-MB质量			
PerkinElmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.1	2.52-3.68
J&J Vitros ECL Johnson&Johnson \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.01	1.57-2.45
(CRP)C反应蛋白			
PerkinElmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	mg/dL	0.176	0.142-0.21
(D-D(DDU)) D-二聚体(DDU)			
Roche ACCU-CHEK series Roche \ Nephelometry	mg/L	0.123	0.099-0.147
(hs-CRP)超敏C反应蛋白			
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	1.42	1.14-1.7
(Myo / Mb)肌红蛋白			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	57.4	45.9-68.9
PerkinElmer SuperFlex PerkinElmer \ Chemiluminescence Immunoassay (CLIA)	μg/L	19.2	15.8-22.6
(NT-proBNP)N-末端脑钠肽前体			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	97	77.6-116
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	11.4	9.12-13.7
(TnT)肌钙蛋白T			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.021	0.017-0.025

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。