

REF	539-2	6×3mL	EXP	2022-08-31	LOT	336C875	更新日期：	2020-09
1 / 4								

批号：336C875			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(AFP)甲胎蛋白			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	127	103-151
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	112	88-136
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	126	102-150
(Anti-Tg)抗甲状腺球蛋白抗体			
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	62.1	50.3-73.9
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	30.2	24-36.4
(Anti-TPO)抗甲状腺过氧化物酶抗体			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	29.3	23.5-35.1
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	35.6	29-42.2
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21.7	16.9-26.5
(CA 125)糖类抗原 125			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	61.2	48.6-73.8
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	59.6	47.8-71.4
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	62.3	50.7-73.9
(CA 15-3)糖类抗原 15-3			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.3	40.7-59.9
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.6	31.8-47.4
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	45.7	36.5-54.9
(CA 19-9)糖类抗原 19-9			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	50.5	40.3-60.7
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.7	28.5-42.9
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	51.7	41.5-61.9
(CEA)癌胚抗原			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	7.9-12.7
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.5	13.7-21.3
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.92	7.96-11.9
(CK-MB)肌酸激酶-MB			
Roche cobas c 501/502/701/702 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.86	4.7-7.02
(Cortisol)皮质醇			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	246	196-296
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	247	197-297
(C-P)C-肽			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.55	2.03-3.07
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.55	2.07-3.03
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.62	2.1-3.14

REF	539-2	6×3mL	EXP	2022-08-31	LOT	336C875	更新日期：	2020-09
2 / 4								

批号：336C875			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(DHEA-S)硫酸脱氢表雄酮			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μmol/L	17.2	13.6-20.8
(E2)雌二醇			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	947	751-1140
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	989	793-1180
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1090	874-1300
(FA)叶酸			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.68	3.76-5.6
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.26	6.62-9.9
(Ferritin)铁蛋白			
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	143	115-171
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	226	174-278
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	201	159-243
(F-PSA)游离前列腺特异性抗原			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.96	3.98-5.94
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.82	3.86-5.78
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.04	4.06-6.02
(FSH)促卵泡刺激素			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	34.2	27.2-41.2
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	32.6	25.6-39.6
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	25.6	21-30.2
(FT3)游离T3			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	12.7	10.1-15.3
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	10.4	8-12.8
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.5	9.1-13.9
(FT4)游离T4			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	36.9	29.7-44.1
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	29.6	22.8-36.4
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	30.4	24.4-36.4
(HCY)同型半胱氨酸			
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μmol/L	24.2	19.2-29.2
(INS)胰岛素			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	35.8	28.6-43
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	35.2	28.4-42
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	34.5	27.3-41.7

REF	539-2	6×3mL	EXP	2022-08-31	LOT	336C875	更新日期：	2020-09
3 / 4								

批号：336C875			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(LH)促黄体生成素			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	20.4	16.6-24.2
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	19.8	15.6-24
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	20.4	16.2-24.6
(PRL)催/泌乳素			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.9	15.3-22.5
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.9	13.7-20.1
(PROG)孕酮			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.1	8.84-13.4
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	8.04-12.6
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.13	7.31-11
(PSA)前列腺特异性抗原			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.6	9.8-15.4
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.5	9.7-15.3
(PTH)甲状旁腺素			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pg/mL	82.6	66-99.2
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pg/mL	78.4	61.8-95
(TES)睾酮			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.531	0.421-0.641
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.43	0.35-0.51
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.56	0.45-0.67
(Tg)甲状腺球蛋白			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.86	7.9-11.8
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.26	5.7-8.82
(TSH)促甲状腺素			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	10.1	7.7-12.5
Abbott ARCHITECT i series Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	9.68	7.72-11.6
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.2	9-13.4
(TT3)总三碘甲状腺原氨酸			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	3.65	2.89-4.41
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	nmol/L	3.36	2.7-4.02
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	2.55	2.11-2.99
(TT4)总T4			
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	133	107-159
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	142	116-168
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	121	99-143

REF	539-2	6×3mL	EXP	2022-08-31	LOT	336C875	更新日期：	2020-09
4 / 4								

批号：336C875			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(VB12)维生素B12			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	583	471-695
(β2M)β-2-微球蛋白			
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/L	1.16	0.9-1.42
(β-hCG)人绒毛膜促性腺激素β亚基			
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	128000	104000-152000
Beckman UniCel Dxi 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	135	107-163

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。