

<b>REF</b>	600-3	30x1.7mL	<b>EXP</b>	2024-06-22	<b>LOT</b>	3940407	更新日期:	2025-02
								1 / 2

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(K)钾	(Na)钠	(Cl)氯
(Ca)钙	(pH)酸碱度	(PO2)氧分压
(PCO2)二氧化碳分压		

<b>REF</b>	600-3	30x1.7mL	<b>EXP</b>	2024-06-22	<b>LOT</b>	3940407	更新日期:	2025-02
								2 / 2

批号: 3940407			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
<b>(Ca)钙</b>			
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	0.716	0.69-0.742
Radiometer ABL Series Dedicated \ ISE (indirect)	mmol/L	0.8	0.608-0.992
Radiometer AQTSeries Dedicated \ ISE (direct)	mmol/L	0.79	0.79-0.79
<b>(Cl)氯</b>			
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	121	119-123
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	121	92-150
Radiometer AQTSeries Dedicated \ ISE (direct)	mmol/L	123	123-123
<b>(K)钾</b>			
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	5.46	5.41-5.51
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	5.6	4.26-6.94
Radiometer AQTSeries Dedicated \ ISE (direct)	mmol/L	5.5	5.5-5.5
<b>(Na)钠</b>			
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	150	148-152
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	154	117-191
Radiometer AQTSeries Dedicated \ ISE (direct)	mmol/L	152	152-152
<b>(PCO2)二氧化碳分压</b>			
Radiometer ABL Series Radiometer \ Electrode	mmHg	18.2	15.8-20.6
Abbott FreeStyle Abbott \ Electrode	mmHg	19.7	16.6-22.8
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	mmHg	17.5	15.6-19.4
Radiometer ABL Series Dedicated \ Electrode	mmHg	18.6	14.1-23.1
Radiometer AQTSeries Dedicated \ Electrode	mmHg	17.4	17.4-17.4
<b>(pH)酸碱度</b>			
Radiometer ABL Series Radiometer \ Electrode	No Units	7.62	7.59-7.65
Abbott FreeStyle Abbott \ Dipstick	Qualitative	NA	NA
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	No Units	7.65	7.63-7.67
Radiometer ABL Series Dedicated \ Electrode	No Units	7.63	5.8-9.46
Radiometer AQTSeries Dedicated \ Electrode	No Units	7.63	7.63-7.63
<b>(PO2)氧分压</b>			
Radiometer ABL Series Radiometer \ Electrode	mmHg	137	124-150
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	mmHg	137	130-144
Radiometer ABL Series Dedicated \ Electrode	mmHg	140	106-174
Abbott FreeStyle Abbott \ Electrode	mmHg	113	90-136
Radiometer AQTSeries Dedicated \ Electrode	mmHg	142	142-142

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。