

REF 670-2	6x1mL	EXP 2025-12-23	LOT 3FEADE2	更新日期: 2024-08
				1 / 1

批号: 3FEADE2					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Sekisui Coapresta 3000 SUNBIO \ Light scattering	Seconds	94	8.4%	7	168
Sysmex CA 1500/6000/7000 Siemens Dade Actin(APTT) \ Light scattering	Seconds	94	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	102	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Dade Actin FS(APTT) \ Light scattering	Seconds	74	10%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	66	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	68	10%	<5	<100
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	62	10%	<5	<100
(AT-III)抗凝血酶III					
Werfen ACL series Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry	%	41	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin (AT□) \ Nephelometry	%	44	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Berichrom Antithrombin III (AT□) \ Nephelometry	%	42	10%	<5	<100
Werfen ACL series Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry	%	42	10%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Werfen ACL series Werfen Hemosil D-Dimer HS \ Nephelometry	mg/L	1.02	10%	<5	<100
Werfen ACL series Werfen Hemosil D-Dimer \ Nephelometry	mg/L	0.997	10%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Stago STA series Stago \ Nephelometry	mg/L	197	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin (AT□) \ Nephelometry	mg/L	3.82	10%	<5	<100
(FIB)纤维蛋白原					
Sekisui Coapresta 3000 SUNBIO \ Light scattering	g/L	1.8	5.2%	7	168
Sysmex CA 1500/6000/7000 Siemens Multifibren U (FIB) \ Light scattering	g/L	2.53	10%	<5	<100
Werfen ACL series Werfen Hemosil Fib-C \ Light scattering	g/L	1.84	10%	<5	<100
Werfen ACL series Werfen Hemosil QFA Thrombin (FIB) \ Light scattering	g/L	1.95	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thrombin (FIB) \ Light scattering	g/L	1.94	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	2.14	10%	<5	<100
(PT)血浆凝血酶原时间					
Sekisui Coapresta 3000 SUNBIO \ Light scattering	Seconds	28.1	3.7%	7	168
Stago STA series Stago \ Magnetic Beads	INR	2.87	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	INR	2.67	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	Seconds	31.1	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S (PT) \ Light scattering	INR	2.51	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	INR	2.41	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	INR	2.45	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	31.8	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	Seconds	29	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	Seconds	40.4	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S (PT) \ Light scattering	Seconds	33.2	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	INR	2.39	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	34.1	10%	<5	<100

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。