

<b>REF</b>	670-2	6x1mL	<b>EXP</b>	2025-09-03	<b>LOT</b>	3FEADDF	更新日期:	2024-08
								1 / 3

		批号: 3FEADDF			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(APTT)活化部分凝血活酶时间</b>					
Dirui BCC-3000 series	Seconds	75.8	1.5%	8	243
DIRUI \ OTHER					
zonic XL Series	Seconds	57.2	9.3%	7	<100
Zonci \ Magnetic Beads					
OTHER	Seconds	66.1	2.7%	7	193
OTHER \ Magnetic Beads					
OTHER	Seconds	93.3	8.6%	7	125
Mindray \ Magnetic Beads					
Sysmex XN series	Seconds	62.8	3%	6	295
sysmex \ Light scattering					
Mindray CX Series	Seconds	102	2.7%	5	114
Mindray \ OTHER					
Sysmex CS 5100	Seconds	84.4	12.3%	5	104
OTHER \ OTHER					
Sekisui Coapresta 3000	Seconds	72.6	0.7%	5	101
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	71.7	0.7%	<5	<100
SUNBIO \ Light scattering					
Mindray EXC810	Seconds	88.3	4.2%	<5	<100
Mindray \ OTHER					
Sysmex CS 5100	Seconds	72.2	7.4%	<5	<100
Dedicated \ Light scattering					
Sysmex CS 5100	Seconds	105	4.1%	<5	<100
Siemens Dade Actin FSL(APTT) \ Light scattering					
OTHER	Seconds	68.7	35.1%	12	121
OTHER \ OTHER					
Sysmex CS 5100	Seconds	71.3	6.3%	11	251
sysmex \ OTHER					
Succeeder SF series	Seconds	68.9	5%	10	220
Succeeder \ Magnetic Beads					
Sysmex CN Series	Seconds	107	4.2%	<5	<100
sysmex \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	74	10%	<5	<100
Siemens Dade Actin FS(APTT) \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	94	10%	<5	<100
Siemens Dade Actin(APTT) \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	102	10%	<5	<100
Siemens Pathrombin SL(APTT) \ Light scattering					
<b>(AT-III)抗凝血酶III</b>					
OTHER	%	59.8	10%	7	186
OTHER \ OTHER					
Sysmex CS 5100	%	38	10.4%	<5	<100
OTHER \ OTHER					
Sysmex CS 5100	%	41.7	8.2%	<5	<100
Siemens Innovance Antithrombin (AT□) \ OTHER					
Werfen ACL series	%	41	10%	<5	<100
Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry					
Sysmex CA 1500/6000/7000	%	44	10%	<5	<100
Siemens Innovance Antithrombin (AT□) \ Nephelometry					
Sysmex CN Series	%	42.9	2.4%	<5	<100
sysmex \ Nephelometry					
Werfen ACL series	%	41	10%	<5	<100
Werfen HemosIL Liquid Antithrombin (AT□) \ Nephelometry					
Sysmex CA 1500/6000/7000	%	42	10%	<5	<100
Siemens Berichrom Antithrombin III (AT□) \ Nephelometry					
<b>(FIB)纤维蛋白原</b>					
Dirui BCC-3000 series	g/L	2.11	4.7%	8	243
DIRUI \ OTHER					
zonic XL Series	g/L	1.79	15.1%	7	<100
Zonci \ Magnetic Beads					
OTHER	g/L	1.7	5.7%	7	197
OTHER \ Magnetic Beads					
OTHER	g/L	2.05	5.5%	7	119
Mindray \ Magnetic Beads					
Sysmex XN series	g/L	2.04	3.5%	6	294
sysmex \ Light scattering					
Mindray CX Series	g/L	1.91	4.2%	5	114
Mindray \ OTHER					
Sysmex CS 5100	g/L	1.71	5.7%	5	104
OTHER \ OTHER					
Sekisui Coapresta 3000	g/L	2.37	2.5%	5	100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	g/L	2.47	2%	<5	<100
SUNBIO \ Light scattering					
Sysmex CS-1300	g/L	1.99	9.9%	<5	<100
Siemens \ OTHER					
Mindray EXC810	g/L	1.99	3.4%	<5	<100
Mindray \ OTHER					
Sysmex CS 5100	g/L	1.78	6.9%	<5	<100
Dedicated \ Light scattering					
Sysmex CS 5100	g/L	1.78	5.6%	<5	<100
Siemens Thrombin (FIB) \ Light scattering					
Sysmex CS 5100	g/L	1.8	9.6%	11	238
sysmex \ OTHER					
OTHER	g/L	1.98	24.5%	11	118
OTHER \ OTHER					
Succeeder SF series	g/L	2.18	4.6%	10	224
Succeeder \ Magnetic Beads					
Sysmex CA 1500/6000/7000	g/L	1.94	10%	<5	<100
Siemens Thrombin (FIB) \ Light scattering					

<b>REF</b> 670-2	6x1mL	<b>EXP</b> 2025-09-03	<b>LOT</b> 3FEADDF	更新日期: 2024-08
				2 / 3

		批号: 3FEADDF			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens Multifibren U (FIB) \ Light scattering	g/L	2.53	10%	<5	<100
Werfen ACL series	g/L	1.91	10%	<5	<100
<b>(PT INR)PT国际标准化比值</b>					
Dirui BCC-3000 series	INR	1.91	3.4%	8	243
Systemex CS 5100	INR	2.47	6.7%	8	197
Mindray \ OTHER	INR	2.02	3.7%	7	118
Systemex XN series	INR	2.45	2.4%	6	295
Mindray CX Series	INR	2.09	2.5%	5	115
Systemex CS 5100	INR	2.75	2.8%	<5	<100
Systemex CS 5100	INR	2.65	4.4%	<5	<100
Mindray EXC810	INR	1.89	2.8%	<5	<100
Systemex CS 5100	INR	2.4	3.6%	<5	<100
OTHER	INR	2.33	30.5%	11	122
Succeeder SF series	INR	2.74	8%	10	299
Werfen ACL series	INR	2.5	10%	<5	<100
<b>(PT)血浆凝血酶原时间</b>					
Dirui BCC-3000 series	Seconds	21.9	4.8%	8	243
Zonci \ Magnetic Beads	Seconds	30.5	3.6%	7	<100
OTHER	Seconds	34.3	9.4%	7	198
Mindray \ Magnetic Beads	Seconds	24.9	3.8%	7	113
Systemex XN series	Seconds	27.9	2.4%	6	295
Mindray CX Series	Seconds	24.4	2.3%	5	114
Systemex CS 5100	Seconds	30.7	6%	5	104
Sekisui Coapresta 3000	Seconds	30.7	0.5%	5	101
Mindray EXC810	Seconds	22.3	2.7%	<5	<100
Systemex CS 5100	Seconds	27.4	3.5%	<5	<100
Siemens Thromborel S (PT) \ Light scattering	Seconds	27.9	3.3%	<5	<100
Systemex CS 5100	Seconds	27.1	6.9%	11	249
OTHER	Seconds	27.3	30.8%	11	122
Succeeder SF series	Seconds	32.7	7%	10	268
Systemex CA 1500/6000/7000	INR	2.82	9.9%	<5	<100
Systemex CA 1500/6000/7000	INR	2.69	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering	Seconds	33.2	10%	<5	<100
Stago STA series	INR	3.23	9.9%	<5	<100
<b>(TT)凝血酶时间</b>					
Dirui BCC-3000 series	Seconds	17.4	3.7%	8	243
Mindray \ Magnetic Beads	Seconds	18.1	2.5%	7	<100
Zonci \ Magnetic Beads	Seconds	13.1	7.9%	7	<100
OTHER	Seconds	17.9	3.8%	7	195
Systemex XN series	Seconds	15.6	3.1%	6	294
Mindray CX Series	Seconds	22.3	2.6%	5	115
Systemex CS 5100	Seconds	20.6	8%	5	104
Sekisui Coapresta 3000	Seconds	18.3	3.9%	5	101
Mindray EXC810	Seconds	18.4	3.4%	<5	<100
Systemex CS 5100	Seconds	16.9	2.7%	<5	<100

<b>REF</b>	670-2	6x1mL	<b>EXP</b>	2025-09-03	<b>LOT</b>	3FEADDF	更新日期:	2024-08
								3 / 3

		批号: 3FEADDF			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CS 5100 Siemens \ Light scattering	Seconds	18.1	1.7%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	14.9	3.3%	11	248
OTHER OTHER \ OTHER	Seconds	21.1	9.9%	11	116
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	15.7	2.6%	10	243
Sysmex CN Series sysmex \ Light scattering	Seconds	34	2.5%	<5	<100
Sysmex CS-1300 Siemens \ OTHER	Seconds	19.7	0%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。