

<b>REF</b> 670-2	6x1mL	<b>EXP</b> 2024-07-28	<b>LOT</b> 3FEAD79	更新日期: 2024-08
				1 / 3

		批号: 3FEAD79			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(APTT)活化部分凝血活酶时间</b>					
Mindray CX Series	Seconds	102	4%	<5	<100
Mindray \ OTHER					
Zonic XL Series	Seconds	63.7	2.2%	<5	108
Zonic \ clauss					
Sekisui Coapresta 3000	Seconds	78.5	3.4%	<5	<100
SUNBIO \ clauss					
Succeeder SF series	Seconds	65.5	5.4%	23	543
Succeeder \ Magnetic Beads					
OTHER	Seconds	85.5	26.8%	<5	<100
OTHER \ OTHER					
Werfen ACL series	Seconds	58.9	4.5%	19	489
OTHER \ OTHER					
Sysmex CS 5100	Seconds	69.2	6%	18	352
sysmex \ OTHER					
Werfen ACL series	Seconds	47.3	4.1%	16	116
SekiSui \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	69.4	4.9%	15	254
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	80.4	6.9%	11	309
SUNBIO \ Light scattering					
Sekisui Coapresta 3000	Seconds	71	1%	<5	<100
SekiSui \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	95.3	12%	<5	<100
Siemens \ Light scattering					
Stago STA series	Seconds	80.4	12%	<5	<100
Dedicated \ Magnetic Beads					
Sysmex CS 5100	Seconds	93.4	12%	<5	<100
Siemens \ Light scattering					
Werfen ACL series	Seconds	64.3	12%	<5	<100
Werfen \ Light scattering					
Sysmex CS 2000i/2100i	Seconds	83.2	10%	<5	<100
sysmex \ Light scattering					
Sysmex CS 5100	Seconds	69.3	3.3%	<5	<100
Siemens Dade Actin FSL(APTT) \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	77.2	0%	<5	<100
Siemens Dade Actin FSL(APTT) \ Light scattering					
Mindray EXC810	Seconds	89.1	0%	<5	<100
Mindray \ OTHER					
<b>(AT-III)抗凝血酶III</b>					
Sekisui Coapresta 3000	%	37	10%	<5	<100
SekiSui \ Nephelometry					
Sysmex CA 1500/6000/7000	%	44.3	12%	<5	<100
Siemens \ Nephelometry					
Werfen ACL series	%	43.6	12%	<5	<100
Dedicated \ Nephelometry					
<b>(D-D(DDU)) D-二聚体(DDU)</b>					
Mindray CX Series	mg/L	2.7	1.5%	<5	<100
Mindray \ Immunoturbidimetry (ITA)					
Sysmex CS 5100	mg/L	2.93	0.5%	<5	<100
sysmex \ Immunoturbidimetry (ITA)					
Sysmex CS 5100	mg/L	2.1	9.6%	12	264
sysmex \ OTHER					
<b>(FIB)纤维蛋白原</b>					
Mindray CX Series	g/L	1.95	2.7%	<5	<100
Mindray \ OTHER					
Zonic XL Series	g/L	2.04	0.9%	<5	108
Zonic \ clauss					
Succeeder SF series	g/L	2.26	7.5%	21	478
Succeeder \ Light scattering					
OTHER	g/L	1.6	11.2%	<5	<100
OTHER \ OTHER					
Sekisui Coapresta 3000	g/L	1.8	4.1%	<5	<100
SUNBIO \ clauss					
Succeeder SF series	g/L	2.45	4.7%	<5	<100
Succeeder \ Magnetic Beads					
Werfen ACL series	g/L	1.76	6.2%	19	467
SUNBIO \ OTHER					
Sysmex CS 5100	g/L	1.94	8%	18	360
sysmex \ OTHER					
Werfen ACL series	g/L	2.15	13.7%	16	126
SekiSui \ OTHER					
Sysmex CA 1500/6000/7000	g/L	2.04	5.5%	15	257
SUNBIO \ OTHER					
Sekisui Coapresta 3000	g/L	1.85	9.1%	11	314
SUNBIO \ Light scattering					
Werfen ACL series	g/L	2.24	12%	<5	<100
Werfen \ Light scattering					
Sekisui Coapresta 3000	g/L	2.24	2%	<5	<100
SekiSui \ Light scattering					
Sysmex CS 2000i/2100i	g/L	1.85	10%	<5	<100
sysmex \ Light scattering					
Sysmex CS 5100	g/L	1.86	12%	<5	<100
Siemens \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	2.15	12%	<5	<100
Siemens \ Light scattering					
Stago STA series	g/L	2.34	12%	<5	<100
Dedicated \ Magnetic Beads					
Sysmex CS 5100	g/L	1.75	15.4%	<5	<100
Siemens Thrombin ( FIB ) \ Light scattering					

<b>REF</b>	670-2	6x1mL	<b>EXP</b>	2024-07-28	<b>LOT</b>	3FEAD79	更新日期:	2024-08
								2 / 3

		批号: 3FEAD79			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
System CA 1500/6000/7000 Siemens Thrombin ( FIB ) \ Light scattering	g/L	2.29	0%	<5	<100
<b>(PT INR)PT国际标准化比值</b>					
Mindray CX Series Mindray \ Calculate	INR	2.06	5.3%	<5	<100
System CS 5100 system \ Calculate	INR	2.56	5.8%	<5	<100
Zonic XL Series Zonci \ Calculate	INR	3.05	2.4%	<5	108
Succeeder SF series Succeeder \ Calculate	INR	2.21	14.2%	23	508
OTHER OTHER \ Calculate	INR	2.48	7%	<5	<100
Werfen ACL series OTHER \ Calculate	INR	3.07	3.5%	19	480
Werfen ACL series SekiSui \ Calculate	INR	2.8	6.2%	16	129
System CA 1500/6000/7000 SUNBIO \ Calculate	INR	2.47	7%	15	247
Sekisui Coapresta 3000 SUNBIO \ Calculate	INR	2.57	7.9%	14	400
System CS 5100 system \ OTHER	INR	2.64	5.2%	14	301
Werfen ACL series Werfen \ Calculate	INR	2.79	12%	<5	<100
System CS 2000i/2100i system \ Calculate	INR	2.16	10%	<5	<100
Sekisui Coapresta 3000 SekiSui \ Calculate	INR	2.6	0.2%	<5	<100
System CA 1500/6000/7000 Siemens \ Calculate	INR	2.68	12%	<5	<100
Stago STA series Stago \ Calculate	INR	2.76	12%	<5	<100
Mindray EXC810 Mindray \ Calculate	INR	2.18	0%	<5	<100
<b>(PT)血浆凝血酶原时间</b>					
Mindray CX Series Mindray \ OTHER	Seconds	24	6.4%	<5	<100
Zonic XL Series Zonci \ clauss	Seconds	18.5	1.4%	<5	108
Sekisui Coapresta 3000 SUNBIO \ clauss	Seconds	27.5	2.7%	<5	<100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	27.5	13.4%	23	532
Werfen ACL series OTHER \ OTHER	Seconds	33.8	4%	19	491
System CS 5100 system \ OTHER	Seconds	28.3	5.3%	18	352
Werfen ACL series SekiSui \ OTHER	Seconds	32.8	6%	16	128
System CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	31.1	6.3%	15	250
Sekisui Coapresta 3000 SUNBIO \ Light scattering	Seconds	29.2	7.9%	11	312
OTHER OTHER \ OTHER	Seconds	29.3	7.5%	<5	<100
System CS 5100 Siemens \ Light scattering	Seconds	33.5	12%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	29.5	12%	<5	<100
System CA 1500/6000/7000 Siemens \ Light scattering	Seconds	31.3	12%	<5	<100
System CS 2000i/2100i system \ Light scattering	Seconds	24.3	10%	<5	<100
Sekisui Coapresta 3000 SekiSui \ Light scattering	Seconds	28.8	0.2%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	32.4	12%	<5	<100
System CS 5100 Siemens Thromborel S ( PT ) \ Light scattering	Seconds	27	2.1%	<5	<100
System CA 1500/6000/7000 Siemens Thromborel S ( PT ) \ Light scattering	Seconds	26.9	0%	<5	<100
Mindray EXC810 Mindray \ OTHER	Seconds	26.7	0%	<5	<100
<b>(TT)凝血酶时间</b>					
Mindray CX Series Mindray \ OTHER	Seconds	21	3.9%	<5	<100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	16.2	5.4%	23	503
Sekisui Coapresta 3000 SUNBIO \ clauss	Seconds	20.1	3.4%	<5	<100
Werfen ACL series OTHER \ OTHER	Seconds	18.6	3.8%	19	494
System CS 5100 system \ OTHER	Seconds	17.1	6.1%	18	351
Werfen ACL series SekiSui \ OTHER	Seconds	18.9	2.5%	16	128
System CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	21.3	9.3%	14	233
Sekisui Coapresta 3000 SUNBIO \ Light scattering	Seconds	19.7	5%	11	313

<b>REF</b>	670-2      6x1mL	<b>EXP</b>	2024-07-28	<b>LOT</b>	3FEAD79	更新日期： 2024-08 3 / 3
------------	------------------	------------	------------	------------	---------	------------------------

		批号： 3FEAD79			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Werfen ACL series SUNBIO \ Light scattering	Seconds	17.4	12%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	17.3	12%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	16.3	12%	<5	<100
Sekisui Coapresta 3000 SekiSui \ Light scattering	Seconds	15.7	10%	<5	<100
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	17.6	10%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	19	3.3%	<5	<100
OTHER OTHER \ OTHER	Seconds	22.8	5.9%	<5	<100
Mindray EXC810 Mindray \ OTHER	Seconds	19.2	0%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	19	0%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。