报告类别： **汇总实验室数据报告**

质控品名： **凝血五项非定值质控品**

REF	670-2	6×1mL	EXP	2024-12-20	LOT	3FEAD7E	更新日期：	2024-05

		批号：3FEAD7E			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Stago STA series DENUO \ Magnetic Beads	Seconds	77.9	3.5%	8	220
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	71.2	4.9%	8	173
Werfen ACL series Werfen \ OTHER	Seconds	56.9	3.3%	8	106
Sysmex CA 1500/6000/7000 sysmex \ Light scattering	Seconds	71.1	5.4%	<5	104
Maccura R01 MAKER \ Magnetic Beads	Seconds	75.5	3.5%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	79.6	12%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	63.7	12%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	92.5	12%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	70	1.9%	<5	<100
Sysmex CS-1300 Siemens \ Light scattering	Seconds	113	10%	<5	<100
Dirui BCA Series DIRUI \ clauss	Seconds	81.6	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	94.3	12%	<5	<100
zonic XL Series Zonci \ Magnetic Beads	Seconds	59.3	10%	<5	<100
(AT-III)抗凝血酶III					
Stago STA series DENUO \ Magnetic Beads	%	41.3	11.1%	8	207
Werfen ACL series Werfen \ OTHER	%	49.6	5.5%	5	<100
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	%	43.9	12%	<5	<100
Werfen ACL series Dedicated \ Nephelometry	%	43.2	12%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
zonic XL Series Zonci \ Immunoturbidimetry (ITA)	mg/L	1.12	10%	<5	<100
Dirui BCA Series DIRUI \ Nephelometry	mg/L	1.29	10%	<5	<100
(FIB)纤维蛋白原					
Stago STA series DENUO \ Magnetic Beads	g/L	1.93	10.2%	8	228
Succeeder SF series Succeeder \ Magnetic Beads	g/L	2.2	6.6%	8	186
Werfen ACL series Werfen \ OTHER	g/L	1.83	10.1%	5	<100
Sysmex CA 1500/6000/7000 sysmex \ Light scattering	g/L	1.92	6.1%	<5	105
Werfen ACL series ABX \ Light scattering	g/L	1.84	6.3%	<5	<100
Maccura R01 MAKER \ Magnetic Beads	g/L	1.94	8.4%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	g/L	2.07	1.6%	<5	<100
Werfen ACL series Werfen \ Light scattering	g/L	2.22	12%	<5	<100
zonic XL Series Zonci \ Magnetic Beads	g/L	2.14	10%	<5	<100
Sysmex CS-1300 Siemens \ Light scattering	g/L	1.69	10%	<5	<100
Dirui BCA Series DIRUI \ clauss	g/L	2.2	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	2.13	12%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	g/L	1.88	12%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	g/L	2.32	12%	<5	<100
(PT INR)PT国际标准化比值					
Werfen ACL series Werfen \ Calculate	INR	2.74	7.2%	9	127
Stago STA series DENUO \ Calculate	INR	2.74	5.5%	8	212
Succeeder SF series Succeeder \ Calculate	INR	2.54	5.6%	8	177
Sysmex CA 1500/6000/7000 sysmex \ Calculate	INR	2.53	5.2%	<5	105
Werfen ACL series ABX \ Calculate	INR	2.66	9.6%	<5	<100
Maccura R01 MAKER \ Calculate	INR	2.32	3%	<5	<100
Sysmex CS-1300 Siemens \ Calculate	INR	2.77	10%	<5	<100
zonic XL Series Zonci \ Calculate	INR	2.48	10%	<5	<100
Dirui BCA Series DIRUI \ Calculate	INR	1.9	10%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **凝血五项非定值质控品**

REF	670-2	6×1mL	EXP	2024-12-20	LOT	3FEAD7E	更新日期：	2024-05

		批号：3FEAD7E			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Stago STA series Stago \ Calculate	INR	2.73	12%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Calculate	INR	2.65	12%	<5	<100
(PT)血浆凝血酶原时间					
Stago STA series DENUO \ Magnetic Beads	Seconds	29.2	4.5%	8	240
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	30.9	5.4%	8	170
Werfen ACL series Werfen \ OTHER	Seconds	30.1	4.1%	8	100
Sysmex CA 1500/6000/7000 sysmex \ Light scattering	Seconds	27.6	4.9%	<5	105
Maccura R01 MAKER \ Magnetic Beads	Seconds	25.9	3.9%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	31	12%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	32.1	12%	<5	<100
Sysmex CS-1300 Siemens \ Light scattering	Seconds	30	10%	<5	<100
zonic XL Series Zonci \ Magnetic Beads	Seconds	28.5	10%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	29.2	12%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	33.2	12%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	26.5	2.3%	<5	<100
Dirui BCA Series DIRUI \ clauss	Seconds	21.7	10%	<5	<100
(TT)凝血酶时间					
Stago STA series DENUO \ Magnetic Beads	Seconds	16.2	2.7%	8	232
Werfen ACL series Werfen \ OTHER	Seconds	19.1	6.8%	8	111
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	15.9	2.8%	6	146
Sysmex CA 1500/6000/7000 sysmex \ Light scattering	Seconds	18.5	7%	<5	103
Werfen ACL series ABX \ Light scattering	Seconds	17.2	2.9%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	18.9	8%	<5	<100
Werfen ACL series SUNBIO \ Light scattering	Seconds	17.2	12%	<5	<100
Dirui BCA Series DIRUI \ clauss	Seconds	17.9	10%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	16.1	12%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	17.1	12%	<5	<100
Sysmex CS-1300 Siemens \ Light scattering	Seconds	17.2	10%	<5	<100
zonic XL Series Zonci \ Magnetic Beads	Seconds	12.3	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。