

| | | | | | | | | |
|------------|-------|-------|------------|------------|------------|---------|-------|---------|
| REF | 670-2 | 6x1mL | EXP | 2024-12-20 | LOT | 3FEAD7E | 更新日期: | 2024-08 |
| | | | | | | | | 1 / 2 |

| | | 批号: 3FEAD7E | | | |
|----------------------------------|---------|-------------|-------|----------|------|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| (APTT)活化部分凝血活酶时间 | | | | | |
| Stago STA series | Seconds | 77.9 | 3.5% | 8 | 220 |
| DENUO \ Magnetic Beads | Seconds | 71.2 | 4.9% | 8 | 173 |
| Succeeder SF series | Seconds | 70.8 | 4.8% | 7 | 176 |
| Succeeder \ Magnetic Beads | Seconds | 72.3 | 5.7% | <5 | <100 |
| Sysmex CA 1500/6000/7000 | Seconds | 75.5 | 3.5% | <5 | <100 |
| sysmex \ Light scattering | Seconds | 57.4 | 3.4% | 11 | 145 |
| Sysmex CS-2400 | Seconds | 79.6 | 12% | <5 | <100 |
| Siemens \ Light scattering | Seconds | 63.7 | 12% | <5 | <100 |
| Maccura R01 | Seconds | 92.5 | 12% | <5 | <100 |
| MAKER \ Magnetic Beads | Seconds | 70 | 1.9% | <5 | <100 |
| Werfen ACL series | Seconds | 113 | 10% | <5 | <100 |
| Werfen \ OTHER | Seconds | 81.6 | 10% | <5 | <100 |
| Stago STA series | Seconds | 94.3 | 12% | <5 | <100 |
| Dedicated \ Magnetic Beads | Seconds | 59.3 | 10% | <5 | <100 |
| Werfen ACL series | Seconds | | | | |
| Werfen \ OTHER | Seconds | | | | |
| Stago STA series | Seconds | | | | |
| DENUO \ Magnetic Beads | Seconds | | | | |
| Sysmex CA 1500/6000/7000 | Seconds | | | | |
| Siemens \ Light scattering | Seconds | | | | |
| zonic XL Series | Seconds | | | | |
| Zonci \ Magnetic Beads | Seconds | | | | |
| (AT-III)抗凝血酶III | | | | | |
| Werfen ACL series | % | 49.3 | 4.7% | 8 | <100 |
| Werfen \ OTHER | % | 41.3 | 11.1% | 8 | 207 |
| Stago STA series | % | 43.9 | 12% | <5 | <100 |
| DENUO \ Magnetic Beads | % | 43.2 | 12% | <5 | <100 |
| Sysmex CA 1500/6000/7000 | % | | | | |
| Siemens \ Nephelometry | % | | | | |
| Werfen ACL series | % | | | | |
| Dedicated \ Nephelometry | % | | | | |
| (D-D(DDU)) D-二聚体(DDU) | | | | | |
| zonic XL Series | mg/L | 1.12 | 10% | <5 | <100 |
| Zonci \ Immunoturbidimetry (ITA) | mg/L | 1.29 | 10% | <5 | <100 |
| Dirui BCA Series | mg/L | | | | |
| DIRUI \ Nephelometry | mg/L | | | | |
| (FIB)纤维蛋白原 | | | | | |
| Werfen ACL series | g/L | 1.89 | 9.8% | 8 | <100 |
| Werfen \ OTHER | g/L | 1.93 | 10.2% | 8 | 228 |
| Stago STA series | g/L | 2.2 | 6.6% | 8 | 186 |
| DENUO \ Magnetic Beads | g/L | 1.92 | 5.9% | 7 | 174 |
| Succeeder SF series | g/L | 1.99 | 4.5% | <5 | <100 |
| Succeeder \ Magnetic Beads | g/L | 1.84 | 6.3% | <5 | <100 |
| Sysmex CA 1500/6000/7000 | g/L | 1.94 | 8.4% | <5 | <100 |
| sysmex \ Light scattering | g/L | 2.07 | 1.6% | <5 | <100 |
| Sysmex CS-2400 | g/L | 2.22 | 12% | <5 | <100 |
| Siemens \ Light scattering | g/L | 2.14 | 10% | <5 | <100 |
| Werfen ACL series | g/L | 2.13 | 12% | <5 | <100 |
| ABX \ Light scattering | g/L | 1.69 | 10% | <5 | <100 |
| Maccura R01 | g/L | 2.2 | 10% | <5 | <100 |
| MAKER \ Magnetic Beads | g/L | 2.2 | 10% | <5 | <100 |
| Sysmex CS 5100 | g/L | 2.13 | 12% | <5 | <100 |
| sysmex \ Light scattering | g/L | 1.88 | 12% | <5 | <100 |
| Werfen ACL series | g/L | 2.22 | 12% | <5 | <100 |
| Werfen \ Light scattering | g/L | 2.13 | 12% | <5 | <100 |
| zonic XL Series | g/L | 2.14 | 10% | <5 | <100 |
| Zonci \ Magnetic Beads | g/L | 1.69 | 10% | <5 | <100 |
| Sysmex CS-1300 | g/L | 1.69 | 10% | <5 | <100 |
| Siemens \ Light scattering | g/L | 2.2 | 10% | <5 | <100 |
| Dirui BCA Series | g/L | 2.2 | 10% | <5 | <100 |
| DIRUI \ clauss | g/L | 2.2 | 10% | <5 | <100 |
| Sysmex CA 1500/6000/7000 | g/L | 2.13 | 12% | <5 | <100 |
| Siemens \ Light scattering | g/L | 2.13 | 12% | <5 | <100 |
| Sysmex CS 5100 | g/L | 1.88 | 12% | <5 | <100 |
| Siemens \ Light scattering | g/L | 2.32 | 12% | <5 | <100 |
| Stago STA series | g/L | 2.32 | 12% | <5 | <100 |
| Dedicated \ Magnetic Beads | g/L | | | | |
| (PT INR)PT国际标准化比值 | | | | | |
| Stago STA series | INR | 2.74 | 5.5% | 8 | 212 |
| DENUO \ Calculate | INR | 2.54 | 5.6% | 8 | 177 |
| Succeeder SF series | INR | 2.49 | 6% | 7 | 176 |
| Succeeder \ Calculate | INR | 2.51 | 6.4% | <5 | <100 |
| Sysmex CA 1500/6000/7000 | INR | 2.51 | 6.4% | <5 | <100 |
| sysmex \ Calculate | INR | 2.66 | 9.6% | <5 | <100 |
| Sysmex CS-2400 | INR | 2.66 | 9.6% | <5 | <100 |
| Siemens \ Calculate | INR | 2.32 | 3% | <5 | <100 |
| Werfen ACL series | INR | 2.72 | 6.8% | 12 | 165 |
| ABX \ Calculate | INR | | | | |
| Maccura R01 | INR | | | | |
| MAKER \ Calculate | INR | | | | |
| Werfen ACL series | INR | | | | |
| Werfen \ Calculate | INR | | | | |

| | | | | | | | | |
|------------|-------|-------|------------|------------|------------|---------|---------------|-------|
| REF | 670-2 | 6x1mL | EXP | 2024-12-20 | LOT | 3FEAD7E | 更新日期： 2024-08 | 2 / 2 |
|------------|-------|-------|------------|------------|------------|---------|---------------|-------|

| | | 批号：3FEAD7E | | | |
|---|---------|------------|------|----------|------|
| 项目\仪器\试剂方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Symex CS-1300 Siemens \ Calculate | INR | 2.77 | 10% | <5 | <100 |
| zonic XL Series Zonci \ Calculate | INR | 2.48 | 10% | <5 | <100 |
| Dirui BCA Series DIRUI \ Calculate | INR | 1.9 | 10% | <5 | <100 |
| Stago STA series Stago \ Calculate | INR | 2.73 | 12% | <5 | <100 |
| Symex CA 1500/6000/7000 Siemens \ Calculate | INR | 2.65 | 12% | <5 | <100 |
| (PT)血浆凝血酶原时间 | | | | | |
| Stago STA series DENUO \ Magnetic Beads | Seconds | 29.2 | 4.5% | 8 | 240 |
| Succeeder SF series Succeeder \ Magnetic Beads | Seconds | 30.9 | 5.4% | 8 | 170 |
| Symex CA 1500/6000/7000 symxex \ Light scattering | Seconds | 27.1 | 5.6% | 7 | 177 |
| Symex CS-2400 Siemens \ Light scattering | Seconds | 28.6 | 6.2% | <5 | <100 |
| Maccura R01 MAKER \ Magnetic Beads | Seconds | 25.9 | 3.9% | <5 | <100 |
| Werfen ACL series Werfen \ OTHER | Seconds | 29.9 | 4% | 11 | 133 |
| Symex CA 1500/6000/7000 Siemens \ Light scattering | Seconds | 31 | 12% | <5 | <100 |
| Werfen ACL series Werfen \ Light scattering | Seconds | 32.1 | 12% | <5 | <100 |
| Symex CS-1300 Siemens \ Light scattering | Seconds | 30 | 10% | <5 | <100 |
| zonic XL Series Zonci \ Magnetic Beads | Seconds | 28.5 | 10% | <5 | <100 |
| Stago STA series Dedicated \ Magnetic Beads | Seconds | 29.2 | 12% | <5 | <100 |
| Symex CS 5100 Siemens \ Light scattering | Seconds | 33.2 | 12% | <5 | <100 |
| Symex CS 5100 symxex \ Light scattering | Seconds | 26.5 | 2.3% | <5 | <100 |
| Dirui BCA Series DIRUI \ clauss | Seconds | 21.7 | 10% | <5 | <100 |
| (TT)凝血酶时间 | | | | | |
| Stago STA series DENUO \ Magnetic Beads | Seconds | 16.2 | 2.7% | 8 | 232 |
| Symex CA 1500/6000/7000 symxex \ Light scattering | Seconds | 19 | 7.9% | 7 | 175 |
| Succeeder SF series Succeeder \ Magnetic Beads | Seconds | 15.9 | 2.8% | 6 | 146 |
| Symex CS-2400 Siemens \ Light scattering | Seconds | 18 | 4.5% | <5 | <100 |
| Werfen ACL series ABX \ Light scattering | Seconds | 17.2 | 2.9% | <5 | <100 |
| Werfen ACL series Werfen \ OTHER | Seconds | 19.1 | 6.2% | 11 | 145 |
| Symex CS 5100 symxex \ Light scattering | Seconds | 18.9 | 8% | <5 | <100 |
| Werfen ACL series SUNBIO \ Light scattering | Seconds | 17.2 | 12% | <5 | <100 |
| Dirui BCA Series DIRUI \ clauss | Seconds | 17.9 | 10% | <5 | <100 |
| Stago STA series Dedicated \ Magnetic Beads | Seconds | 16.1 | 12% | <5 | <100 |
| Symex CA 1500/6000/7000 Dedicated \ Light scattering | Seconds | 17.1 | 12% | <5 | <100 |
| Symex CS-1300 Siemens \ Light scattering | Seconds | 17.2 | 10% | <5 | <100 |
| zonic XL Series Zonci \ Magnetic Beads | Seconds | 12.3 | 10% | <5 | <100 |

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。