

REF	670-2	6x1mL	EXP	2026-04-06	LOT	3FEADE6	更新日期:	2024-08
								1 / 2

		批号: 3FEADE6				
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数		
(APTT)活化部分凝血活酶时间						
Mindray CX Series Mindray \ Light scattering	Seconds	95.4	4.8%	6	181	
OTHER Mindray \ Magnetic Beads	Seconds	101	4.2%	<5	<100	
Mindray CX Series Mindray \ OTHER	Seconds	97.7	4.3%	<5	106	
Stago STA series Stago \ Magnetic Beads	Seconds	93.2	21.4%	<5	<100	
Sysmex CS 5100 sysmex \ Light scattering	Seconds	74.2	13.4%	<5	<100	
Sysmex CS-2500 Siemens Pathrombin SL(APTT) \ OTHER	Seconds	60.8	19%	<5	<100	
Sysmex CS 5100 sysmex \ OTHER	Seconds	75.2	7.6%	<5	<100	
Sysmex XN series sysmex \ Light scattering	Seconds	63.7	3.1%	12	605	
Sysmex CS 5100 SUNBIO \ OTHER	Seconds	83.1	3.4%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Dade Actin FS(APTT) \ Light scattering	Seconds	76.2	10%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	105	10%	<5	<100	
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	63.9	10%	<5	<100	
Mindray EXC810 Mindray \ OTHER	Seconds	96	4.7%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Dade Actin(APTT) \ Light scattering	Seconds	96.8	10%	<5	<100	
Sysmex CS-2500 sysmex \ OTHER	Seconds	64	5.9%	<5	<100	
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	68	10%	<5	<100	
Werfen ACL series OTHER \ OTHER	Seconds	60.5	5.5%	<5	<100	
Sekisui Coapresta 3000 SUNBIO \ OTHER	Seconds	68.5	2%	<5	<100	
Dymind DH Series DYMIND \ Light scattering	Seconds	66.5	2.2%	<5	<100	
(AT-III)抗凝血酶III						
Werfen ACL series Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry	%	42	10%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Berichrom Antithrombin III (AT□) \ Nephelometry	%	43	10%	<5	<100	
Stago STA series Stago \ Nephelometry	%	43	10%	<5	<100	
Sysmex CS 5100 sysmex \ Nephelometry	%	36.8	2.6%	<5	<100	
Werfen ACL series Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry	%	43	10%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin (AT□) \ Nephelometry	%	45	10%	<5	<100	
(FIB)纤维蛋白原						
Mindray CX Series Mindray \ claus	g/L	1.97	4.1%	6	364	
Mindray CX Series Mindray \ OTHER	g/L	1.89	6.3%	<5	212	
OTHER Mindray \ Magnetic Beads	g/L	2.06	2.8%	<5	154	
Sysmex CS-2500 Siemens Thrombin (FIB) \ OTHER	g/L	1.27	26.6%	<5	<100	
Stago STA series Stago \ Magnetic Beads	g/L	1.83	22.7%	<5	174	
Sysmex CS 5100 sysmex \ Light scattering	g/L	0.715	141.5%	<5	142	
Sysmex XN series sysmex \ Light scattering	g/L	2.06	4.3%	12	1212	
Werfen ACL series Werfen Hemosil Fib-C \ Light scattering	g/L	1.9	9.9%	<5	<100	
Werfen ACL series Werfen Hemosil QFA Thrombin (FIB) \ Light scattering	g/L	2.01	9.9%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Multifibren U (FIB) \ Light scattering	g/L	2.61	9.9%	<5	<100	
Mindray EXC810 Mindray \ OTHER	g/L	1.95	1.1%	<5	<100	
Sysmex CA 1500/6000/7000 Siemens Thrombin (FIB) \ Light scattering	g/L	2	9.9%	<5	<100	
Werfen ACL series SUNBIO \ OTHER	g/L	1.7	6.8%	<5	<100	
Sekisui Coapresta 3000 SUNBIO \ OTHER	g/L	2.06	1.6%	<5	<100	
Dymind DH Series DYMIND \ Magnetic Beads	g/L	1.95	3.6%	<5	<100	
Sysmex CS-2500 sysmex \ OTHER	g/L	1.48	0%	<5	<100	
Succeeder SA series Succeeder \ Magnetic Beads	g/L	2.29	1.3%	<5	<100	
Sysmex CS 5100 OTHER \ OTHER	g/L	1.81	4.6%	<5	<100	
Sysmex CS 5100 sysmex \ OTHER	g/L	1.76	0%	<5	<100	

REF 670-2	6x1mL	EXP 2026-04-06	LOT 3FEADE6	更新日期: 2024-08
				2 / 2

		批号: 3FEADE6			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(PT)PT国际标准化比值					
OTHER					
Mindray \ OTHER	INR	1.97	2.4%	<5	<100
Sysmex XN series	INR	2.35	3.4%	12	603
sysmex \ Calculate					
Mindray CX Series	INR	2.01	3.3%	10	287
Mindray \ Calculate					
Sysmex CS 5100	INR	2.4	1.9%	<5	<100
SUNBIO \ Calculate					
Mindray EXC810	INR	1.85	2.6%	<5	<100
Mindray \ Calculate					
Stago STA series	INR	3.67	10%	<5	<100
Stago \ Calculate					
Werfen ACL series	INR	2.98	4.9%	<5	<100
OTHER \ Calculate					
Sysmex CS 5100	INR	2.83	5.1%	<5	<100
OTHER \ Calculate					
(PT)血浆凝血酶原时间					
OTHER					
Mindray \ Magnetic Beads	Seconds	23.1	2.3%	<5	<100
Stago STA series	Seconds	35.4	9.9%	<5	<100
Stago \ Magnetic Beads					
Sysmex CS 5100	Seconds	26.2	16.4%	<5	<100
sysmex \ Light scattering					
Sysmex CS-2500	Seconds	29.9	12%	<5	<100
Siemens Thromborel S (PT) \ OTHER					
Sysmex XN series	Seconds	26.8	3.3%	12	604
sysmex \ Light scattering					
Mindray CX Series	Seconds	22.7	3.4%	10	287
Mindray \ OTHER					
Sysmex CS 5100	Seconds	28.8	1.9%	<5	<100
SUNBIO \ OTHER					
Werfen ACL series	Seconds	32	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Werfen ACL series	Seconds	32.8	10%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Werfen ACL series	INR	2.48	10%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Sysmex CA 1500/6000/7000	INR	2.59	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Mindray EXC810	Seconds	21.8	2.5%	<5	<100
Mindray \ OTHER					
Stago STA series	INR	2.96	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	INR	2.52	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	34.2	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Werfen ACL series	Seconds	41.6	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Werfen ACL series	INR	2.75	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Sysmex CS-2500	Seconds	27.4	0.4%	<5	<100
sysmex \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	29.9	10%	<5	<100
Siemens PT Innovin \ Light scattering					
(TT)凝血酶时间					
OTHER					
Mindray \ Magnetic Beads	Seconds	16.5	1.6%	<5	152
Sysmex CS-2500	Seconds	20.5	7%	<5	<100
Siemens \ OTHER					
Sysmex CS 5100	Seconds	18.1	6.3%	<5	140
sysmex \ Light scattering					
Sysmex XN series	Seconds	16.6	2.9%	12	1216
sysmex \ Light scattering					
Mindray CX Series	Seconds	20.2	6.7%	10	556
Mindray \ OTHER					
Mindray EXC810	Seconds	17.1	1.5%	<5	<100
Mindray \ OTHER					
Sysmex CS-2500	Seconds	19.2	9.7%	<5	<100
sysmex \ OTHER					
Werfen ACL series	Seconds	19.2	2.7%	<5	<100
OTHER \ OTHER					
Sekisui Coapresta 3000	Seconds	20.1	0.6%	<5	<100
SUNBIO \ OTHER					
Dymind DH Series	Seconds	16.1	1.1%	<5	<100
DYMIND \ Magnetic Beads					
Succeeder SA series	Seconds	14.9	4.1%	<5	<100
Succeeder \ Magnetic Beads					
Sysmex CS 5100	Seconds	18.2	3.8%	<5	<100
OTHER \ OTHER					
Sysmex CS 5100	Seconds	16.7	0%	<5	<100
sysmex \ OTHER					
Sysmex CS 5100	Seconds	19.3	1.7%	<5	<100
SUNBIO \ OTHER					

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。