报告类别： **汇总实验室数据报告**

质控品名： **凝血五项非定值质控品**

REF	670-2	6×1mL	EXP	2023-11-20	LOT	3FEAD19	更新日期：	2025-06

本报告含以下项目的示范实验室数据

参数（Parameters）	参数（Parameters）	参数（Parameters）
(PT)血浆凝血酶原时间	(APTT)活化部分凝血活酶时间	(FIB)纤维蛋白原
(TT)凝血酶时间	(AT-III)抗凝血酶III	

		批号：3FEAD19	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	58	54.4-61.6
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	84.4	70.1-98.7
Sysmex CS-2400 Siemens \ Light scattering	Seconds	74.4	58.3-90.5
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	69.5	61.7-77.3
Sysmex CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	70.2	66.7-73.7
Werfen ACL series OTHER \ OTHER	Seconds	54.5	48.9-60.1
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	88.7	71-106.4
Werfen ACL series Werfen \ Light scattering	Seconds	69.3	55.7-82.9
Sysmex CS-1300 SUNBIO \ OTHER	Seconds	73.6	70.8-76.4
(AT-III)抗凝血酶III			
Sysmex CS-1300 SUNBIO \ OTHER	%	44.5	40-49
Sysmex CA 1500/6000/7000 SUNBIO \ OTHER	%	45.9	39.3-52.5
Werfen ACL series Werfen \ Nephelometry	%	43.8	35.4-52.2
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	%	46.7	37.5-55.9
Stago STA series Dedicated \ Nephelometry	%	48.2	39-57.4
(FIB)纤维蛋白原			
Werfen ACL series Werfen HemosIL Fib-C \ Light scattering	g/L	2.23	1.88-2.58
Succeeder SF series Succeeder \ Light scattering	g/L	2.03	1.83-2.23
Sysmex CS-2400 Siemens \ Light scattering	g/L	2.01	1.68-2.34
Sysmex CS 2000i/2100i sysmex \ Light scattering	g/L	1.59	1.25-1.93
Sysmex CS-1300 SUNBIO \ OTHER	g/L	1.88	1.78-1.98
Sysmex CA 1500/6000/7000 SUNBIO \ OTHER	g/L	2.01	1.95-2.07
Succeeder SF series Succeeder \ Magnetic Beads	g/L	2.22	1.84-2.6
Werfen ACL series SUNBIO \ OTHER	g/L	1.86	1.7-2.02
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	1.83	1.47-2.19
Werfen ACL series Werfen \ Light scattering	g/L	1.72	1.4-2.04
(PT)血浆凝血酶原时间			
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	29	26.2-31.8
Succeeder SF series Succeeder \ Light scattering	Seconds	30.3	27.1-33.5
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	27.6	23.4-31.8
Sysmex CS-2400 Siemens \ Light scattering	Seconds	27.4	24.1-30.7
Sysmex CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	29.6	28.5-30.7
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	32.5	24.2-40.8
Werfen ACL series OTHER \ OTHER	Seconds	31.7	28.8-34.6
Stago STA series Dedicated \ Magnetic Beads	Seconds	24.3	19.5-29.1
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	25.6	20.8-30.4
Werfen ACL series Werfen \ Light scattering	Seconds	29.6	24.2-35
Sysmex CS-1300 SUNBIO \ OTHER	Seconds	29.7	28.6-30.8
(TT)凝血酶时间			
Sysmex CS-2400 Siemens \ Light scattering	Seconds	18.6	16.1-21.1
Sysmex CS-1300 SUNBIO \ OTHER	Seconds	18.6	17.6-19.6
Sysmex CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	17.6	16.5-18.7
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	16	14.7-17.3
Werfen ACL series OTHER \ OTHER	Seconds	18.8	17.2-20.4
Werfen ACL series Werfen \ Light scattering	Seconds	18.8	15.7-21.9
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	18.3	15.1-21.5

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。