

REF	670-2	6x1mL	EXP	2023-11-20	LOT	3FEAD19	更新日期:	2024-08
								1 / 2

		批号: 3FEAD19			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	58	3.1%	9	201
Symx CS 2000i/2100i symx \ Light scattering	Seconds	84.4	8.5%	6	100
Symx CS-2400 Siemens \ Light scattering	Seconds	74.4	10.8%	5	<100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	69.5	5.6%	22	431
Symx CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	70.2	2.5%	<5	<100
Werfen ACL series OTHER \ OTHER	Seconds	54.5	5.1%	13	251
Symx CA 1500/6000/7000 Siemens \ Light scattering	Seconds	88.7	10%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	69.3	9.8%	<5	<100
Symx CS-1300 SUNBIO \ OTHER	Seconds	73.6	1.9%	<5	<100
(AT-III)抗凝血酶III					
Symx CS-1300 SUNBIO \ OTHER	%	44.5	5.1%	<5	<100
Symx CA 1500/6000/7000 SUNBIO \ OTHER	%	45.9	7.2%	<5	<100
Werfen ACL series Werfen \ Nephelometry	%	43.8	9.6%	<5	<100
Symx CA 1500/6000/7000 Siemens \ Nephelometry	%	46.7	9.9%	<5	<100
Stago STA series Dedicated \ Nephelometry	%	48.2	9.5%	<5	<100
(FIB)纤维蛋白原					
Werfen ACL series Werfen Hemosil Fib-C \ Light scattering	g/L	2.23	7.9%	9	171
Succeeder SF series Succeeder \ Light scattering	g/L	2.03	4.9%	8	189
Symx CS-2400 Siemens \ Light scattering	g/L	2.01	8.1%	7	<100
Symx CS 2000i/2100i symx \ Light scattering	g/L	1.59	10.8%	6	<100
Symx CS-1300 SUNBIO \ OTHER	g/L	1.88	2.7%	<5	<100
Symx CA 1500/6000/7000 SUNBIO \ OTHER	g/L	2.01	1.6%	<5	<100
Succeeder SF series Succeeder \ Magnetic Beads	g/L	2.22	8.6%	14	235
Werfen ACL series SUNBIO \ OTHER	g/L	1.86	4.4%	13	256
Symx CA 1500/6000/7000 Siemens \ Light scattering	g/L	1.83	9.8%	<5	<100
Werfen ACL series Werfen \ Light scattering	g/L	1.72	9.3%	<5	<100
(PT)血浆凝血酶原时间					
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	29	4.9%	9	162
Succeeder SF series Succeeder \ Light scattering	Seconds	30.3	5.3%	8	190
Symx CS 2000i/2100i symx \ Light scattering	Seconds	27.6	7.6%	6	<100
Symx CS-2400 Siemens \ Light scattering	Seconds	27.4	6%	6	<100
Symx CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	29.6	1.8%	<5	<100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	32.5	12.8%	13	255
Werfen ACL series OTHER \ OTHER	Seconds	31.7	4.5%	13	254
Stago STA series Dedicated \ Magnetic Beads	Seconds	24.3	9.9%	<5	<100
Symx CA 1500/6000/7000 Siemens \ Light scattering	Seconds	25.6	9.4%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	29.6	9.1%	<5	<100
Symx CS-1300 SUNBIO \ OTHER	Seconds	29.7	1.9%	<5	<100
(TT)凝血酶时间					
Symx CS-2400 Siemens \ Light scattering	Seconds	18.6	6.6%	7	<100
Symx CS-1300 SUNBIO \ OTHER	Seconds	18.6	2.7%	<5	<100
Symx CA 1500/6000/7000 SUNBIO \ OTHER	Seconds	17.6	3.2%	<5	<100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	16	4.1%	14	231
Werfen ACL series OTHER \ OTHER	Seconds	18.8	4.3%	13	270
Werfen ACL series Werfen \ Light scattering	Seconds	18.8	8.3%	10	214
Symx CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	18.3	8.7%	<5	<100

报告类别： 汇总实验室数据报告



质控品名： 凝血五项非定值质控品

REF	670-2	6x1mL	EXP	2023-11-20	LOT	3FEAD19	更新日期:	2024-08
								2 / 2

		批号: 3FEAD19	
项目\仪器\试剂\方法	单位	均值 \ CV	Lab \ 点数

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。