

REF 670-2	6x1mL	EXP 2023-12-31	LOT 3FEAD1A	更新日期: 2024-08
				1 / 2

		批号: 3FEAD1A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Werfen ACL series	Seconds	55.9	5.4%	8	127
Werfen \ OTHER					
Sekisui Coapresta 3000	Seconds	71.8	5.5%	6	182
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	71.7	6.1%	6	116
Siemens \ OTHER					
Sysmex CS-1300	Seconds	73.1	2.3%	<5	<100
SUNBIO \ OTHER					
Steellex CLA1	Seconds	77	0.6%	<5	<100
Steellex \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	71.6	3%	<5	<100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	67	1.1%	21	200
SekiSui \ Light scattering					
Werfen ACL series	Seconds	65.3	10.4%	<5	<100
Werfen \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	68.7	10.6%	<5	<100
Siemens \ Light scattering					
(AT-III)抗凝血酶III					
Werfen ACL series	%	45.2	7.9%	8	127
Werfen \ OTHER					
Sysmex CA 1500/6000/7000	%	41.8	9%	7	<100
SUNBIO \ OTHER					
Sysmex CS-1300	%	42.7	5.7%	7	<100
SUNBIO \ OTHER					
Stago STA series	%	46.2	10%	<5	<100
Dedicated \ Nephelometry					
Werfen ACL series	%	43.3	9.7%	<5	<100
Werfen \ Nephelometry					
Sysmex CA 1500/6000/7000	%	44.7	10.3%	<5	<100
Siemens \ Nephelometry					
(FIB)纤维蛋白原					
Werfen ACL series	g/L	2.02	10.9%	8	133
Werfen \ OTHER					
Sekisui Coapresta 3000	g/L	1.75	2.3%	6	181
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	1.93	5.9%	6	117
Siemens \ OTHER					
Sysmex CS-1300	g/L	1.89	3.4%	<5	<100
SUNBIO \ OTHER					
Steellex CLA1	g/L	1.47	9.8%	<5	<100
Steellex \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	1.96	4.3%	<5	<100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	g/L	1.05	1%	19	183
SekiSui \ Light scattering					
Werfen ACL series	g/L	2.14	7.3%	<5	<100
MEDCONN \ Light scattering					
Werfen ACL series	g/L	1.82	9.3%	<5	<100
Werfen \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	1.88	9%	<5	<100
Siemens \ Light scattering					
(PT)血浆凝血酶原时间					
Werfen ACL series	Seconds	29.5	6%	8	127
Werfen \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	26.7	6.8%	7	140
Siemens \ OTHER					
Sekisui Coapresta 3000	Seconds	26.9	2%	6	181
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	29	2.8%	5	<100
SUNBIO \ OTHER					
Sysmex CS-1300	Seconds	29.3	2.4%	<5	<100
SUNBIO \ OTHER					
Steellex CLA1	Seconds	21	2.6%	<5	<100
Steellex \ Light scattering					
Sekisui Coapresta 3000	Seconds	38.3	0.6%	21	209
SekiSui \ Light scattering					
Werfen ACL series	Seconds	28.6	9.4%	<5	<100
Werfen \ Light scattering					
Stago STA series	Seconds	24.6	9.8%	<5	<100
Dedicated \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	25.2	9.5%	<5	<100
Siemens \ Light scattering					
(TT)凝血酶时间					
Werfen ACL series	Seconds	19.3	5.2%	8	110
Werfen \ OTHER					
Sekisui Coapresta 3000	Seconds	19.5	2.3%	7	182
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	18.1	5.1%	6	120
Siemens \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	18.1	4.9%	5	<100
SUNBIO \ OTHER					
Sysmex CS-1300	Seconds	18.8	2.9%	<5	<100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	20.2	0.9%	21	204
SekiSui \ Light scattering					
Werfen ACL series	Seconds	16.5	1.7%	<5	<100
MEDCONN \ Light scattering					

REF	670-2	6x1mL	EXP	2023-12-31	LOT	3FEAD1A	更新日期:	2024-08
								2 / 2

		批号: 3FEAD1A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	16.4	9.8%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	18.3	9.8%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。