

<b>REF</b>	510-1	6x3mL	<b>EXP</b>	2027-02-10	<b>LOT</b>	30A626C	更新日期:	2024-07
								1 / 2

		批号: 30A626C			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
<b>(AFP)甲胎蛋白</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.48	16.1%	72	1685
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.24	9.4%	12	283
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.33	17.2%	10	209
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.15	10%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.26	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.17	0.5%	<5	<100
<b>(CA 125)糖类抗原 125</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	39.2	5.8%	67	1516
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.7	2.5%	15	340
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.8	0.9%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	30.3	10%	<5	<100
<b>(CA 15-3)糖类抗原 15-3</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	14.5	7.5%	67	1450
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.3	3.5%	16	267
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15	1.5%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	14.9	10%	<5	<100
<b>(CA 19-9)糖类抗原 19-9</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	10.8	3.4%	14	283
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	10.5	2.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	10.5	10%	<5	<100
<b>(CA 50)糖类抗原 50</b>					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/mL	8.78	10%	<5	<100
<b>(CA 72-4)糖类抗原 72-4</b>					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	4.88	6.7%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.43	4.1%	12	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	2.97	12%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	3.39	10%	<5	<100
<b>(CA242)糖类抗原 242</b>					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	U/mL	6.09	10%	<5	<100
<b>(CEA)癌胚抗原</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.47	13.9%	71	1678
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.85	5.2%	15	335
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.9	9.1%	10	211
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.89	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.88	0.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.11	12%	<5	<100
<b>(Cyfra 21-1)细胞角蛋白19片段</b>					
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.851	37.9%	9	193
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.96	25%	7	163
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.44	7.4%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.35	3.2%	12	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.98	10%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.7	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.05	2.3%	<5	<100
<b>(Ferritin)铁蛋白</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	84.8	4.1%	18	399

<b>REF</b>	510-1	6x3mL	<b>EXP</b>	2027-02-10	<b>LOT</b>	30A626C	更新日期:	2024-07
								2 / 2

		批号: 30A626C			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray CL-2000i/1000i/6000i series Mindray \ Immunonephelometry (INA)	ng/mL	769	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	65.3	10%	<5	<100
<b>(F-PSA)游离前列腺特异性抗原</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.732	8.4%	54	1268
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.582	4%	15	358
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.676	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.698	1.8%	<5	<100
<b>(hCG)人绒毛膜促性腺激素</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	17	10%	<5	<100
<b>(HE4)人附睾蛋白4</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	32.5	10.8%	5	<100
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	34.6	7.2%	10	227
<b>(proGRP)胃泌素释放肽前体</b>					
Abbott ARCHITECT c series Abbott \ Electro Chemi luminescence (ECL)	ng/L	193582493	11.1%	9	160
Abbott ARCHITECT i series Abbott \ Electro chemi luminescence (ECL)	ng/L	168522641	8.9%	5	106
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/L	192000000	10%	<5	<100
<b>(PSA)前列腺特异性抗原</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.46	6.5%	68	1540
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.39	3.1%	15	290
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.84	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.85	0.5%	<5	<100
<b>(SCC)鳞状细胞癌抗原</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.18	2.9%	8	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.39	5.9%	53	1147
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.23	25.7%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.24	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.23	10%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.017	10%	<5	<100

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。