

REF	670-3	6x1mL	EXP	2025-12-23	LOT	3FED4F2	更新日期:	2024-08
								1 / 2

		批号: 3FED4F2			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Succeeder SF series Succeeder \ clauss	Seconds	39.5	4.3%	5	148
BE Thrombolyzer X/XRC/XRM Behnk Elektronik (BE) \ clauss	Seconds	27.3	1.9%	5	100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	94.5	5.2%	<5	161
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	101	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	85	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	148	10%	<5	<100
Sysmex CN Series sysmex \ Light scattering	Seconds	158	3.6%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	94	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Dade Actin(APTT) \ Light scattering	Seconds	143	10%	<5	<100
(AT-III)抗凝血酶III					
Werfen ACL series Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry	%	25	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Berichrom Antithrombin III (AT□) \ Nephelometry	%	27	10%	<5	<100
Sysmex CN Series sysmex \ Nephelometry	%	21.2	4%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Werfen ACL series Werfen Hemosil D-Dimer HS \ Nephelometry	mg/L	1.12	10%	<5	<100
Werfen ACL series Werfen Hemosil D-Dimer \ Nephelometry	mg/L	1.07	10%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin (AT□) \ Nephelometry	mg/L	3.67	10%	<5	<100
Succeeder SF series Succeeder \ Nephelometry	mg/L	1.96	0%	<5	<100
(FIB)纤维蛋白原					
Succeeder SF series Succeeder \ clauss	g/L	2.97	5.8%	5	148
BE Thrombolyzer X/XRC/XRM Behnk Elektronik (BE) \ clauss	g/L	3.23	2.3%	5	100
Succeeder SF series Succeeder \ Magnetic Beads	g/L	1.25	5.7%	<5	168
Werfen ACL series Werfen Hemosil QFA Thrombin (FIB) \ Light scattering	g/L	1.22	10%	<5	<100
Werfen ACL series Werfen Hemosil Fib-C \ Light scattering	g/L	1.11	10%	<5	<100
Sysmex CN Series sysmex \ Light scattering	g/L	1.96	2.1%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Multifibren U (FIB) \ Light scattering	g/L	1.75	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	1.36	10%	<5	<100
(PT)血浆凝血酶原时间					
Succeeder SF series Succeeder \ clauss	Seconds	12.8	4.2%	5	147
BE Thrombolyzer X/XRC/XRM Behnk Elektronik (BE) \ clauss	Seconds	11.6	1.4%	5	100
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	55.5	6%	<5	163
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	52.1	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	INR	4.43	9.9%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	INR	4.04	9.9%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	Seconds	47.9	10%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	51.9	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	Seconds	62.7	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	INR	5.1	10%	<5	<100
Sysmex CN Series sysmex \ Light scattering	Seconds	54.2	1.7%	<5	<100
Stago STA series Stago \ Magnetic Beads	INR	5.8	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S (PT) \ Light scattering	Seconds	49.5	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S (PT) \ Light scattering	INR	4.55	10.1%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	Seconds	42.8	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	INR	4.38	10%	<5	<100

报告类别： 汇总实验室数据报告



质控品名： 凝血五项非定值质控品

REF	670-3	6x1mL	EXP	2025-12-23	LOT	3FED4F2	更新日期:	2024-08
								2 / 2

		批号: 3FED4F2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TT)凝血酶时间					
Sysmex CN Series sysmex \ Light scattering	Seconds	119	5.6%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。