

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期: 2025-02
							1 / 7

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(Myo / Mb)肌红蛋白	(TnI)肌钙蛋白I	(NT-proBNP)N-末端脑钠肽前体
(TT3)总三碘甲状腺原氨酸	(FT3)游离T3	(TT4)总T4
(FT4)游离T4	(TSH)促甲状腺素	(Cortisol)皮质醇
(FSH)促卵泡刺激素	(LH)促黄体生成素	(PROG)孕酮
(PRL)催乳素	(TES)睾酮	(E2)雌二醇
(C-P)C-肽	(FA)叶酸	(INS)胰岛素
(VB12)维生素B12	(IgE)免疫球蛋白 IgE	(AFP)甲胎蛋白
(CEA)癌胚抗原	(hCG)人绒毛膜促性腺激素	(PSA)前列腺特异性抗原
(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3	(CA 19-9)糖类抗原 19-9
(β2M)β-2-微球蛋白	(Ferritin)铁蛋白	(β-hCG)人绒毛膜促性腺激素β亚基
(F-PSA)游离前列腺特异性抗原	(HCY)同型半胱氨酸	(ALD)醛固酮
(Anti-Tg)抗甲状腺球蛋白抗体	(Anti-TPO)抗甲状腺过氧化物酶抗体	(DHEA-S)硫酸脱氢表雄酮
(HGH)人生长激素	(IGF1)胰岛素生长因子1	(PTH)甲状旁腺素
(Tg)甲状腺球蛋白	(PCT)降钙素原	

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期: 2025-02	2 / 7
------------	-------	-------	------------	------------	------------	---------	---------------	-------

批号: 336A289			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(AFP)甲胎蛋白			
Beckman UniCel Dxl 600/800	ng/mL	5.92	4.31-7.53
Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.02	7.17-8.87
Autolumo A2000/A2000 PLUS	ng/mL	8.51	6.42-10.6
Autobio \ OTHER	ng/mL	6.63	6.13-7.13
Maccura i3000	ng/mL	8.37	7.11-9.63
MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.21	5.23-7.19
Mindray CL-2000i/1000i/6000i series	ng/mL	7.42	7.09-7.75
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.19	5.46-8.92
maglumi system	ng/mL	6.74	5.39-8.09
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.81	5.82-7.8
maglumi system	ng/mL	7.59	7.21-7.97
Snibe \ OTHER	ng/mL		
Roche cobas 4000/6000/8000	ng/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL		
Roche cobas E601/602	ng/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL		
Beckman UniCel Dxl 600/800	ng/mL		
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL		
Roche cobas e 801	ng/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL		
Autolumo A2000/A2000 PLUS	ng/mL		
Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL		
(ALD)醛固酮			
maglumi system	pg/mL	144	132-156
Snibe \ Chemiluminescence Immunoassay (CLIA)	pg/mL		
(Anti-Tg)抗甲状腺球蛋白抗体			
Autolumo A2000/A2000 PLUS	IU/mL	26.8	9.5-44.1
Autobio \ OTHER	IU/mL	93.1	74.5-111.7
maglumi system	IU/mL	25.2	22.9-27.5
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.4	12.3-18.5
Autolumo A2000/A2000 PLUS	IU/mL	9.8	7.84-11.76
Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.7	10.9-12.5
Dirui CM Series	IU/mL	24.9	18.9-30.9
DIRUI \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Beckman UniCel Dxl 600/800	IU/mL		
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Mindray CL-2000i/1000i/6000i series	IU/mL		
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Roche cobas E601/602	IU/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL		
(Anti-TPO)抗甲状腺过氧化物酶抗体			
Autolumo A2000/A2000 PLUS	IU/mL	39.3	34.8-43.8
Autobio \ OTHER	IU/mL	41.6	26.5-56.7
maglumi system	IU/mL	15.3	13.4-17.2
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.6	9.3-13.9
Mindray CL-2000i/1000i/6000i series	IU/mL	23.3	17.7-28.9
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.3	11.4-17.2
Beckman UniCel Dxl 600/800	IU/mL	24.7	23.3-26.1
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Roche cobas E601/602	IU/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL		
Dirui CM Series	IU/mL		
DIRUI \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Autolumo A2000/A2000 PLUS	IU/mL		
Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
(CA 125)糖类抗原 125			
Maccura i3000	IU/mL	7.29	5.67-8.91
MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.3	8.8-13.8
Autolumo A2000/A2000 PLUS	IU/mL	8.65	7.23-10.07
Autobio \ OTHER	IU/mL	7.84	3.26-12.42
Mindray CL-2000i/1000i/6000i series	IU/mL	7.42	6.19-8.65
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.94	8.4-9.48
maglumi system	IU/mL	9.69	8.2-11.18
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7	5.6-8.4
maglumi system	IU/mL	9.63	7.32-11.94
Snibe \ OTHER	IU/mL	9.38	7.64-11.12
Roche cobas 4000/6000/8000	IU/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL		
Roche cobas e 801	IU/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL		
Beckman UniCel Dxl 600/800	IU/mL		
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Roche cobas E601/602	IU/mL		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL		
Autolumo A2000/A2000 PLUS	IU/mL		
Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
(CA 15-3)糖类抗原 15-3			
maglumi system	IU/mL	11.5	9.2-13.8
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	9.94	8.27-11.61
Mindray CL-2000i/1000i/6000i series	IU/mL	16.5	7.4-25.6
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL		
Autolumo A2000/A2000 PLUS	IU/mL		
Autobio \ OTHER	IU/mL		

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期:	2025-02
								3 / 7

Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15	0.1-29.9
maglumi system Snibe \ OTHER	IU/mL	12	9.6-14.4
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.4	11.2-13.6
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	12.4	11.1-13.7
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.6	9.6-15.6
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12	11.1-12.9
Beckman UniCel Dxl 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	8.89	7.65-10.13
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.9	7.12-10.68

(CA 19-9)糖类抗原 19-9

maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5.74	4.01-7.47
Autolumo A2000/A2000 PLUS Autobio \ OTHER	IU/mL	3.19	2.6-3.78
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.57	2.78-4.36
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5	4.03-5.97
maglumi system Snibe \ OTHER	IU/mL	4.75	4.02-5.48
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.03	4.62-5.44
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.1	2.48-3.72
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	4.82	3.66-5.98
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	4.91	3.84-5.98
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.87	2.94-4.8

(CEA)癌胚抗原

Autolumo A2000/A2000 PLUS Autobio \ OTHER	ng/mL	3.53	2.92-4.14
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.02	2.2-3.84
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.71	2.46-2.96
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.22	2.61-3.83
maglumi system Snibe \ OTHER	ng/mL	3.21	2.83-3.59
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.8	2.24-3.36
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.25	2.47-4.03
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.45	2.82-4.08
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	2.6-3.36
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.27	1.27-1.27

(Cortisol)皮质醇

maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	141	126-156
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	118	103-133
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	119	90-148

(C-P)C-肽

Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.827	0.686-0.968
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.26	1.07-1.45
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.25	1.22-1.28
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.24	0.94-1.54

(DHEA-S)硫酸脱氢表雄酮

Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3780	2873-4687
---	-------	------	-----------

(E2)雌二醇

maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pmol/L	235	177-293
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	230	157-303
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	377	295-459
maglumi system Snibe \ OTHER	pmol/L	203	147-259
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	252	181-323

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期:	2025-02
								4 / 7

Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	266	202-330
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	244	229-259
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	pmol/L	338	270-406
(FA)叶酸			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.76	2.48-3.04
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.15	2.39-3.91
(Ferritin)铁蛋白			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	120	109-131
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	78.3	62.6-94
(F-PSA)游离前列腺特异性抗原			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.096	0.086-0.106
Autolumo A2000/A2000 PLUS Autobio \ OTHER	ng/mL	0.136	0-0.435
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.132	0.104-0.16
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1	0.8-1.2
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.113	0.086-0.14
(FSH)促卵泡刺激素			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.35	4.43-6.27
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	5.2	3.73-6.67
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.23	4.35-6.11
maglumi system Snibe \ OTHER	IU/L	4.94	4.29-5.59
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.88	4.4-5.36
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.24	4.27-6.21
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.07	3.85-6.29
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.57	3.66-5.48
(FT3)游离T3			
Autolumo A2000/A2000 PLUS Autobio \ OTHER	pmol/L	8.27	7.15-9.39
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	7.23	4.8-9.66
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	5.9	5.56-6.24
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	7.11	5.33-8.89
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.87	5.47-6.27
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pmol/L	4.9	4.23-5.57
maglumi system Snibe \ OTHER	pmol/L	4.99	4.61-5.37
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.1	4.54-5.66
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.11	4.89-7.33
Mindray BC-6000 series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5	4-6
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.13	5.77-6.49
(FT4)游离T4			
Autolumo A2000/A2000 PLUS Autobio \ OTHER	pmol/L	16.3	13.3-19.3
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.4	13.1-25.7
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	17.4	16.6-18.2
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.4	15.2-19.6
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	22.3	20.1-24.5
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pmol/L	16.5	13.3-19.7
maglumi system Snibe \ OTHER	pmol/L	17	14.8-19.2
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	16.2	15-17.4
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	7.96	7.69-8.23
Mindray BC-6000 series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	106	85-127
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	pmol/L	19.3	15.4-23.2

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期:	2025-02
								5 / 7

(hCG)人绒毛膜促性腺激素			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	1.91	1.48-2.34
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	6.75	6.03-7.47
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.62	2.69-4.55
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.09	4.12-6.06
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/L	7.65	6.12-9.18
(HCY)同型半胱氨酸			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	μmol/L	13.4	12.2-14.6
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	13.4	10.2-16.6
(HGH)人生长激素			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	1.82	1.38-2.26
(IgE)免疫球蛋白 IgE			
Beckman Image 800 Roche \ Immunoturbidimetry (ITA)	IU/mL	32	24.3-39.7
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.6	21.5-35.7
(IGF1)胰岛素生长因子1			
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	223	169-277
(INS)胰岛素			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	11.8	11.1-12.5
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	10.4	9.2-11.6
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.8	9-14.6
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	11	9.7-12.3
(LH)促黄体生成素			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.37	4.52-6.22
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.47	2.61-4.33
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.99	4.18-5.8
maglumi system Snibe \ OTHER	IU/L	5.12	4.25-5.99
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.03	4.94-5.12
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.91	3.73-6.09
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.81	3.85-5.77
(Myo / Mb)肌红蛋白			
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	μg/L	37	31.2-42.8
(NT-proBNP)N-末端脑钠肽前体			
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pg/mL	153	128-178
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	18	15.1-20.9
(PCT)降钙素原			
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.23	1.23-1.23
(PRL)催乳素			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.22	5-7.44
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.8	4.71-6.89
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.9	7.35-10.45
maglumi system Snibe \ OTHER	ng/mL	6.2	5.64-6.76
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.96	6.7-7.22
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.49	6.45-10.53
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.07	7.3-8.84
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.38	6.7-10.06
(PROG)孕酮			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.716	0.371-1.061
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.44	1.21-1.67

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期:	2025-02
								6 / 7

Ortho Vitros 3600/5600 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.49	2.18-2.8
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.4	0.85-1.95
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29	0.93-1.65
maglumi system Snibe \ OTHER	ng/mL	0.94	0.436-1.444
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.41	1.18-1.64
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.94	0.752-1.128
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.37	1.04-1.7
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29	1.18-1.4
(PSA)前列腺特异性抗原			
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.586	0.464-0.708
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.434	0.247-0.621
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.54	0.507-0.573
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.526	0.454-0.598
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.48	0.33-0.63
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.56	0.448-0.672
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.594	0.451-0.737
(PTH)甲状旁腺素			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.43	1.42-1.44
(TES)睾酮			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.57	4.6-6.54
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.2	4.65-7.75
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.35	7.06-9.64
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.75	6.13-7.37
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.62	5.71-7.53
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.65	5.05-8.25
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.4	5.92-8.88
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.63	6.93-8.33
(Tg)甲状腺球蛋白			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.1	8.4-13.8
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.89	5.51-8.27
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.1	3.01-3.19
(Tni)肌钙蛋白I			
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.019	0.013-0.025
(TSH)促甲状腺素			
Autolumo A2000/A2000 PLUS Autobio \ OTHER	mIU/L	2.78	2.44-3.12
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	2.71	1.58-3.84
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	2.93	2.78-3.08
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.01	2.75-3.27
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.16	2.58-3.74
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.02	2.64-3.4
maglumi system Snibe \ OTHER	mIU/L	2.96	2.78-3.14
Mindray BC-6000 series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	1.69	1.35-2.03
Dirui CM Series DIRUI \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.71	2.17-3.25
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	3.6	3.15-4.05
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	0.317	0.28-0.354
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.85	2.4-3.3

REF	539-1	6x3mL	EXP	2025-01-01	LOT	336A289	更新日期:	2025-02
								7 / 7

(TT3)总三碘甲状腺原氨酸			
Autolumo A2000/A2000 PLUS			
Autobio \ OTHER	ng/mL	1.09	0.97-1.21
Roche cobas 4000/6000/8000			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.6	1.47-1.73
Mindray CL-2000i/1000i/6000i series			
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.19	0.7-1.68
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.14	1-1.28
maglumi system			
Snibe \ OTHER	ng/mL	1.17	1-1.34
Mindray BC-6000 serices			
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.32	1.06-1.58
Dirui CM Series			
DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.06	0.85-1.27
Roche cobas E601/602			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.43	1.09-1.77
Abbott ARCHITECT i series			
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.926	0.854-0.998
Autolumo A2000/A2000 PLUS			
Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.41	1.14-1.68
(TT4)总T4			
Autolumo A2000/A2000 PLUS			
Autobio \ OTHER	ng/mL	71.6	58-85.2
Roche cobas 4000/6000/8000			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	65.3	60.3-70.3
Mindray CL-2000i/1000i/6000i series			
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.1	59.8-76.4
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	72.5	61.3-83.7
maglumi system			
Snibe \ OTHER	ng/mL	76.1	68.5-83.7
Autolumo A2000/A2000 PLUS			
Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	61.5	56.6-66.4
Roche cobas E601/602			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	62.1	47.2-77
Dirui CM Series			
DIRUI \ Chemiluminescence Immunoassay (CLIA)	ng/mL	74.7	59.8-89.6
Mindray BC-6000 serices			
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	108	86-130
Abbott ARCHITECT i series			
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	28.7	23.8-33.6
(VB12)维生素B12			
Mindray CL-2000i/1000i/6000i series			
Mindray \ Chemiluminescence Immunoassay (CLIA)	pg/mL	432	423-441
Autolumo A2000/A2000 PLUS			
Autobio \ Chemiluminescence Immunoassay (CLIA)	pg/mL	447	396-498
Roche cobas E601/602			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	436	331-541
(\beta2M)\beta-2-微球蛋白			
Roche cobas 4000/6000/8000			
Roche \ Immunoturbidimetry (ITA)	ng/mL	890	676-1104
Autolumo A2000/A2000 PLUS			
Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	720	648-792
(\beta-hCG)人绒毛膜促性腺激素\beta亚基			
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.007	0.006-0.008
Mindray CL-2000i/1000i/6000i series			
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.002	0.002-0.002
maglumi system			
Snibe \ OTHER	IU/mL	0.007	0.006-0.008
Dirui CM Series			
DIRUI \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.006	0.005-0.007
Roche cobas E601/602			
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	0.002-0.004

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。