报告类别： **汇总实验室数据报告**

质控品名： **特定蛋白液体非定值质控品**

REF	521-1	6×3mL	EXP	2024-10-17	LOT	31B2AEE	更新日期：	2024-03

		批号：31B2AEE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Alb)白蛋白					
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	20.3	12%	<5	<100
(ASO)抗链球菌溶血素 O 抗体					
Siemens BN ProSpec Dedicated \ Immunonephelometry (INA)	IU/mL	133	3.1%	16	344
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	82.5	12%	<5	<100
(AT-III)抗凝血酶III					
Stago STA series Dedicated \ Nephelometry	%	7	12%	<5	<100
(C3)补体 C3					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	73.6	8.1%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	78.8	4.9%	16	346
(C4)补体 C4					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	18.5	9.3%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	18.7	4.6%	13	273
(Cer)铜蓝蛋白					
Siemens BN ProSpec Dedicated \ Immunonephelometry (INA)	g/L	0.165	2.2%	6	130
Siemens BN II Dedicated \ Immunonephelometry (INA)	g/L	0.169	4%	10	209
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	0.132	12%	<5	<100
(CRP)C反应蛋白					
Beckman AU series DongOu \ Immunoturbidimetry (ITA)	mg/dL	1.5	2.2%	7	138
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	1.36	12%	<5	<100
(Ferritin)铁蛋白					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	66.4	12%	<5	<100
(IgA)免疫球蛋白 IgA					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	121	8.3%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	136	5.9%	16	343
(IgE)免疫球蛋白 IgE					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	28.7	12.4%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	IU/mL	31.3	4.1%	10	212
(IgG)免疫球蛋白 IgG					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	683	9%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	751	4.3%	16	344
(IgM)免疫球蛋白 IgM					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	83.3	9%	<5	<100
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	87.6	4.4%	16	342
(Kappa Light Chain)κ轻链抗体					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	5.85	12%	<5	<100
(Lambda Light Chain)λ 轻链抗体					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	2.97	12%	<5	<100
(PAB)前白蛋白					
Beckman AU series DongOu \ Immunoturbidimetry (ITA)	g/L	0.165	2.5%	27	564
Roche cobas 4000/6000/8000 DiaSys \ Immunonephelometry (INA)	g/L	0.161	12%	<5	<100
(RBP)视黄醇结合蛋白					
Beckman AU series DongOu \ Immunoturbidimetry (ITA)	mg/dL	2.78	3%	27	561
Roche cobas 4000/6000/8000 WanTai \ Immunonephelometry (INA)	mg/dL	2.76	12%	<5	<100
(RF)类风湿因子					
Siemens BN ProSpec Dedicated \ Immunonephelometry (INA)	IU/mL	38.6	5%	16	346
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	36.1	12%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **特定蛋白液体非定值质控品**

REF	521-1	6×3mL	EXP	2024-10-17	LOT	31B2AEE	更新日期：	2024-03

		批号：31B2AEE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TRF)转铁蛋白					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	mg/dL	127	12%	<5	<100
(β2M)β-2-微球蛋白					
Beckman AU series LEADMAN \ Immunoturbidimetry (ITA)	ng/mL	2625	2.8%	27	561
Roche cobas 4000/6000/8000 JS (Shanghai) \ Immunonephelometry (INA)	ng/mL	2410	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。