

<b>REF</b>	760-3 6x3mL	<b>EXP</b> 2027-02-15	<b>LOT</b> 488292F	更新日期: 2025-02 1 / 3
------------	----------------	-----------------------	--------------------	------------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(K)钾	(Na)钠	(Ca)钙
(P)磷	(Fe)铁	(Mg)镁
(Li)锂	(Cu)铜	(Zn)锌
(Al)铝	(Sb)锑	(As)砷
(Be)铍	(Bi)铋	(Cd)镉
(Cr)铬	(Co)钴	(I)碘
(Pb)铅	(Mn)锰	(Hg)汞
(Mo)钼	(Ni)镍	(Se)硒
(Ti)钛	(Sn)锡	(V)钒
(Au)金	(B)硼	(Rb)铷
(Ag)银	(Sr)锶	(Ba)钡
(Br)溴	(Cs)铯	(Si)硅
(S)硫		

<b>REF</b>	760-3	6x3mL	<b>EXP</b>	2027-02-15	<b>LOT</b>	488292F	更新日期:	2025-02
								2 / 3

批号: 488292F			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
<b>(Ag)银</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.044	0.035-0.053
<b>(Al)铝</b>			
PerkinElmer NexION Series PerkinElmer \ MS	mg/L	0.09	0.09-0.09
OTHER Dedicated \ ICP-MS	mg/L	0.11	0.088-0.132
<b>(As)砷</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.034	0.027-0.041
<b>(Au)金</b>			
OTHER Dedicated \ ICP-MS	µg/L	0.04	0.031-0.049
<b>(B)硼</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.033	0.026-0.04
<b>(Ba)钡</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.176	0.141-0.211
<b>(Be)铍</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.003	0.002-0.004
<b>(Bi)铋</b>			
OTHER Dedicated \ ICP-MS	µg/L	59.4	47.5-71.3
<b>(Br)溴</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.99	0.792-1.188
<b>(Ca)钙</b>			
OTHER Dedicated \ ICP-MS	mmol/L	2.12	1.7-2.54
<b>(Cd)镉</b>			
OTHER Dedicated \ ICP-MS	µmol/L	0.127	0.102-0.152
OTHER Dedicated \ ICP-MS	µg/L	14.3	11.4-17.2
<b>(Co)钴</b>			
OTHER Dedicated \ ICP-MS	µmol/L	0.205	0.164-0.246
OTHER Dedicated \ ICP-MS	µg/L	12.1	9.7-14.5
<b>(Cr)铬</b>			
OTHER Dedicated \ ICP-MS	µg/L	44	35.2-52.8
<b>(Cs)铯</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.003	0.002-0.004
<b>(Cu)铜</b>			
OTHER Dedicated \ ICP-MS	µmol/L	39.8	31.8-47.8
<b>(Fe)铁</b>			
OTHER Dedicated \ ICP-MS	µmol/L	6301	5041-7561
<b>(Hg)汞</b>			
OTHER Dedicated \ ICP-MS	µmol/L	0.192	0.154-0.23
<b>(I)碘</b>			
OTHER Dedicated \ ICP-MS	µg/L	198	158-238
<b>(K)钾</b>			
OTHER Dedicated \ ICP-MS	mmol/L	31	24.8-37.2
<b>(Li)锂</b>			
OTHER Dedicated \ ICP-MS	mmol/L	0.006	0.005-0.007
<b>(Mg)镁</b>			
OTHER Dedicated \ ICP-MS	mmol/L	2.85	2.28-3.42

<b>REF</b>	760-3	6x3mL	<b>EXP</b>	2027-02-15	<b>LOT</b>	488292F	更新日期：	2025-02
								3 / 3

<b>(Mn)锰</b>			
OTHER Dedicated \ ICP-MS	µg/L	44	35.2-52.8
<b>(Mo)钼</b>			
OTHER Dedicated \ ICP-MS	µg/L	7.7	6.16-9.24
<b>(Na)钠</b>			
OTHER Dedicated \ ICP-MS	mmol/L	62.2	49.8-74.6
<b>(Ni)镍</b>			
OTHER Dedicated \ ICP-MS	µg/L	16.5	13.2-19.8
<b>(P)磷</b>			
OTHER Dedicated \ ICP-MS	mmol/L	6.75	5.4-8.1
<b>(Pb)铅</b>			
OTHER Dedicated \ ICP-MS	µmol/L	2.82	2.26-3.38
<b>(Rb)铷</b>			
OTHER Dedicated \ ICP-MS	mg/L	1.65	1.32-1.98
<b>(S)硫</b>			
OTHER Dedicated \ ICP-MS	mg/L	1100	880-1320
<b>(Sb)锑</b>			
PerkinElmer NexlON Series PerkinElmer \ MS	mg/L	0.03	0.03-0.03
OTHER Dedicated \ ICP-MS	mg/L	0.034	0.027-0.041
<b>(Se)硒</b>			
OTHER Dedicated \ ICP-MS	µmol/L	0.391	0.313-0.469
<b>(Si)硅</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.66	0.528-0.792
<b>(Sn)锡</b>			
OTHER Dedicated \ ICP-MS	µg/L	11	8.8-13.2
<b>(Sr)锶</b>			
OTHER Dedicated \ ICP-MS	mg/L	0.104	0.083-0.125
<b>(Ti)钛</b>			
OTHER Dedicated \ ICP-MS	µg/L	4.4	3.52-5.28
<b>(V)钒</b>			
OTHER Dedicated \ ICP-MS	µg/L	6.6	5.28-7.92
<b>(Zn)锌</b>			
OTHER Dedicated \ ICP-MS	µmol/L	126	101-151

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。