

REF	510-1 6x3mL	EXP 2024-06-18	LOT 30A6143	更新日期： 2025-02 1 / 3
------------	----------------	-----------------------	--------------------	------------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(Alb)白蛋白	(AST/ SGOT)天门冬氨酸氨基转移酶	(ALP)碱性磷酸酶
(DGXN)地高辛	(AFP)甲胎蛋白	(CEA)癌胚抗原
(PSA)前列腺特异性抗原	(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3
(CA 19-9)糖类抗原 19-9	(Ferritin)铁蛋白	(β-hCG)人绒毛膜促性腺激素β亚基
(F-PSA)游离前列腺特异性抗原	(GLU)葡萄糖	(Anti -HAV IgM)抗-甲肝病毒 IgM
(Cyfra 21-1)细胞角蛋白19片段	(SCC)鳞状细胞癌抗原	(CA 50)糖类抗原 50
(CA 72-4)糖类抗原 72-4	(ADA)腺苷脱氨酶	

REF	510-1	6x3mL	EXP	2024-06-18	LOT	30A6143	更新日期:	2025-02
								2 / 3

批号: 30A6143			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
(ADA)腺苷脱氨酶			
AB Sciex API Series AllergScreen \ Peroxidase	U/L	2	2-2
(AFP)甲胎蛋白			
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.21	3.97-4.45
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.05	3.81-4.29
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.31	2.52-4.1
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.14	3.91-4.37
(Alb)白蛋白			
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	1.05	0.91-1.19
(ALP)碱性磷酸酶			
AB Sciex API Series ABX \ dry chemistry	U/L	11.2	11.2-11.2
(Anti -HAV IgM)抗-甲肝病毒 IgM			
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	5	5-5
(AST/ SGOT)天门冬氨酸氨基转移酶			
AB Sciex API Series ABX \ dry chemistry	U/L	11.2	11.2-11.2
(CA 125)糖类抗原 125			
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	8.78	0-25.01
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23	22.3-23.7
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.7	29.1-34.3
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	54.4	47.2-61.6
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.8	14.3-23.3
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.3	32-36.6
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.6	27.5-27.7
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	70.9	70.6-71.2
(CA 15-3)糖类抗原 15-3			
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.9	2-17.8
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.8	14.4-17.2
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.8	13.5-16.1
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	17.5	16.3-18.7
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.8	9-14.6
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.1	22.1-22.1
(CA 19-9)糖类抗原 19-9			
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.07	1.98-16.16
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.2	22-40.4
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.3	11.3-13.3
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.4	25.2-31.6
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.4	29.3-29.5
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.9	23.3-30.5
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.6	8.8-14.4
(CA 50)糖类抗原 50			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.7	10.2-13.2
(CA 72-4)糖类抗原 72-4			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.87	1.69-2.05
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.03	1.86-2.2

REF	510-1	6x3mL	EXP	2024-06-18	LOT	30A6143	更新日期:	2025-02
								3 / 3

(CEA)癌胚抗原				
Roche cobas E601/602				
Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.43		4.94-5.92
Roche cobas 4000/6000/8000				
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.41		2.14-2.68
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.76		2.52-3
Roche cobas E601/602				
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.72		3.65-3.79
Mindray CL-2000i/1000i/6000i series				
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.18		0.9-1.46
Abbott ARCHITECT i series				
Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.79		2.56-3.02
(Cyfra 21-1)细胞角蛋白19片段				
Roche cobas E601/602				
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.02		1.79-2.25
Abbott ARCHITECT i series				
Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29		1.14-1.44
(DGXN)地高辛				
ABX Pentra Series				
ImmunoCAP \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.56		1.56-1.56
(Ferritin)铁蛋白				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	62.8		58.3-67.3
Roche cobas E601/602				
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	58.5		58.5-58.5
Abbott ARCHITECT i series				
Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	66.6		63.1-70.1
(F-PSA)游离前列腺特异性抗原				
Roche cobas E601/602				
Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.994		0.861-1.127
Roche cobas 4000/6000/8000				
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.898		0.86-0.936
Mindray CL-2000i/1000i/6000i series				
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.85		0.646-1.054
(GLU)葡萄糖				
AB Sciex API Series				
ABX \ dry chemistry	mmol/L	62.2		62.2-62.2
(PSA)前列腺特异性抗原				
Roche cobas 4000/6000/8000				
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.86		1.78-1.94
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.75		1.63-1.87
Mindray CL-2000i/1000i/6000i series				
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.71		1.3-2.12
Abbott ARCHITECT i series				
Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.57		1.49-1.65
(SCC)鳞状细胞癌抗原				
Roche cobas E601/602				
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.921		0.391-1.451
maglumi system				
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.39		1.24-1.54
Roche cobas E601/602				
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.95		0.889-1.011
Mindray CL-2000i/1000i/6000i series				
Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.12		0.85-1.39
Abbott ARCHITECT i series				
Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.04		0.88-1.2
(β-hCG)人绒毛膜促性腺激素β亚基				
Roche cobas E601/602				
Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	0.01		0.009-0.011
Roche cobas 4000/6000/8000				
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.013		0.012-0.014

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。