

<b>REF</b>	510-1 6x3mL	<b>EXP</b> 2025-09-26	<b>LOT</b> 30A61AA	更新日期:	2025-02 1 / 3
------------	----------------	-----------------------	--------------------	-------	------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(BNP)B型利钠肽	(AFP)甲胎蛋白	(CEA)癌胚抗原
(hCG)人绒毛膜促性腺激素	(PSA)前列腺特异性抗原	(CA 125)糖类抗原 125
(CA 15-3)糖类抗原 15-3	(CA 19-9)糖类抗原 19-9	(Ferritin)铁蛋白
(β-hCG)人绒毛膜促性腺激素β亚基	(F-PSA)游离前列腺特异性抗原	(SCC)鳞状细胞癌抗原
(TPA)组织多肽抗原	(TPS)组织多肽特异抗原	(CA242)糖类抗原 242

<b>REF</b>	510-1	6x3mL	<b>EXP</b>	2025-09-26	<b>LOT</b>	30A61AA	更新日期:	2025-02
								2 / 3

批号: 30A61AA			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
<b>(AFP)甲胎蛋白</b>			
Roche cobas E601/602	ng/mL	3.68	2.97-4.39
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
Siemens ADVIA Centaur XP/XPT	ng/mL	2.76	2.22-3.3
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
Beckman UniCel Dxl 600/800	ng/mL	4.23	3.41-5.05
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)			
<b>(BNP)B型利钠肽</b>			
AB Sciex API Series	pg/mL	1.82	0.96-2.68
ABX \ Chemiluminescence enzyme immunoassay (CLEIA)			
<b>(CA 125)糖类抗原 125</b>			
Abbott ARCHITECT i series	IU/mL	32.6	26-39.2
Abbott \ Chemiluminescence Immunoassay (CLIA)			
Siemens ADVIA Centaur XP/XPT	IU/mL	28.3	22.8-33.8
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
DiaSorin Liaison /XL	IU/mL	32.6	26.3-38.9
DiaSorin \ Chemiluminescence Immunoassay (CLIA)			
Roche cobas E601/602	IU/mL	29.4	23.5-35.3
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
<b>(CA 15-3)糖类抗原 15-3</b>			
Abbott ARCHITECT i series	IU/mL	15.3	12.4-18.2
Abbott \ Chemiluminescence Immunoassay (CLIA)			
DiaSorin Liaison /XL	IU/mL	14.8	11.9-17.7
DiaSorin \ Chemiluminescence Immunoassay (CLIA)			
Roche cobas E601/602	IU/mL	14.9	12-17.8
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
Siemens ADVIA Centaur XP/XPT	IU/mL	14.9	11.9-17.9
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
<b>(CA 19-9)糖类抗原 19-9</b>			
Roche cobas E601/602	IU/mL	14.2	11.5-16.9
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
Siemens ADVIA Centaur XP/XPT	IU/mL	26.2	21-31.4
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
DiaSorin Liaison /XL	IU/mL	27.6	22.6-32.6
DiaSorin \ Chemiluminescence Immunoassay (CLIA)			
Abbott ARCHITECT i series	IU/mL	45.3	36.5-54.1
Abbott \ Chemiluminescence Immunoassay (CLIA)			
<b>(CA242)糖类抗原 242</b>			
Manual	U/mL	13.7	11.1-16.3
GLD \ Enzyme linked immunosorbent assays (ELISA)			
<b>(CEA)癌胚抗原</b>			
Siemens ADVIA Centaur XP/XPT	ng/mL	1.54	1.24-1.84
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
Roche cobas E601/602	ng/mL	2.73	2.17-3.29
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
Abbott ARCHITECT i series	ng/mL	3.04	2.43-3.65
Abbott \ Chemiluminescence Immunoassay (CLIA)			
Beckman UniCel Dxl 600/800	ng/mL	2.25	1.83-2.67
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)			
<b>(Ferritin)铁蛋白</b>			
Siemens ADVIA Centaur XP/XPT	ng/mL	53.7	43.3-64.1
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
Roche cobas c 501/502/701/702	ng/mL	72.6	58.8-86.4
Roche \ Immunoturbidimetry (ITA)			
Beckman Access	ng/mL	49.7	40.2-59.2
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)			
Abbott ARCHITECT i series	ng/mL	62.8	50.4-75.2
Abbott \ Chemiluminescence Immunoassay (CLIA)			
<b>(F-PSA)游离前列腺特异性抗原</b>			
Beckman Access	ng/mL	1.58	1.27-1.89
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)			
Siemens ADVIA Centaur XP/XPT	ng/mL	0.94	0.733-1.147
Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)			
Roche cobas E601/602	ng/mL	1.11	0.91-1.31
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)			
Abbott ARCHITECT i series	ng/mL	0.95	0.785-1.115
Abbott \ Chemiluminescence Immunoassay (CLIA)			
<b>(hCG)人绒毛膜促性腺激素</b>			
Siemens ADVIA Centaur XP/XPT	IU/L	10.6	8.4-12.8
Siemens \ Enzyme linked immunosorbent assays (ELISA)			
<b>(PSA)前列腺特异性抗原</b>			
Abbott ARCHITECT i series	ng/mL	2.37	1.91-2.83
Abbott \ Chemiluminescence Immunoassay (CLIA)			

<b>REF</b>	510-1	6x3mL	<b>EXP</b>	2025-09-26	<b>LOT</b>	30A61AA	更新日期:	2025-02
								3 / 3

Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.28	1.73-2.83
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.24	1.77-2.71
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.26	1.81-2.71
<b>(SCC)鳞状细胞癌抗原</b>			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.48	1.24-1.72
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.69	1.37-2.01
<b>(TPA)组织多肽抗原</b>			
DiaSorin Liaison /XL DiaSorin \ Chemiluminescence Immunoassay (CLIA )	U/L	392	314-470
<b>(TPS)组织多肽特异抗原</b>			
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	16.4	13.2-19.6
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.014	0.011-0.017

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。