

REF	670-1	6x1mL	EXP	2024-12-20	LOT	3FE866E	更新日期:	2024-08
								1 / 2

批号: 3FE866E						
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数		
(APTT)活化部分凝血活酶时间						
Stago STA series	Seconds	39.4	4.8%	8	218	
DENUO \ Magnetic Beads	Seconds	31.3	4.1%	7	175	
Sysmex CA 1500/6000/7000	Seconds	31.3	6.3%	23	564	
sysmex \ Light scattering	Seconds	33.4	3%	<5	<100	
Sysmex CS 2000i/2100i	Seconds	41.8	36.1%	<5	<100	
sysmex \ Light scattering	Seconds	33.6	2.3%	<5	<100	
Sysmex CS-2400	Seconds	38	4.6%	11	152	
Siemens \ Light scattering	Seconds	38.3	12%	<5	<100	
Sekisui Coapresta 3000	Seconds	38.7	12%	<5	<100	
SekiSui \ Light scattering	Seconds	29.5	10%	<5	<100	
Maccura R01	Seconds	37.5	12%	<5	<100	
MAKER \ Magnetic Beads	Seconds	37.1	12%	<5	<100	
Werfen ACL series	Seconds					
Werfen \ OTHER	Seconds					
Werfen ACL series	Seconds					
Werfen \ Light scattering	Seconds					
Stago STA series	Seconds					
Stago \ Magnetic Beads	Seconds					
zonic XL Series	Seconds					
Zonci \ Magnetic Beads	Seconds					
Sysmex CS 5100	Seconds					
Siemens \ Light scattering	Seconds					
Sysmex CA 1500/6000/7000	Seconds					
Siemens \ Light scattering	Seconds					
(AT-III)抗凝血酶III						
Werfen ACL series	%	102	4.7%	8	<100	
Werfen \ OTHER	%	88.3	6.9%	8	206	
Stago STA series	%	86	10%	<5	<100	
DENUO \ Magnetic Beads	%	94.7	12%	<5	<100	
Sekisui Coapresta 3000	%	96.5	12%	<5	<100	
SekiSui \ Nephelometry	%	95.5	12%	<5	<100	
Sysmex CA 1500/6000/7000	%					
Siemens \ Nephelometry	%					
Stago STA series	%					
Stago \ Magnetic Beads	%					
Werfen ACL series	%					
Dedicated \ Nephelometry	%					
(D-D(DDU)) D-二聚体(DDU)						
zonic XL Series	mg/L	0.43	10%	<5	<100	
Zonci \ Immunoturbidimetry (ITA)	mg/L					
(FIB)纤维蛋白原						
Werfen ACL series	g/L	3.25	8.3%	8	<100	
Werfen \ OTHER	g/L	3.6	7.9%	8	222	
Stago STA series	g/L	3.51	5.2%	7	175	
DENUO \ Magnetic Beads	g/L	2.37	24.8%	23	574	
Sysmex CA 1500/6000/7000	g/L	3.51	6.2%	<5	<100	
sysmex \ Light scattering	g/L	3.51	6.2%	<5	<100	
Sysmex CS 2000i/2100i	g/L	2.87	21.8%	<5	<100	
sysmex \ Light scattering	g/L	3.52	8.9%	<5	<100	
Sysmex CS-2400	g/L	3.92	12%	<5	<100	
Siemens \ Light scattering	g/L	3.4	10%	<5	<100	
Sekisui Coapresta 3000	g/L	3.89	12%	<5	<100	
SekiSui \ Light scattering	g/L	3.8	12%	<5	<100	
Maccura R01	g/L					
MAKER \ Magnetic Beads	g/L					
Werfen ACL series	g/L					
Werfen \ Light scattering	g/L					
zonic XL Series	g/L					
Zonci \ Magnetic Beads	g/L					
Stago STA series	g/L					
Stago \ Magnetic Beads	g/L					
Sysmex CA 1500/6000/7000	g/L					
Siemens \ Light scattering	g/L					
(PT INR)PT国际标准化比值						
Stago STA series	INR	1.07	4.2%	8	212	
DENUO \ Calculate	INR	1.11	5.2%	7	175	
Sysmex CA 1500/6000/7000	INR	1.6	29.5%	20	459	
sysmex \ Calculate	INR	1.02	1.6%	<5	<100	
Sysmex CS 2000i/2100i	INR	1	5.8%	<5	<100	
sysmex \ Calculate	INR	1	5.8%	<5	<100	
Sysmex CS-2400	INR	1.08	5.6%	12	158	
Siemens \ Calculate	INR	1.05	1%	<5	<100	
Maccura R01	INR	1.02	12%	<5	<100	
MAKER \ Calculate	INR	1.02	12%	<5	<100	
Werfen ACL series	INR					
Werfen \ Calculate	INR					
Sekisui Coapresta 3000	INR					
SekiSui \ Calculate	INR					
Stago STA series	INR					
Stago \ Calculate	INR					
zonic XL Series	INR					
Zonci \ Calculate	INR					
Sysmex CA 1500/6000/7000	INR					
Siemens \ Calculate	INR					
(PT)血浆凝血酶原时间						
Stago STA series	Seconds	14.1	3.4%	8	239	
DENUO \ Magnetic Beads	Seconds					

REF	670-1	6x1mL	EXP	2024-12-20	LOT	3FE866E	更新日期：	2024-08
								2 / 2

		批号：3FE866E			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CA 1500/6000/7000	Seconds	12.7	4.9%	7	176
sysmex \ Light scattering					
Sysmex CS 2000i/2100i	Seconds	11.8	3.6%	23	538
sysmex \ Light scattering					
Sekisui Coapresta 3000	Seconds	20.3	52.1%	<5	<100
SekiSui \ Light scattering					
Sysmex CS-2400	Seconds	11.9	1.6%	<5	<100
Siemens \ Light scattering					
Maccura R01	Seconds	12	2.9%	<5	<100
MAKER \ Magnetic Beads					
Werfen ACL series	Seconds	11.8	3.3%	11	140
Werfen \ OTHER					
Stago STA series	Seconds	13.2	12%	<5	<100
Stago \ Magnetic Beads					
Sysmex CS 5100	Seconds	13.3	12%	<5	<100
Siemens \ Light scattering					
Werfen ACL series	Seconds	13	12%	<5	<100
Werfen \ Light scattering					
zonic XL Series	Seconds	12.5	10%	<5	<100
Zonci \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	13	12%	<5	<100
Siemens \ Light scattering					
(TT)凝血酶时间					
Stago STA series	Seconds	19.7	2.7%	8	230
DENUO \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	20.2	6.6%	7	174
sysmex \ Light scattering					
Sysmex CS 2000i/2100i	Seconds	18.2	3.6%	23	571
sysmex \ Light scattering					
Sekisui Coapresta 3000	Seconds	19.7	5.2%	<5	<100
SekiSui \ Light scattering					
Sysmex CS-2400	Seconds	19.2	5.3%	<5	<100
Siemens \ Light scattering					
Werfen ACL series	Seconds	14.4	3.6%	11	157
Werfen \ OTHER					
Werfen ACL series	Seconds	17	12%	<5	<100
Werfen \ Light scattering					
zonic XL Series	Seconds	13.1	10%	<5	<100
Zonci \ Magnetic Beads					
Sysmex CS 5100	Seconds	17.5	12%	<5	<100
Dedicated \ Light scattering					
Stago STA series	Seconds	17.2	12%	<5	<100
Stago \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	18.5	12%	<5	<100
Dedicated \ Light scattering					

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。