

REF 670-1	6x1mL	EXP 2026-06-03	LOT 3FE86D8	更新日期： 2024-08
				1 / 1

批号： 3FE86D8					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	37.2	4.9%	<5	161
Mindray CX Series Mindray \ OTHER	Seconds	31.1	3.2%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ clauss	Seconds	32.1	3.2%	<5	<100
Sysmex CS-2500 Siemens Pathrombin SL(APTT) \ OTHER	Seconds	29.2	10.8%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	31.5	4%	<5	<100
Sysmex CS 5100 SUNBIO \ OTHER	Seconds	34.4	2.8%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	30	10%	<5	<100
Sysmex CS-2500 sysmex \ OTHER	Seconds	30	1.2%	<5	<100
Werfen ACL series OTHER \ OTHER	Seconds	35.9	3%	<5	<100
(AT-III)抗凝血酶III					
Werfen ACL series Werfen Hemosil Liquid Antithrombin (AT \square) \ Nephelometry	%	104	10%	<5	<100
(FIB)纤维蛋白原					
Succeeder SF series Succeeder \ Magnetic Beads	g/L	3.56	4.7%	<5	340
Mindray CX Series Mindray \ OTHER	g/L	3.58	3.5%	<5	142
Werfen ACL series Werfen Hemosil Fib-C \ clauss	g/L	3.17	5.1%	<5	<100
Sysmex CS-2500 Siemens Thrombin (FIB) \ OTHER	g/L	2.73	13.1%	<5	<100
Werfen ACL series Werfen Hemosil QFA Thrombin (FIB) \ Light scattering	g/L	3.7	9.9%	<5	<100
Werfen ACL series SUNBIO \ OTHER	g/L	3	3.8%	<5	<100
Sysmex CS-2500 sysmex \ OTHER	g/L	2.78	8.9%	<5	<100
Succeeder SA series Succeeder \ Magnetic Beads	g/L	3.74	2.4%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	g/L	3.27	0%	<5	<100
Sysmex CS 5100 SUNBIO \ OTHER	g/L	3.22	1.2%	<5	<100
(PT INR)PT国际标准化比值					
Succeeder SF series Succeeder \ Calculate	INR	1.07	3.7%	<5	158
Mindray CX Series Mindray \ Calculate	INR	0.918	2.5%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Calculate	INR	1.03	5.1%	<5	<100
Sysmex CS 5100 SUNBIO \ Calculate	INR	1.11	4.2%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Calculate	INR	0.9	10%	<5	<100
Werfen ACL series OTHER \ Calculate	INR	1.09	1.5%	<5	<100
(TT)凝血酶时间					
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	17.2	2.5%	<5	322
Mindray CX Series Mindray \ OTHER	Seconds	17	1.2%	<5	142
Sysmex CS-2500 Siemens \ OTHER	Seconds	18.9	7.5%	<5	<100
Werfen ACL series Werfen \ clauss	Seconds	19.2	3.6%	<5	106
Sysmex CS-2500 sysmex \ OTHER	Seconds	19.4	6.9%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	19.3	9.9%	<5	<100
Werfen ACL series OTHER \ OTHER	Seconds	14.7	2.1%	<5	<100
Succeeder SA series Succeeder \ Magnetic Beads	Seconds	14.8	3.3%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	18.7	0%	<5	<100
Sysmex CS 5100 SUNBIO \ OTHER	Seconds	16.8	1.7%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。