

REF	670-1	6x1mL	EXP	2026-04-06	LOT	3FE86D6	更新日期:	2024-08
								1 / 2

		批号: 3FE86D6			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
OTHER					
Mindray \ Magnetic Beads	Seconds	30.8	3.1%	9	222
Mindray CX Series	Seconds	31.5	3.9%	6	178
Mindray \ Light scattering					
Stago STA series	Seconds	34.5	10.7%	<5	<100
Stago \ Magnetic Beads					
Mindray CX Series	Seconds	30.7	3.2%	<5	<100
Mindray \ OTHER					
Sysmex XN series	Seconds	29.7	2.2%	12	609
sysmex \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	34	10%	<5	<100
Siemens Dade Actin(APTT) \ Light scattering					
Werfen ACL series	Seconds	28	10%	<5	<100
Werfen Hemosil APTT-SP \ Light scattering					
Mindray EXC810	Seconds	31.2	2.6%	<5	<100
Mindray \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	33	10%	<5	<100
Siemens Dade Actin FS(APTT) \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	37	10%	<5	<100
Siemens Pathrombin SL(APTT) \ Light scattering					
Werfen ACL series	Seconds	32	10%	<5	<100
Werfen Hemosil SynthAsil APTT \ Light scattering					
Sekisui Coapresta 3000	Seconds	32	1.9%	<5	<100
SUNBIO \ OTHER					
Sysmex CS 5100	Seconds	44.4	2.7%	<5	<100
sysmex \ Light scattering					
Dymind DH Series	Seconds	32.6	2%	<5	<100
DYMIND \ Light scattering					
Sysmex CS 5100	Seconds	33.7	4.8%	<5	<100
OTHER \ OTHER					
(AT-III)抗凝血酶III					
Stago STA series	%	99	10%	<5	<100
Stago \ Nephelometry					
Werfen ACL series	%	100	10%	<5	<100
Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry					
Werfen ACL series	%	100	10%	<5	<100
Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry					
Sysmex CS 5100	%	96	1.5%	<5	<100
sysmex \ Nephelometry					
Sysmex CA 1500/6000/7000	%	102	10%	<5	<100
Siemens Innovance Antithrombin (AT□) \ Nephelometry					
Sysmex CA 1500/6000/7000	%	99	10%	<5	<100
Siemens Berichrom Antithrombin III (AT□) \ Nephelometry					
(FIB)纤维蛋白原					
OTHER					
Mindray \ Magnetic Beads	g/L	3.22	10%	9	226
Mindray CX Series	g/L	3.44	3.4%	6	178
Mindray \ clauss					
Stago STA series	g/L	3.51	10.1%	<5	<100
Stago \ Magnetic Beads					
Mindray CX Series	g/L	3.01	2.3%	<5	<100
Mindray \ OTHER					
Sysmex XN series	g/L	3.65	4.4%	12	607
sysmex \ Light scattering					
Werfen ACL series	g/L	3.3	10%	<5	<100
Werfen Hemosil Fib-C \ Light scattering					
Mindray EXC810	g/L	3.32	4.6%	<5	<100
Mindray \ OTHER					
Sysmex CS 5100	g/L	0.035	6.4%	<5	<100
sysmex \ Light scattering					
Werfen ACL series	g/L	3.39	10%	<5	<100
Werfen Hemosil QFA Thrombin (FIB) \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	3.3	10%	<5	<100
Siemens Thrombin (FIB) \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	3.58	10%	<5	<100
Siemens Multifibren U (FIB) \ Light scattering					
Sekisui Coapresta 3000	g/L	3.86	3.4%	<5	<100
SUNBIO \ OTHER					
Dymind DH Series	g/L	3.4	3%	<5	<100
DYMIND \ Magnetic Beads					
Sysmex CS 5100	g/L	3.25	5.8%	<5	<100
OTHER \ OTHER					
(PT)血浆凝血酶原时间					
OTHER					
Mindray \ Magnetic Beads	Seconds	13.3	6.2%	9	223
Mindray CX Series	Seconds	11.1	4.2%	8	216
Mindray \ OTHER					
Sysmex XN series	Seconds	11.7	2.5%	12	608
sysmex \ Light scattering					
Werfen ACL series	Seconds	16	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Stago STA series	Seconds	14.7	10%	<5	<100
STA-Neoplastine CI Plus \ Magnetic Beads					
Stago STA series	Seconds	13.2	10%	<5	<100
Stago \ Magnetic Beads					
Sysmex CS 5100	Seconds	13.9	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Werfen ACL series	Seconds	12.4	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					

REF	670-1	6x1mL	EXP	2026-04-06	LOT	3FE86D6	更新日期:	2024-08
								2 / 2

		批号: 3FE86D6			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray EXC810 Mindray \ OTHER	Seconds	12.1	2%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	12.6	4.2%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	12.3	10%	<5	<100
Sekisui Coapresta 3000 SUNBIO \ OTHER	Seconds	11.8	1.5%	<5	<100
Dymind DH Series DYMIND \ Magnetic Beads	Seconds	13	2.5%	<5	<100
Sysmex CS 5100 OTHER \ OTHER	Seconds	13	3.5%	<5	<100
(TT)凝血酶时间					
OTHER Mindray \ Magnetic Beads	Seconds	14.4	5.1%	9	225
Mindray CX Series Mindray \ OTHER	Seconds	18.1	6.1%	8	210
Sysmex XN series sysmex \ Light scattering	Seconds	18.5	3.1%	12	609
Mindray EXC810 Mindray \ OTHER	Seconds	15.6	4%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	19.4	10%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	17.3	2.9%	<5	<100
Dymind DH Series DYMIND \ Magnetic Beads	Seconds	20.5	2.3%	<5	<100
Sekisui Coapresta 3000 SUNBIO \ OTHER	Seconds	14.6	1.5%	<5	<100
Sysmex CS 5100 OTHER \ OTHER	Seconds	19.9	3%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。