

<b>REF</b>	670-1	6x1mL	<b>EXP</b>	2023-11-20	<b>LOT</b>	3FE8609	更新日期:	2024-08
								1 / 3

		批号: 3FE8609			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(APTT)活化部分凝血活酶时间</b>					
Werfen ACL series	Seconds	38	4.6%	8	126
Werfen \ OTHER					
Sysmex CS-2400	Seconds	31.8	8.6%	6	<100
Siemens \ Light scattering					
Sekisui Coapresta 3000	Seconds	36.1	3.9%	6	182
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	31	2.6%	6	116
Siemens \ OTHER					
Sysmex CS 2000i/2100i	Seconds	32.2	3.5%	6	100
sysmex \ Light scattering					
Sysmex CS-1300	Seconds	36.1	2.6%	5	<100
SUNBIO \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	36.4	2.4%	5	<100
SUNBIO \ OTHER					
Succeeder SF series	Seconds	39.8	4.4%	22	415
Succeeder \ Magnetic Beads					
Sekisui Coapresta 3000	Seconds	31.3	0.6%	21	199
SekiSui \ Light scattering					
Sysmex CS 5100	Seconds	35.4	4.8%	<5	<100
SUNBIO \ Light scattering					
Mindray CL-2000i/1000i/6000i series	Seconds	31.1	2.1%	<5	<100
Mindray \ Light scattering					
Sysmex CS-1300	Seconds	25.7	7.1%	18	478
sysmex \ Light scattering					
Werfen ACL series	Seconds	37.8	3.2%	13	251
OTHER \ OTHER					
Sysmex CS 5100	Seconds	37.1	12%	<5	<100
Siemens \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	36.7	12%	<5	<100
Siemens \ Light scattering					
Stago STA series	Seconds	37.9	12%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	Seconds	37.9	12%	<5	<100
Werfen \ Light scattering					
Sysmex CS 5100	Seconds	33.6	3.6%	<5	<100
sysmex \ Magnetic Beads					
<b>(AT-III)抗凝血酶III</b>					
Sysmex CA 1500/6000/7000	%	91.7	3.7%	8	<100
SUNBIO \ OTHER					
Sysmex CS-1300	%	92.5	3.5%	8	<100
SUNBIO \ OTHER					
Werfen ACL series	%	100	5.1%	8	125
Werfen \ OTHER					
Sysmex CS 5100	%	91.2	7.9%	<5	<100
SUNBIO \ Nephelometry					
Sysmex CA 1500/6000/7000	%	93.8	12%	<5	<100
Siemens \ Nephelometry					
Werfen ACL series	%	94.6	12%	<5	<100
Dedicated \ Nephelometry					
Stago STA series	%	98	12%	<5	<100
Stago \ Magnetic Beads					
<b>(FIB)纤维蛋白原</b>					
Succeeder SF series	g/L	3.23	5.6%	8	189
Succeeder \ Light scattering					
Werfen ACL series	g/L	3.31	8.3%	8	131
Werfen \ OTHER					
Sysmex CS-2400	g/L	3.46	6.4%	7	<100
Siemens \ Light scattering					
Sysmex CS 2000i/2100i	g/L	2.73	11.6%	6	<100
sysmex \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	3.48	5.4%	6	117
Siemens \ OTHER					
Sysmex CS-1300	g/L	3.24	3%	5	<100
SUNBIO \ OTHER					
Sysmex CA 1500/6000/7000	g/L	3.36	4%	5	<100
SUNBIO \ OTHER					
Sysmex CS 5100	g/L	2.98	14%	<5	<100
sysmex \ Magnetic Beads					
Mindray CL-2000i/1000i/6000i series	g/L	2.72	1.9%	<5	<100
Mindray \ Light scattering					
Sysmex CS-1300	g/L	2.39	4.5%	18	476
sysmex \ Light scattering					
Sekisui Coapresta 3000	g/L	2.6	2.6%	18	178
SekiSui \ Light scattering					
Succeeder SF series	g/L	3.52	8.8%	14	233
Succeeder \ Magnetic Beads					
Werfen ACL series	g/L	3.36	5.9%	13	266
SUNBIO \ OTHER					
Sekisui Coapresta 3000	g/L	2.67	12.8%	10	257
SUNBIO \ Light scattering					
Werfen ACL series	g/L	3.52	10.9%	<5	<100
MEDCONN \ Light scattering					
Werfen ACL series	g/L	3.88	12%	<5	<100
Werfen \ Light scattering					
Stago STA series	g/L	3.79	12%	<5	<100
Stago \ Magnetic Beads					
Sysmex CA 1500/6000/7000	g/L	3.76	12%	<5	<100
Siemens \ Light scattering					

<b>REF</b> 670-1	6x1mL	<b>EXP</b> 2023-11-20	<b>LOT</b> 3FE8609	更新日期: 2024-08
				2 / 3

		批号: 3FE8609			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(PT INR)PT国际标准化比值</b>					
Werfen ACL series	INR	1.05	6.5%	9	157
Werfen \ Calculate					
Sysmex CA 1500/6000/7000	INR	1.05	1.6%	6	<100
SUNBIO \ Calculate					
Sysmex CS-2400	INR	1.08	3%	6	<100
Siemens \ Calculate					
Sekisui Coapresta 3000	INR	1	1.9%	6	181
SUNBIO \ Calculate					
Sysmex CA 1500/6000/7000	INR	1.09	4.3%	6	116
Siemens \ OTHER					
Sysmex CS 2000i/2100i	INR	1.11	7.1%	6	101
sysmex \ Calculate					
Sysmex CS-1300	INR	1.07	1.5%	5	<100
SUNBIO \ Calculate					
Succeeder SF series	INR	1.07	5.2%	22	437
Succeeder \ Calculate					
Stago STA series	INR	1.02	11.9%	<5	<100
Stago \ Calculate					
Sysmex CS 5100	INR	1.23	2.9%	<5	<100
sysmex \ Calculate					
Mindray CL-2000i/1000i/6000i series	INR	0.961	1.1%	<5	<100
Mindray \ Calculate					
Sysmex CS-1300	INR	1.03	3.3%	18	475
sysmex \ Calculate					
Werfen ACL series	INR	1.09	2.2%	12	241
OTHER \ Calculate					
Werfen ACL series	INR	1.07	1.4%	<5	<100
MEDCONN \ Calculate					
Sysmex CA 1500/6000/7000	INR	1.01	12%	<5	<100
Siemens \ Calculate					
<b>(PT)血浆凝血酶原时间</b>					
Succeeder SF series	Seconds	13.2	4.2%	8	189
Succeeder \ Light scattering					
Werfen ACL series	Seconds	11.8	4.1%	8	125
Werfen \ OTHER					
Sysmex CS 2000i/2100i	Seconds	12.9	6.5%	6	<100
sysmex \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	12.6	1.3%	6	<100
SUNBIO \ OTHER					
Sysmex CS-2400	Seconds	12.6	3%	6	<100
Siemens \ Light scattering					
Sekisui Coapresta 3000	Seconds	12.4	6.8%	6	181
SUNBIO \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	12.3	3.3%	6	109
Siemens \ OTHER					
Sysmex CS-1300	Seconds	12.9	1.5%	5	<100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	12.1	1.1%	21	207
SekiSui \ Light scattering					
Sysmex CS 5100	Seconds	13.5	6.9%	<5	<100
SUNBIO \ Light scattering					
Sysmex CS 5100	Seconds	14.3	2.8%	<5	<100
sysmex \ Magnetic Beads					
Mindray CL-2000i/1000i/6000i series	Seconds	10.5	2.5%	<5	<100
Mindray \ Light scattering					
Sysmex CS-1300	Seconds	12	4.8%	18	475
sysmex \ Light scattering					
Succeeder SF series	Seconds	13.4	8.7%	13	251
Succeeder \ Magnetic Beads					
Werfen ACL series	Seconds	11.8	2.1%	13	250
OTHER \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	12.9	12%	<5	<100
Siemens \ Light scattering					
Stago STA series	Seconds	13.2	12%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	Seconds	12.9	12%	<5	<100
Werfen \ Light scattering					
Sysmex CS 5100	Seconds	13.2	12%	<5	<100
Siemens \ Light scattering					
<b>(TT)凝血酶时间</b>					
Werfen ACL series	Seconds	14.6	4.1%	7	<100
Werfen \ OTHER					
Sysmex CS-2400	Seconds	19.5	3.7%	7	<100
Siemens \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	17	3.6%	6	<100
SUNBIO \ OTHER					
Sysmex CA 1500/6000/7000	Seconds	19.7	4.7%	6	119
Siemens \ OTHER					
Sysmex CS-1300	Seconds	16.7	2.5%	<5	<100
SUNBIO \ OTHER					
Sekisui Coapresta 3000	Seconds	16.9	1.1%	21	201
SekiSui \ Light scattering					
Sysmex CS 5100	Seconds	17.3	13.4%	<5	<100
sysmex \ Magnetic Beads					
Mindray CL-2000i/1000i/6000i series	Seconds	15.4	10.4%	<5	<100
Mindray \ Light scattering					
Sysmex CS-1300	Seconds	22.5	6.3%	18	476
sysmex \ Light scattering					
Succeeder SF series	Seconds	17.6	3.8%	14	236
Succeeder \ Magnetic Beads					

<b>REF</b>	670-1	6x1mL	<b>EXP</b>	2023-11-20	<b>LOT</b>	3FE8609	更新日期：	2024-08
								3 / 3

		批号： 3FE8609			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Werfen ACL series OTHER \ OTHER	Seconds	14.3	3.9%	13	270
Sekisui Coapresta 3000 SUNBIO \ Light scattering	Seconds	18.4	5%	11	258
Werfen ACL series MEDCONN \ Light scattering	Seconds	13.5	1.2%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	16.8	12%	<5	<100
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	17.3	12%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	18.3	12%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	16.6	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。