

<b>REF</b>	539-2	6x3mL	<b>EXP</b>	2026-11-29	<b>LOT</b>	336CA08	更新日期: 2025-02
							1 / 5

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(CK-MB)肌酸激酶-MB (活性)	(Myo / Mb)肌红蛋白	(TnT)肌钙蛋白T
(NT-proBNP)N-末端脑钠肽前体	(DGXN)地高辛	(THEO)茶碱
(TT3)总三碘甲状腺氨酸	(FT3)游离T3	(TT4)总T4
(FT4 )游离T4	(TSH)促甲状腺素	(Cortisol)皮质醇
(FSH)促卵泡刺激素	(LH)促黄体生成素	(PROG)孕酮
(PRL)催 / 泌乳素	(TES)睾酮	(E2)雌二醇
(C-P)C-肽	(FA)叶酸	(INS)胰岛素
(VB12)维生素B12	(IgE)免疫球蛋白 IgE	(AFP)甲胎蛋白
(CEA)癌胚抗原	(hCG)人绒毛膜促性腺激素	(PSA)前列腺特异性抗原
(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3	(CA 19-9)糖类抗原 19-9
(β2M)β-2-微球蛋白	(Ferritin)铁蛋白	(β-hCG)人绒毛膜促性腺激素β亚基
(F-PSA)游离前列腺特异性抗原	(HCY)同型半胱氨酸	(Cyfra 21-1)细胞角蛋白19片段
(SCC)鳞状细胞癌抗原	(ASD)雄烯二酮	(ALD)醛固酮
(Anti-Tg)抗甲状腺球蛋白抗体	(Anti-TPO)抗甲状腺过氧化物酶抗体	(DHEA-S)硫酸脱氢表雄酮
(EPO)促红细胞生成素	(HGH)人生长激素	(IGF1)胰岛素生长因子1
(PTH)甲状旁腺素	(Tg)甲状腺球蛋白	(25-OH-Vit D)25羟基维生素D
(HE4)人附睾蛋白4	(IntactPTH)全段甲状旁腺激素	(HS-CTnl) (高) 超敏肌钙蛋白I

<b>REF</b>	539-2	6x3mL	<b>EXP</b>	2026-11-29	<b>LOT</b>	336CA08	更新日期:	2025-02
								2 / 5

批号: 336CA08			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
<b>(25-OH-Vit D)25羟基维生素D</b>			
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	44.5	35.6-53.4
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	nmol/L	111	89-133
<b>(AFP)甲胎蛋白</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	118	104-132
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	118	108-128
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	119	104-134
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	119	95-143
<b>(ALD)醛固酮</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	pg/mL	163	158-168
Agilent agilent \ Chemiluminescence Immunoassay (CLIA )	pg/mL	182	146-218
<b>(Anti-Tg)抗甲状腺球蛋白抗体</b>			
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	IU/mL	34.3	32-36.6
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	21.6	14-29.2
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.2	40.2-60.2
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	IU/mL	7.35	5.85-8.85
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	7.51	6.01-9.01
<b>(Anti-TPO)抗甲状腺过氧化物酶抗体</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	37.7	33.5-41.9
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	32.7	26.1-39.3
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	IU/mL	21.5	17.2-25.8
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	43.5	34.9-52.1
<b>(ASD)雄烯二酮</b>			
Agilent agilent \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.35	2.68-4.02
<b>(CA 125)糖类抗原 125</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	51.8	48.5-55.1
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	73.8	70.1-77.5
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	66.9	57.9-75.9
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/mL	80.9	64.7-97.1
<b>(CA 15-3)糖类抗原 15-3</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	63.9	58.8-69
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.2	45.7-54.7
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	60.3	54.1-66.5
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/mL	62.6	50.1-75.1
<b>(CA 19-9)糖类抗原 19-9</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	55.8	51.7-59.9
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	53.4	46-60.8
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/mL	39.3	31.4-47.2
<b>(CEA)癌胚抗原</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.2	9.4-11
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	7.2	6.55-7.85
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	13.6	11.2-16
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	20.6	16.5-24.7
<b>(CK-MB)肌酸激酶-MB (活性)</b>			
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	15.9	12.7-19.1

<b>REF</b>	539-2	6x3mL	<b>EXP</b>	2026-11-29	<b>LOT</b>	336CA08	更新日期:	2025-02
								3 / 5

<b>(Cortisol)皮质醇</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	189	180-198
Tosoh AIA Series TOSOH \ Enzyme immunoassay (EIA)	ng/mL	214	171-257
<b>(C-P)C-肽</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.31	2.16-2.46
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.42	1.31-1.53
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.95	1.67-2.23
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.16	1.73-2.59
<b>(Cyfra 21-1)细胞角蛋白19片段</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.734	0.599-0.869
<b>(DGXN)地高辛</b>			
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.93	1.54-2.32
<b>(DHEA-S)硫酸脱氢表雄酮</b>			
Tosoh AIA Series TOSOH \ Enzyme immunoassay (EIA)	ng/mL	5822	4658-6986
<b>(E2)雌二醇</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	pmol/L	928	846-1010
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	831	633-1029
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	837	721-953
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	960	768-1152
<b>(EPO)促红细胞生成素</b>			
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA )	mIU/mL	51	40.8-61.2
<b>(FA)叶酸</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.9	9.7-12.1
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	11.1	8.8-13.4
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.1	8.1-12.1
<b>(Ferritin)铁蛋白</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	276	237-315
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	276	236-316
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	186	149-223
<b>(F-PSA)游离前列腺特异性抗原</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.95	3.73-4.17
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.32	3.61-5.03
<b>(FSH)促卵泡刺激素</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/L	24.2	22.4-26
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	26.1	23.9-28.3
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	21.4	16.9-25.9
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/L	30.6	24.5-36.7
<b>(FT3)游离T3</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	pmol/L	12.1	11-13.2
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	pmol/L	11.7	11.3-12.1
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	9.93	8.6-11.26
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	10.6	8.5-12.7
<b>(FT4 )游离T4</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	pmol/L	34.6	32.2-37
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	pmol/L	25.6	24.7-26.5
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	23.7	21-26.4
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	33.5	26.8-40.2

<b>REF</b>	539-2	6x3mL	<b>EXP</b>	2026-11-29	<b>LOT</b>	336CA08	更新日期:	2025-02
								4 / 5

<b>(hCG)人绒毛膜促性腺激素</b>			
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/L	203	162-244
<b>(HCY)同型半胱氨酸</b>			
Tosoh AIA Series TOSOH \ Enzyme immunoassay (EIA)	μmol/L	21.9	17.5-26.3
<b>(HE4)人附睾蛋白4</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	66	50.7-81.3
<b>(HGH)人生长激素</b>			
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	μg/L	3.83	3.06-4.6
<b>(HS-CTnI) (高) 超敏肌钙蛋白I</b>			
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.315	0.252-0.378
<b>(IgE)免疫球蛋白 IgE</b>			
Tosoh AIA Series TOSOH \ Enzyme linked immunosorbent assays (ELISA)	IU/mL	85.5	68.4-102.6
<b>(IGF1)胰岛素生长因子1</b>			
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA )	nmol/L	18.4	14.7-22.1
<b>(INS)胰岛素</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	mIU/L	27.3	24.9-29.7
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	mIU/L	31.7	28.8-34.6
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	24.8	22.8-26.8
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	mIU/L	31.6	25.3-37.9
<b>(IntactPTH)全段甲状旁腺激素</b>			
Abbott ARCHITECT i series Abbott \ OTHER	pg/mL	15	9-21
<b>(LH)促黄体生成素</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/L	23.2	20-26.4
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	15.5	13.8-17.2
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	16.2	13.4-19
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	IU/L	17.7	14.2-21.2
<b>(Myo / Mb)肌红蛋白</b>			
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	μg/L	150	120-180
<b>(NT-proBNP)N-末端脑钠肽前体</b>			
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	1037	830-1244
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	122	98-146
<b>(PRL)催/泌乳素</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.6	13.3-15.9
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.8	17-18.6
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	16.1	15.2-17
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	19.3	15.4-23.2
<b>(PROG)孕酮</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	11.1	10.1-12.1
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.1	9.6-12.6
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.7	9.4-12
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11.6	9.3-13.9
<b>(PSA)前列腺特异性抗原</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.9	10.2-11.6
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	11.4	10.7-12.1
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.5	10.6-14.4
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	12.9	10.3-15.5
<b>(PTH)甲状旁腺素</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	pmol/L	7.85	6.63-9.07

<b>REF</b>	539-2	6x3mL	<b>EXP</b>	2026-11-29	<b>LOT</b>	336CA08	更新日期:	2025-02
								5 / 5

maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	pmol/L	9.16	7.57-10.75
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	9.03	7.22-10.84
<b>(SCC)鳞状细胞癌抗原</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.918	0.749-1.087
<b>(TES)睾酮</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.412	0.331-0.493
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.406	0.245-0.567
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.459	0.333-0.585
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.487	0.39-0.584
<b>(Tg)甲状腺球蛋白</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	9.05	0-20.82
<b>(THEO)茶碱</b>			
Roche cobas c 501/502/701/702 Roche \ Chemiluminescence Immunoassay (CLIA )	µg/mL	23.6	18.9-28.3
<b>(TnT)肌钙蛋白T</b>			
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.098	0.078-0.118
<b>(TSH)促甲状腺素</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	mIU/L	10.6	9.8-11.4
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	mIU/L	9.47	8.58-10.36
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	8.82	7.87-9.77
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	mIU/L	10.4	8.3-12.5
<b>(TT3)总三碘甲状腺原氨酸</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.14	1.99-2.29
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.92	1.84-2
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.04	1.66-2.42
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.15	2.75-3.55
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.66	2.26-3.06
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.27	1.77-2.77
<b>(TT4)总T4</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	130	119-141
Caris200 INNODX \ Chemiluminescence Immunoassay (CLIA )	ng/mL	82.3	79.5-85.1
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	129	109-149
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	132	106-158
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	130	104-156
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	100	80-120
<b>(VB12)维生素B12</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	pg/mL	731	706-756
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pg/mL	608	510-706
Tosoh AIA Series TOSOH \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	711	569-853
<b>(β2M)β-2-微球蛋白</b>			
Tosoh AIA Series TOSOH \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1550	1240-1860
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>			
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.227	0.203-0.251

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。