

|       |       |       |     |            |     |         |       |         |
|-------|-------|-------|-----|------------|-----|---------|-------|---------|
| REF   | 551-3 | 6×1mL | EXP | 2023-09-02 | LOT | 3493F69 | 更新日期： | 2024-04 |
| 1 / 1 |       |       |     |            |     |         |       |         |

|                                                                                               |       | 批号：3493F69 |       |          |      |
|-----------------------------------------------------------------------------------------------|-------|------------|-------|----------|------|
| 项目\仪器\试剂\方法                                                                                   | 单位    | 均值 \ CV    |       | Lab \ 点数 |      |
| (AFP)甲胎蛋白                                                                                     |       |            |       |          |      |
| Beckman UniCel DxI 600/800<br>Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)        | ng/mL | 39.8       | 6%    | 18       | 104  |
| Siemens Immulite1000/2000<br>Dedicated \ Chemiluminescence Immunoassay (CLIA )                | ng/mL | 142        | 3.3%  | 13       | 216  |
| Tosoh AIA 900/1800<br>Dedicated \ Chemiluminescence Immunoassay (CLIA )                       | ng/mL | 127        | 10.5% | <5       | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)        | ng/mL | 136        | 9.7%  | <5       | <100 |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )                   | ng/mL | 123        | 9.8%  | <5       | <100 |
| (FE3/uE3)游离雌三醇                                                                                |       |            |       |          |      |
| Beckman UniCel DxI 600/800<br>Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 4.97       | 3.4%  | 8        | <100 |
| Beckman UniCel DxI 600/800<br>Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)        | ng/mL | 1.93       | 9.3%  | 16       | <100 |
| Siemens Immulite1000/2000<br>Dedicated \ Chemiluminescence Immunoassay (CLIA )                | ng/mL | 7.66       | 5.9%  | 14       | 216  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)        | ng/mL | 3.05       | 10.2% | <5       | <100 |
| Tosoh AIA 900/1800<br>Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)                | ng/mL | 4.12       | 9.8%  | <5       | <100 |
| (Fβ-hCG)游离β-绒毛膜促性腺激素                                                                          |       |            |       |          |      |
| Siemens Immulite1000/2000<br>Dedicated \ Chemiluminescence Immunoassay (CLIA )                | ng/mL | 117        | 6.2%  | 13       | 208  |
| Tosoh AIA 900/1800<br>Dedicated \ Electro chemi luminescence (ECL)                            | ng/mL | 103        | 11.7% | <5       | <100 |
| (hCG)人绒毛膜促性腺激素                                                                                |       |            |       |          |      |
| Beckman UniCel DxI 600/800<br>Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA) | IU/L  | 70635      | 0.9%  | 7        | 178  |
| Roche cobas e411<br>Roche \ Enzyme linked immunosorbent assays (ELISA)                        | IU/L  | 75.8       | 10.4% | <5       | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)        | IU/L  | 3721       | 10%   | <5       | <100 |
| (PAPP-A)妊娠相关血浆蛋白-A                                                                            |       |            |       |          |      |
| Siemens Immulite1000/2000<br>Dedicated \ Chemiluminescence Immunoassay (CLIA )                | IU/L  | 2.26       | 4.5%  | 13       | 212  |
| Tosoh AIA 900/1800<br>Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)                | IU/L  | 1.05       | 10.7% | <5       | <100 |
| Roche cobas e411<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)                     | IU/L  | 983        | 10%   | <5       | <100 |
| (β-hCG)人绒毛膜促性腺激素β亚基                                                                           |       |            |       |          |      |
| Beckman UniCel DxI 600/800<br>Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)        | IU/mL | 3.09       | 5.6%  | 17       | 104  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )                   | IU/mL | 2.06       | 10.7% | <5       | <100 |

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。