

REF	510-2	6x3mL	EXP	2026-07-22	LOT	30A891C	更新日期:	2024-07
								1 / 2

批号: 30A891C						
项目\仪器\试剂\方法	单位	均值	CV	Lab	点数	
(AFP)甲胎蛋白						
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	72.8	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	56.9	12%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.25	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	76.1	12%	<5	<100	
(CA 125)糖类抗原 125						
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	268	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	165	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	193	12%	<5	<100	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	241	12%	<5	<100	
(CA 15-3)糖类抗原 15-3						
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	54	12%	<5	<100	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	61.2	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	67.5	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	63	12%	<5	<100	
(CA 19-9)糖类抗原 19-9						
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	82	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	366	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	176	12%	<5	<100	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	166	12%	<5	<100	
(CA 72-4)糖类抗原 72-4						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.32	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.4	12%	<5	<100	
(CEA)癌胚抗原						
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	44	12%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.82	12%	<5	<100	
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	32.6	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	40.5	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	26.4	12%	<5	<100	
(Cyfra 21-1)细胞角蛋白19片段						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.85	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.3	12%	<5	<100	
(Ferritin)铁蛋白						
Mindray CL-2000i/1000i/6000i series Mindray \ Immunonephelometry (INA)	ng/mL	685	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	404	12%	<5	<100	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	207	12%	<5	<100	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	266	12%	<5	<100	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	310	12%	<5	<100	
(F-PSA)游离前列腺特异性抗原						
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.88	12%	<5	<100	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.65	12%	<5	<100	

REF	510-2	6x3mL	EXP	2026-07-22	LOT	30A891C	更新日期： 2024-07	2 / 2
------------	-------	-------	------------	------------	------------	---------	---------------	-------

		批号： 30A891C			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.39	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.8	12%	<5	<100
(PSA)前列腺特异性抗原					
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.2	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.4	12%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.6	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.3	12%	<5	<100
(SCC)鳞状细胞癌抗原					
Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.2	12%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.63	12%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	7340	12%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	75	12%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.154	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.09	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。