

| | | | | | | | | |
|------------|-------|-------|------------|------------|------------|---------|-------|---------|
| REF | 510-2 | 6x3mL | EXP | 2026-01-08 | LOT | 30A8915 | 更新日期: | 2024-07 |
| | | | | | | | | 1 / 2 |

| | | 批号: 30A8915 | | | |
|--|-------|-------------|-------|----------|------|
| 项目\仪器\试剂方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| (AFP)甲胎蛋白 | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 69.1 | 2.3% | 8 | 204 |
| Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 69.8 | 10.9% | <5 | <100 |
| Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 73.5 | 11.2% | <5 | <100 |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 80.7 | 11.3% | <5 | <100 |
| Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 72.8 | 12% | <5 | <100 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 75.4 | 10% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | ng/mL | 87.9 | 10.2% | <5 | <100 |
| AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 78.6 | 0% | <5 | <100 |
| (CA 125)糖类抗原 125 | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 260 | 3.6% | 8 | 203 |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 263 | 8.4% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 243 | 9.5% | <5 | <100 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 174 | 10% | <5 | <100 |
| Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 112 | 9.8% | <5 | <100 |
| AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA) | IU/mL | 235 | 0% | <5 | <100 |
| (CA 15-3)糖类抗原 15-3 | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 67.9 | 5.2% | 8 | 205 |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 88.4 | 10.2% | <5 | <100 |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 109 | 10.1% | <5 | <100 |
| Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 108 | 10.2% | <5 | <100 |
| DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 96.9 | 9.3% | <5 | <100 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 59.8 | 10% | <5 | <100 |
| (CA 19-9)糖类抗原 19-9 | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 362 | 5.5% | 8 | 202 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 81.8 | 10% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 121 | 9.1% | <5 | <100 |
| Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | IU/mL | 94.2 | 8.5% | <5 | <100 |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 268 | 10.8% | <5 | <100 |
| DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 145 | 11% | <5 | <100 |
| AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA) | IU/mL | 356 | 0% | <5 | <100 |
| (CA 50)糖类抗原 50 | | | | | |
| Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 72.7 | 10% | <5 | <100 |
| (CA 72-4)糖类抗原 72-4 | | | | | |
| Mindray BC-6000 series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 6.92 | 3.7% | <5 | 106 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 13 | 10% | <5 | <100 |
| DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 28.9 | 10.7% | <5 | <100 |
| (CA242)糖类抗原 242 | | | | | |
| maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA) | U/mL | 23.7 | 10% | <5 | <100 |
| (CEA)癌胚抗原 | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 42.6 | 2.3% | 8 | 204 |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | ng/mL | 37.6 | 8.5% | <5 | <100 |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 35.8 | 10.6% | <5 | <100 |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 26.2 | 10% | <5 | <100 |

| | | | | | | | | |
|------------|-------|-------|------------|------------|------------|---------|-------|---------|
| REF | 510-2 | 6x3mL | EXP | 2026-01-08 | LOT | 30A8915 | 更新日期: | 2024-07 |
| | | | | | | | | 2 / 2 |

| | | 批号: 30A8915 | | | | | |
|--|-------|-------------|-------|----------|------|--|--|
| 项目\仪器\试剂方法 | 单位 | 均值 \ CV | | Lab \ 点数 | | | |
| Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 38.6 | 10.9% | <5 | <100 | | |
| Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 21.7 | 10.1% | <5 | <100 | | |
| AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 37.7 | 0% | <5 | <100 | | |
| (Cyfra 21-1)细胞角蛋白19片段 | | | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 13.3 | 4.6% | 7 | 202 | | |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 15.4 | 10% | <5 | <100 | | |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 13.2 | 9.1% | <5 | <100 | | |
| (Ferritin)铁蛋白 | | | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 390 | 3.4% | 7 | 201 | | |
| Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 266 | 10.2% | <5 | <100 | | |
| Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA) | ng/mL | 313 | 10% | <5 | <100 | | |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 349 | 8.3% | <5 | <100 | | |
| (F-PSA)游离前列腺特异性抗原 | | | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 4.47 | 3.9% | 8 | 205 | | |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 4.35 | 10% | <5 | <100 | | |
| Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 5.26 | 11.6% | <5 | <100 | | |
| AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 4.65 | 0% | <5 | <100 | | |
| (PSA)前列腺特异性抗原 | | | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 13.5 | 4.4% | 8 | 205 | | |
| Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA) | ng/mL | 11.8 | 9.3% | <5 | <100 | | |
| Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 15.9 | 10.1% | <5 | <100 | | |
| Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 21.3 | 1% | <5 | <100 | | |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 17.9 | 10% | <5 | <100 | | |
| Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 15.4 | 10.4% | <5 | <100 | | |
| (SCC)鳞状细胞癌抗原 | | | | | | | |
| Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 15.2 | 4.1% | 7 | 201 | | |
| OTHER Abbott \ Chemiluminescence Immunoassay (CLIA) | ng/mL | 15.8 | 14.3% | 33 | 672 | | |
| Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 9.73 | 10% | <5 | <100 | | |
| Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 12.9 | 9.3% | <5 | <100 | | |
| (TPA)组织多肽抗原 | | | | | | | |
| DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA) | U/L | 6047 | 11.3% | <5 | <100 | | |
| (TPS)组织多肽特异抗原 | | | | | | | |
| Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA) | U/L | 2218 | 10.5% | <5 | <100 | | |
| (β-hCG)人绒毛膜促性腺激素β亚基 | | | | | | | |
| Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 0.156 | 9.6% | <5 | <100 | | |
| Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA) | IU/mL | 0.163 | 10% | <5 | <100 | | |

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。