

REF	760-1	6x3mL	EXP	2024-05-01	LOT	487D9E5	更新日期:	2025-02
								1 / 2

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(Cu)铜	(Zn)锌	(Al)铝
(Sb)锑	(As)砷	(Cd)镉
(Cr)铬	(Co)钴	(I)碘
(Pb)铅	(Mn)锰	(Hg)汞
(Mo)钼	(Ni)镍	(Se)硒
(Sn)锡	(V)钒	

REF	760-1	6x3mL	EXP	2024-05-01	LOT	487D9E5	更新日期：	2025-02
								2 / 2

批号：487D9E5			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(Al)铝			
OTHER Dedicated \ ICP-MS	mg/L	0.011	0.009-0.013
PerkinElmer NexION Series PerkinElmer \ MS	mg/L	0.02	0.02-0.02
(As)砷			
OTHER Dedicated \ ICP-MS	mg/L	0.002	0.002-0.002
(Cd)镉			
OTHER Dedicated \ ICP-MS	μmol/L	0.003	0.002-0.004
OTHER Dedicated \ ICP-MS	μg/L	0.283	0.226-0.34
(Co)钴			
OTHER Dedicated \ ICP-MS	μmol/L	0.004	0.003-0.005
OTHER Dedicated \ ICP-MS	μg/L	0.222	0.178-0.266
(Cr)铬			
OTHER Dedicated \ ICP-MS	μg/L	0.778	0.622-0.934
(Cu)铜			
OTHER Dedicated \ ICP-MS	μmol/L	10.2	8.2-12.2
PerkinElmer NexION Series PerkinElmer \ MS	μmol/L	7.4	6.22-8.58
AWSA AS9000 Series AWSA \ ICP-MS	μmol/L	14.1	11.3-16.9
(Hg)汞			
OTHER Dedicated \ ICP-MS	μmol/L	7.9	6.32-9.48
(I)碘			
OTHER Dedicated \ ICP-MS	μg/L	24.8	19.8-29.8
(Mn)锰			
OTHER Dedicated \ ICP-MS	μg/L	19.4	15.5-23.3
(Mo)钼			
OTHER Dedicated \ ICP-MS	μg/L	0.949	0.759-1.139
(Ni)镍			
OTHER Dedicated \ ICP-MS	μg/L	2.2	1.76-2.64
(Pb)铅			
OTHER Dedicated \ ICP-MS	μmol/L	0.05	0.04-0.06
AWSA AS9000 Series AWSA \ ICP-MS	μmol/L	48.7	39-58.4
(Sb)锑			
OTHER Dedicated \ ICP-MS	mg/L	0.003	0.002-0.004
(Se)硒			
OTHER Dedicated \ ICP-MS	μmol/L	0.076	0.061-0.091
(Sn)锡			
OTHER Dedicated \ ICP-MS	μg/L	0.212	0.17-0.254
(V)钒			
OTHER Dedicated \ ICP-MS	μg/L	0.263	0.21-0.316
(Zn)锌			
AWSA AS9000 Series AWSA \ ICP-MS	μmol/L	120	96-144
PerkinElmer NexION Series PerkinElmer \ MS	μmol/L	47.5	41.8-53.2
OTHER Dedicated \ ICP-MS	μmol/L	71	56.8-85.2

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。