

<b>REF</b>	513-2 6x3mL	<b>EXP</b> 2024-03-22	<b>LOT</b> 30F1C30	更新日期: 2025-03 1 / 4
------------	----------------	-----------------------	--------------------	------------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(AFP)甲胎蛋白	(CEA)癌胚抗原	(PSA)前列腺特异性抗原
(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3	(CA 19-9)糖类抗原 19-9
(β2M)β-2-微球蛋白	(Ferritin)铁蛋白	(β-hCG)人绒毛膜促性腺激素β亚基
(F-PSA)游离前列腺特异性抗原	(NSE)神经元特异性烯醇化酶	(Cyfra 21-1)细胞角蛋白19片段
(SCC)鳞状细胞癌抗原	(CA 50)糖类抗原 50	(CA 72-4)糖类抗原 72-4
(25-OH-Vit D)25羟基维生素D	(CA242)糖类抗原 242	(HE4)人附睾蛋白4
(cPSA)复合前列腺特异性抗原		

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2024-03-22	<b>LOT</b>	30F1C30	更新日期:	2025-03
								2 / 4

批号: 30F1C30			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
<b>(25-OH-Vit D)25羟基维生素D</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	79.1	72.6-85.6
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	nmol/L	198	182-214
<b>(AFP)甲胎蛋白</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	62.9	26.3-99.5
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52.1	38.7-65.5
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	107	91-123
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	106	93-119
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	77.9	72.1-83.7
<b>(CA 125)糖类抗原 125</b>			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	169	124-214
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	288	246-330
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA )	IU/mL	202	186-218
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	309	292-326
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	350	335-365
<b>(CA 15-3)糖类抗原 15-3</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	117	97-137
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	81.3	63.9-98.7
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	38.2	31.8-44.6
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	116	104-128
<b>(CA 19-9)糖类抗原 19-9</b>			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	90.2	79-101.4
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	443	333-553
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	163	139-187
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	107	98-116
<b>(CA 50)糖类抗原 50</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	66.6	58.9-74.3
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	70.4	68.9-71.9
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	48.7	47.1-50.3
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	68.1	51.8-84.4
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	49.7	43-56.4
<b>(CA 72-4)糖类抗原 72-4</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	11.9	10.4-13.4
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.39	4.78-6
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.15	3.75-6.55
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	4.99	4.46-5.52
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.17	4.36-5.98
<b>(CA242)糖类抗原 242</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	U/mL	35.3	31.9-38.7
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/mL	40.5	23.4-57.6
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/mL	56.8	55.3-58.3

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2024-03-22	<b>LOT</b>	30F1C30	更新日期:	2025-03
								3 / 4

Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA ) maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	U/mL	56.6	43-70.2
<b>(CEA)癌胚抗原</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA ) Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	51.1	19.9-82.3
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.5	16.3-30.7
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA ) DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	29.2	23.7-34.7
	ng/mL	48.8	42.5-55.1
	ng/mL	29.9	28.2-31.6
<b>(cPSA)复合前列腺特异性抗原</b>			
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/L	8366	7730-9002
<b>(Cyfra 21-1)细胞角蛋白19片段</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA ) Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA ) Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.7	15.9-19.5
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	13.7	12.5-14.9
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.1	13.4-14.8
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.3	10.3-12.3
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.5	8.8-14.2
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	10.7-12.5
	ng/mL	10.5	9.5-11.5
<b>(Ferritin)铁蛋白</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	398	327-469
<b>(F-PSA)游离前列腺特异性抗原</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA ) Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.85	3.11-10.59
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.59	4.01-5.17
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.5	4.75-6.25
	ng/mL	4.62	4.42-4.82
<b>(HE4)人附睾蛋白4</b>			
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	274	247-301
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA ) Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	pmol/L	412	341-483
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	362	337-387
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	311	236-386
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	318	233-403
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	320	305-335
	pmol/L	260	232-288
<b>(NSE)神经元特异性烯醇化酶</b>			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.845	0.818-0.872
<b>(PSA)前列腺特异性抗原</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA ) Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15	12-18
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	16.1	11.7-20.5
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.5	12.8-16.2
	ng/mL	15.7	14.9-16.5
<b>(SCC)鳞状细胞癌抗原</b>			
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA ) Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA ) Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA ) Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	13.5	11.6-15.4
	ng/mL	17	15-19
	ng/mL	14.2	13.7-14.7
	ng/mL	16.3	13.9-18.7

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2024-03-22	<b>LOT</b>	30F1C30	更新日期:	2025-03
								4 / 4

Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.9	9.9-13.9
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.3	10.9-17.7
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.2	13.9-14.5
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	17.3	15.8-18.8
<b>(β2M)β-2-微球蛋白</b>			
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	941	841-1041
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>			
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.208	0.203-0.213

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。