

REF	510-2	6x3mL	EXP	2025-09-26	LOT	30A88BA	更新日期:	2024-07
								1 / 2

		批号: 30A88BA					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
(AFP)甲胎蛋白							
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	83.2	9.7%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	60.3	10.1%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	70.5	10.2%	<5	<100		
(CA 125)糖类抗原 125							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	179	11.4%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	288	9.8%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	229	9.9%	<5	<100		
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	276	9.9%	<5	<100		
(CA 15-3)糖类抗原 15-3							
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	105	10%	6	105		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	69.3	10.7%	<5	<100		
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	95.3	9.9%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	96.3	10%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	104	11.3%	<5	<100		
(CA 19-9)糖类抗原 19-9							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	76.8	10.6%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	398	10.1%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	178	10.1%	<5	<100		
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	187	9.5%	<5	<100		
(CA 72-4)糖类抗原 72-4							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	8.23	12.3%	<5	<100		
(CEA)癌胚抗原							
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	26.5	9.1%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.3	9.5%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	35.7	10.9%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	46.3	9.7%	<5	<100		
(Cyfra 21-1)细胞角蛋白19片段							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.3	10.8%	<5	<100		
(Ferritin)铁蛋白							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	352	10.5%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	268	9.2%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	306	9.2%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	409	9.7%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.23	10.3%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.86	10%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.12	9.3%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.03	11.4%	<5	<100		
(hCG)人绒毛膜促性腺激素							
Siemens ADVIA Centaur XP/XPT Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	98.6	10%	<5	<100		

REF	510-2	6x3mL	EXP	2025-09-26	LOT	30A88BA	更新日期：	2024-07 2 / 2
------------	-------	-------	------------	------------	------------	---------	-------	------------------

		批号：30A88BA			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(PSA)前列腺特异性抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.3	9.9%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.3	10.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	16.2	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.9	9.2%	<5	<100
(SCC)鳞状细胞癌抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.85	10.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.8	10.3%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL DiaSorin \ Chemiluminescence Immunoassay (CLIA)	U/L	3826	10.2%	<5	<100
(TPS)组织多肽特异抗原					
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	104	8.7%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.157	8.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。