

质控品名： 特殊生化与蛋白非定值质控品

REF	583-2	6×3mL	EXP	2026-02-19	LOT	379ED7F	更新日期：	2024-05
1 / 3								

		批号：379ED7F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(5'NT)5`核苷酸酶					
Beckman AU series BSBE \ Enzymatic	U/L	179	12.3%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic	U/L	40	7.4%	<5	<100
Beckman AU series chongqiong zybio \ Enzymatic	U/L	90.9	1.6%	<5	<100
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ Enzymatic	U/L	134	2.7%	<5	<100
Abbott ARCHITECT ci series Abbott \ Enzymatic	U/L	186	10%	<5	<100
Beckman AU series Reebio \ Enzymatic	U/L	139	1%	<5	<100
(ACE)血管紧张素转化酶					
Beckman AU series Iprocom \ Enzymatic	U/L	11	10%	<5	<100
Siemens Atellica Siemens \ Enzymatic	U/L	19.2	9.6%	<5	<100
(ADA)腺苷脱氨酶					
Mindray BS series OTHER \ Peroxidase	U/L	147	1.6%	9	230
Mindray BS series Mindray \ Trinder reaction	U/L	129	3.2%	9	227
Roche Hitachi/Hitachi Systems MAKER \ Trinder reaction	U/L	144	2.3%	9	221
Mindray BS series MAKER \ Trinder reaction	U/L	129	1.3%	8	161
Roche Hitachi/Hitachi Systems MAKER \ Peroxidase	U/L	144	2.4%	7	205
Beckman AU series Whitman \ Peroxidase	U/L	165	1.3%	6	<100
Beckman AU series Beckman Coulter \ Peroxidase	U/L	154	8.8%	5	192
Beckman AU series BSBE \ Peroxidase	U/L	134	10.4%	5	124
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	148	7.3%	<5	<100
Siemens ADVIA Chemistry XPT KHB \ Trinder reaction	U/L	194	1.5%	<5	<100
Beckman AU series Iprocom \ Peroxidase	U/L	163	10%	<5	<100
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ Peroxidase	U/L	126	2.2%	<5	<100
Mindray BS series Mindray \ Peroxidase	U/L	142	10%	<5	<100
Abbott ARCHITECT ci series Abbott \ Peroxidase	U/L	135	10%	<5	<100
Siemens Atellica Siemens \ Peroxidase	U/L	149	0.9%	<5	<100
Beckman AU series chongqiong zybio \ Peroxidase	U/L	133	0.8%	<5	<100
Beckman AU series MedicalSystem \ Peroxidase	U/L	152	0.6%	<5	<100
Roche Hitachi/Hitachi Systems XJJC \ Peroxidase	U/L	158	0.7%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Peroxidase	U/L	157	0.7%	<5	<100
(AFU)α-L-岩藻糖苷酶					
Mindray BS series OTHER \ OTHER	U/L	40.4	2.9%	9	231
Beckman AU series BSBE \ CNPF	U/L	24.5	8.4%	7	194
Beckman AU series Beckman Coulter \ CNPF	U/L	36.4	18%	<5	<100
Mindray BS series Mindray \ CNPF	U/L	36.4	4.2%	10	254
Roche Hitachi/Hitachi Systems MedicalSystem \ CNPF	U/L	39.8	1.9%	<5	<100
Beckman AU series Iprocom \ CNPF	U/L	26.5	10%	<5	<100
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ CNPF	U/L	25	3.5%	<5	<100
Roche cobas c 501/502/701/702 Roche \ CNPF	U/L	34.3	3%	<5	<100
Mindray BC-5100/BC-5300 Series LEADMAN \ CNPF	U/L	28.3	3.6%	<5	<100
Roche Hitachi/Hitachi Systems XJJC \ CNPF	U/L	42.8	2.3%	<5	<100
Beckman AU series MedicalSystem \ CNPF	U/L	24	10%	<5	<100
Beckman AU series Reebio \ CNPF	U/L	24.3	0.6%	<5	<100
Roche cobas e 801 MAKER \ CNPF	U/L	23.4	2.2%	<5	<100
Abbott ARCHITECT ci series Abbott \ CNPF	U/L	36	10%	<5	<100
(anti CCP)抗环瓜氨酸抗体					
Beckman AU series OTHER \ immunoturbidimetric	U/mL	51.3	3.2%	<5	<100

质控品名： 特殊生化与蛋白非定值质控品

REF	583-2	6×3mL	EXP	2026-02-19	LOT	379ED7F	更新日期：	2024-05
2 / 3								

		批号：379ED7F			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas e411 Roche \ Chemiluminescence Immunoassay (CLIA)	U/mL	42	10%	<5	<100
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	U/mL	23.7	4.8%	<5	<100
Beckman AU series Beckman Coulter \ immunoturbidimetric	U/mL	43.3	1.1%	<5	<100
Beckman AU series MedicalSystem \ immunoturbidimetric	U/mL	57	1.2%	<5	<100
(ASO)抗链球菌溶血素 O 抗体					
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	IU/mL	111	0.4%	<5	<100
(CG)甘胆酸					
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/L	2	4.5%	<5	<100
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Mindray BS series MedicalSystem \ OTHER	mmol/L	29.6	6.2%	9	231
Mindray BS series Mindray \ PEPC	mmol/L	28.8	4%	9	225
Mindray BS series MAKER \ Colorimetry	mmol/L	28.2	4.8%	8	164
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	26.2	6.1%	7	210
Beckman AU series Beckman Coulter \ PEPC	mmol/L	29.6	17.1%	<5	118
Roche Hitachi/Hitachi Systems MAKER \ Colorimetry	mmol/L	28.5	5.8%	11	238
Siemens ADVIA Chemistry XPT DiaSys \ Electrode	mmol/L	25.1	13.1%	<5	<100
Abbott ARCHITECT ci series Abbott \ Colorimetry	mmol/L	28	10%	<5	<100
Mindray BS series Mindray \ Colorimetry	mmol/L	30.3	10%	<5	<100
Roche cobas e 801 DYMIND \ Electrode	mmol/L	27.8	1.1%	<5	<100
Beckman AU series chongqiong zybio \ PEPC	mmol/L	27.6	6.4%	<5	<100
Roche cobas c 501/502/701/702 Roche \ PEPC	mmol/L	29.9	7.2%	<5	<100
(Cys C)半胱氨酸蛋白酶抑制剂C/ 胱抑素C					
Mindray BS series MedicalSystem \ OTHER	mg/L	1.75	2.2%	9	231
Beckman AU series OTHER \ OTHER	mg/L	1.52	3.7%	6	<100
Beckman AU series chongqiong zybio \ Immunoturbidimetry (ITA)	mg/L	1.95	10.2%	6	154
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	2.07	12.4%	<5	<100
Roche Hitachi/Hitachi Systems DiaSys \ Immunonephelometry (INA)	mg/L	1.59	7.2%	<5	<100
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/L	2.18	5.2%	10	227
Siemens ADVIA Chemistry XPT Dedicated \ Immunonephelometry (INA)	mg/L	2.17	3%	<5	<100
Beckman AU series Iprocom \ Immunoturbidimetry (ITA)	mg/L	2.2	10%	<5	<100
Beckman AU series MedicalSystem \ Immunoturbidimetry (ITA)	mg/L	1.78	0.5%	<5	<100
Roche cobas e 801 DYMIND \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/L	1.67	1.5%	<5	<100
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/L	2.25	10.2%	<5	<100
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	mg/L	1.75	1.4%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Immunonephelometry (INA)	mg/L	1.86	0.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ Immunoturbidimetry (ITA)	mg/L	1.83	10%	<5	<100
Mindray BC-5100/BC-5300 Series LEADMAN \ Immunoturbidimetry (ITA)	mg/L	2.15	3.8%	<5	<100
(FRUC (GSP))果糖胺 (糖化血清蛋白)					
Beckman AU series Beckman Coulter \ NBT	mmol/L	2.38	62.3%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ NBT	mmol/L	2.92	2.5%	<5	<100
Abbott ARCHITECT c series Abbott \ NBT	mmol/L	0.537	4.3%	<5	<100
Roche cobas e411 Roche \ NBT	mmol/L	0.433	10%	<5	<100
(GA)糖化血清白蛋白					
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	%	11.3	2.3%	<5	<100

REF	583-2	6×3mL	EXP	2026-02-19	LOT	379ED7F	更新日期：	2024-05
3 / 3								

批号: 379ED7F					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(HCY)同型半胱氨酸					
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	μmol/L	7.88	0.5%	<5	<100
(Lp(a))脂蛋白a					
Mindray BS series MedicalSystem \ OTHER	mg/L	283	1.7%	7	141
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	306	7.9%	<5	<100
Beckman AU series Iprocom \ Immunoturbidimetry (ITA)	mg/L	347	10%	<5	<100
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	mg/L	286	1.6%	<5	<100
Abbott ARCHITECT ci series Abbott \ Immunoturbidimetry (ITA)	mg/L	318	10%	<5	<100
(NEFA)游离脂肪酸					
Beckman AU series Beckman Coulter \ ACS - ACOD	mmol/L	3.47	9.1%	<5	<100
(RF)类风湿因子					
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	IU/mL	10.1	2%	<5	<100
(TBA)总胆汁酸					
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	9.82	6%	<5	<100
(β-Hb)Beta-羟丁酸					
Siemens ADVIA Chemistry XPT KHB \ Dehydrogenase	mmol/L	0.958	2.3%	<5	<100
Beckman AU series chongqiong zybio \ Dehydrogenase	mmol/L	0.99	10%	<5	<100
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	0.83	9.6%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。