

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								1 / 20

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(K)钾	(Na)钠	(Cl)氯
(Ca)钙	(P)磷	(Urea)尿素
(UA)尿酸	(Cr)肌酐	(Alb)白蛋白
(CHOL/ TC)总胆固醇	(TG)甘油三酯	(ALT/ SGPT)丙氨酸氨基转移酶
(AST/ SGOT)天门冬氨酸氨基转移酶	(TBIL)总胆红素	(ALP)碱性磷酸酶
(AMY)淀粉酶	(CK)肌酸激酶	(LDH)乳酸脱氢酶
(DBIL)直接胆红素	(Fe)铁	(TIBC)总铁结合力
(Mg)镁	(Li)锂	(Cu)铜
(Zn)锌	(GGT)L-γ-谷氨酰基转移酶	(α-HBDH)α-羟丁酸脱氢酶
(HDL-C)高密度脂蛋白胆固醇	(LDL-C)低密度脂蛋白胆固醇	(Lp(a))脂蛋白a
(Apo-A I)载脂蛋白A I	(Apo-B)载脂蛋白B	(CK-MB)肌酸激酶-MB (活性)
(Myo / Mb)肌红蛋白	(Tnl)肌钙蛋白I	(TT3)总三碘甲状原氨酸
(FT3)游离T3	(TT4)总T4	(FT4)游离T4
(TSH)促甲状腺素	(FSH)促卵泡刺激素	(LH)促黄体生成素
(PROG)孕酮	(PRL)催/ 泌乳素	(TES)睾酮
(E2)雌二醇	(IgG)免疫球蛋白 IgG	(IgM)免疫球蛋白 IgM
(IgA)免疫球蛋白 IgA	(IgE)免疫球蛋白 IgE	(C3)补体 C3
(C4)补体 C4	(CRP)C反应蛋白	(RF)类风湿因子
(ASO)抗链球菌溶血素 O 抗体	(TRF)转铁蛋白	(PAB)前白蛋白
(β2M)β-2-微球蛋白	(PRO/ TP)(总) 蛋白	(GLU)葡萄糖
(LA)乳酸	(HCY)同型半胱氨酸	(Cys C)半胱氨酸蛋白酶抑制剂C/ 胱抑素C
(CHE)胆碱酯酶	(LAP)亮氨酸氨基肽酶	(PAMY)胰淀粉酶
(LIP)脂肪酶	(ACP)酸性磷酸酶	(Anti-Tg)抗甲状腺球蛋白抗体
(BUN)尿素氮	(UIBC)不饱和铁结合力	(CO2 (HCO3-))二氧化碳 (碳酸氢根)
(AFU)α-L-岩藻糖苷酶	(TBA)总胆汁酸	(GA)糖化血清白蛋白
(NEFA)游离脂肪酸	(α1M)α1微球蛋白	(ACE)血管紧张素转化酶
(CG)甘胆酸	(ADA)腺苷脱氨酶	(SOD)超氧化物歧化酶
(5`NT)5`核苷酸酶	(GLDH)谷氨酸脱氢酶	(Hap (HPT))触珠蛋白 (结合珠蛋白)
(IV-Col)IV型胶原	(Lp-PLA2)脂蛋白磷脂酶A2	(LDH-I)乳酸脱氢酶同工酶1
(H-FABP)心型脂肪酸结合蛋白	(hs-CRP)超敏C反应蛋白	(sd LDL-C)小而密低密度脂蛋白胆固醇
(IMA)缺血修饰白蛋白	(GR)谷胱甘肽还原酶	(CK-MB mass)肌酸激酶-MB质量

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
2 / 20								

		批号：376E040	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(5'NT)5'核苷酸酶			
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic	U/L	1.62	1.19-2.05
(ACE)血管紧张素转化酶			
Abbott ARCHITECT c series BSBE \ OTHER	U/L	23.7	19.1-28.3
Beckman AU series BSBE \ OTHER	U/L	20.7	13.7-27.7
(ACP)酸性磷酸酶			
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	-1.998	0--1.187
(ADA)腺苷脱氨酶			
Beckman AU series BSBE \ Peroxidase	U/L	3.97	3.24-4.7
URIT 8000 series Uritest \ Peroxidase	U/L	6.99	6.67-7.31
Roche Hitachi/Hitachi Systems chongqiong zybio \ Peroxidase	U/L	188	131-245
Beckman AU series OTHER \ Peroxidase	U/L	6.97	4.8-9.14
Abbott ARCHITECT c series BIOSINO \ Peroxidase	U/L	139	133-145
Roche cobas c 501/502/701/702 MedicalSystem \ Peroxidase	U/L	5.82	5.16-6.48
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	5.2	2.71-7.69
TOSHIBA TBA OTHER \ Peroxidase	U/L	5.8	3.91-7.69
TOSHIBA TBA BSBE \ Peroxidase	U/L	5.16	4.77-5.55
Beckman AU series Reebio \ Peroxidase	U/L	6.67	6.55-6.79
(AFU)α-L-岩藻糖苷酶			
Beckman AU series Autobio \ OTHER	U/L	11.8	6.2-17.4
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	13.6	12.3-14.9
TOSHIBA TBA OTHER \ CNPF	U/L	12.4	10.9-13.9
TOSHIBA TBA BSBE \ OTHER	U/L	13.6	12.8-14.4
(a-HBDH)α-羟丁酸脱氢酶			
Beckman AU series Reebio \ Enzymatic Cycling Assays Method	U/L	508	491-525
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	483	457-509
Beckman AU series OTHER \ OTHER	U/L	476	456-496
Beckman AU series MedicalSystem \ OTHER	U/L	534	491-577
Dirui CS series DIRUI \ OTHER	U/L	481	456-506
Mindray BS series chongqiong zybio \ OTHER	U/L	517	488-546
Mindray BS series Mindray \ DGKC	U/L	451	410-492
Abbott ARCHITECT c series BSBE \ OTHER	U/L	462	408-516
Beckman AU series chongqiong zybio \ OTHER	U/L	487	468-506
Mindray BS series Mindray \ immunoturbidimetric	U/L	482	455-509
Roche Hitachi/Hitachi Systems Roche \ Enzymatic Cycling Assays Method	U/L	402	179-625
Beckman AU series BSBE \ OTHER	U/L	469	420-518
Beckman AU series Reebio \ OTHER	U/L	517	493-541
TOSHIBA TBA OTHER \ OTHER	U/L	528	480-576
Beckman AU series Beckman Coulter \ immunoturbidimetric	U/L	480	409-551
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	473	426-520
Abbott ARCHITECT c series BIOSINO \ DGKC	U/L	514	498-530
Mindray BS series OTHER \ OTHER	U/L	415	354-476
Mindray BS series KangHua(Weifang) \ DGKC	U/L	413	382-444
(Alb)白蛋白			
URIT-500B/330/180/55 Uritest \ OTHER	g/L	44.3	39.8-48.8
Beckman AU series MAKER \ Bromocresol green	g/L	45	41.6-48.4
Roche Hitachi/Hitachi Systems MAKER \ Bromocresol green	g/L	45.9	44.1-47.7



REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
3 / 20								

TOSHIBA TBA	g/L	44.6	39.1-50.1
Autobio \ Bromocresol green			
Beckman AU series	g/L	43.5	42.5-44.5
Dedicated \ Bromocresol green			
ALIEX CA400	g/L	45.3	44.2-46.4
SHLY \ Bromocresol green			
Mindray BS series	g/L	42.1	41.3-42.9
chongqiong zybio \ Bromocresol green			
URIT-500B/330/180/55	g/L	44.5	43-46
Uritest \ Bromocresol green			
TOSHIBA TBA	g/L	45.3	41.7-48.9
chongqiong zybio \ OTHER			
TOSHIBA TBA	g/L	46.4	45-47.8
MedicalSystem \ Bromocresol green			
TOSHIBA TBA	g/L	42.9	39.4-46.4
Mindray \ Bromocresol green			
Beckman AU series	g/L	44.8	42.6-47
Beckman Coulter \ Bromocresol green			
Abbott ARCHITECT c series	g/L	44.2	42.7-45.7
BSBE \ Bromocresol green			
Beckman AU series	g/L	43.2	41.6-44.8
chongqiong zybio \ Bromocresol green			
Beckman AU series	g/L	44.8	43.9-45.7
XJKC \ Bromocresol green			
Mindray BS series	g/L	44.2	42.7-45.7
OTHER \ Bromocresol green			
Beckman AU series	g/L	44.2	36.2-52.2
BSBE \ Bromocresol green			
Roche cobas c 501/502/701/702	g/L	46.3	42.9-49.7
DENUO \ Bromocresol green			
Dirui CS series	g/L	47.4	43.2-51.6
DIRUI \ Bromocresol green			

(ALP)碱性磷酸酶			
URIT-500B/330/180/55	U/L	426	400-452
Uritest \ OTHER			
TOSHIBA TBA	U/L	462	386-538
MedicalSystem \ Rate -AMP buffer			
Beckman AU series	U/L	452	417-487
Reebio \ Rate -AMP buffer			
Roche Hitachi/Hitachi Systems	U/L	469	431-507
MAKER \ OTHER			
Beckman AU series	U/L	508	487-529
Dedicated \ Rate -AMP buffer			
TOSHIBA TBA	U/L	421	343-499
Autobio \ Rate -AMP buffer			
Dirui CS series	U/L	462	434-490
DIRUI \ OTHER			
Mindray BS series	U/L	427	407-447
chongqiong zybio \ OTHER			
URIT-500B/330/180/55	U/L	441	409-473
Uritest \ Rate -AMP buffer			
TOSHIBA TBA	U/L	457	420-494
chongqiong zybio \ OTHER			
Mindray BS series	U/L	483	471-495
Mindray \ Rate -AMP buffer			
TOSHIBA TBA	U/L	461	450-472
Mindray \ Rate -AMP buffer			
Abbott ARCHITECT c series	U/L	451	438-464
BIOSINO \ dry chemistry			
Abbott ARCHITECT c series	U/L	470	422-518
BSBE \ OTHER			
Beckman AU series	U/L	453	357-549
OTHER \ OTHER			
Beckman AU series	U/L	474	417-531
Beckman Coulter \ Rate -AMP buffer			
Beckman AU series	U/L	580	560-600
XJKC \ Rate -AMP buffer			
Mindray BS series	U/L	500	466-534
OTHER \ Rate -AMP buffer			
Roche Hitachi/Hitachi Systems	U/L	480	444-516
BSBE \ Rate -AMP buffer			

(ALT/ SGPT)丙氨酸氨基转移酶			
Beckman AU series	U/L	178	167-189
MAKER \ OTHER			
Beckman AU series	U/L	176	168-184
chongqiong zybio \ OTHER			
Roche Hitachi/Hitachi Systems	U/L	195	173-217
Roche \ UV-rate without P5P			
TOSHIBA TBA	U/L	181	160-202
Autobio \ OTHER			
Beckman AU series	U/L	189	183-195
Dedicated \ UV-rate without P5P			
ALIEX CA400	U/L	183	177-189
SHLY \ OTHER			
Mindray BS series	U/L	187	180-194
chongqiong zybio \ OTHER			
TOSHIBA TBA	U/L	171	158-184
chongqiong zybio \ OTHER			
Roche Hitachi/Hitachi Systems	U/L	183	168-198
MAKER \ OTHER			
TOSHIBA TBA	U/L	179	174-184
KHB \ OTHER			
TOSHIBA TBA	U/L	181	174-188
Mindray \ OTHER			
Abbott ARCHITECT c series	U/L	168	147-189
BSBE \ OTHER			

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								4 / 20

Beckman AU series	U/L	169	128-210
BSBE \ OTHER			
Beckman AU series	U/L	163	149-177
KHB \ OTHER			
Beckman AU series	U/L	170	162-178
Beckman Coulter \ OTHER			
Roche cobas c 501/502/701/702	U/L	166	156-176
DENUO \ UV-rate without P5P			
Dirui CS series	U/L	179	169-189
DIRUI \ OTHER			
Mindray BS series	U/L	174	166-182
OTHER \ UV-rate without P5P			
TOSHIBA TBA	U/L	187	175-199
OTHER \ OTHER			

(AMY)淀粉酶

Beckman AU series	U/L	381	368-394
Dedicated \ PNP-G7			
Beckman AU series	U/L	434	411-457
Reebio \ PNP-G7			
Beckman AU series	U/L	493	457-529
BSBE \ OTHER			
Dirui CS series	U/L	642	597-687
DIRUI \ OTHER			
TOSHIBA TBA	U/L	381	373-389
Autobio \ CNP_G3			
Roche Hitachi/Hitachi Systems	U/L	461	444-478
MAKER \ OTHER			
Beckman AU series	U/L	420	319-521
BSBE \ PNP-G7			
Roche Hitachi/Hitachi Systems	U/L	450	429-471
MAKER \ Maltotetraose-rate			
Roche Hitachi/Hitachi Systems	U/L	453	440-466
WanTai DRD \ PNP-G7			
Beckman AU series	U/L	518	496-540
Beckman Coulter \ OTHER			
Roche cobas c 501/502/701/702	U/L	473	371-575
DENUO \ Maltotetraose-rate			
Beckman AU series	U/L	461	373-549
Beckman Coulter \ PNP-G7			
Abbott ARCHITECT c series	U/L	505	492-518
BIOSINO \ CNP_G3			
Mindray BS series	U/L	530	519-541
WanTai DRD \ PNP-G7			
Abbott ARCHITECT c series	U/L	521	325-717
BSBE \ OTHER			
Roche Hitachi/Hitachi Systems	U/L	529	224-834
Roche \ PNP-G7			
Mindray BS series	U/L	546	520-572
WanTai \ PNP-G7			
Beckman AU series	U/L	493	465-521
Beckman Coulter \ CNP_G3			
TOSHIBA TBA	U/L	458	432-484
WanTai \ PNP-G7			

(Anti-Tg)抗甲状腺球蛋白抗体

Beckman Access	IU/mL	2.05	1.91-2.19
Abbott \ Chemiluminescence Immunoassay (CLIA)			

(Apo-A I)载脂蛋白A I

Beckman AU series	g/L	2.12	1.83-2.41
BSBE \ immunoturbidimetric			
Dirui CS series	g/L	2.37	1.98-2.76
DIRUI \ immunoturbidimetric			
ALIEX CA400	g/L	2.3	2.23-2.37
SHLY \ immunoturbidimetric			
TOSHIBA TBA	g/L	2.13	2.03-2.23
Autobio \ immunoturbidimetric			
Mindray BS series	g/L	1.86	1.51-2.21
Mindray \ immunoturbidimetric			
Beckman AU series	g/L	2.3	2.01-2.59
Reebio \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	2.14	1.81-2.47
MAKER \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	2.12	1.96-2.28
MAKER \ OTHER			
Roche cobas c 501/502/701/702	g/L	1.84	1.39-2.29
DENUO \ immunoturbidimetric			
KHB ZY series	g/L	2.18	2.01-2.35
KHB \ immunoturbidimetric			
Beckman AU series	g/L	2.13	2.05-2.21
chongqiong zybio \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	1.98	1.77-2.19
BIOSINO \ immunoturbidimetric			
Mindray BS series	g/L	2.25	2.1-2.4
MAKER \ immunoturbidimetric			
Beckman AU series	g/L	1.89	1.53-2.25
Beckman Coulter \ immunoturbidimetric			
Roche cobas c 501/502/701/702	g/L	2.02	1.89-2.15
Dedicated \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	μmol/L	2.31	1.78-2.84
Roche \ immunoturbidimetric			
Mindray BS series	g/L	2.17	1.98-2.36
Mindray \ OTHER			
TOSHIBA TBA	g/L	2.2	2.06-2.34
OTHER \ OTHER			
TOSHIBA TBA	g/L	2.25	2.16-2.34
OTHER \ immunoturbidimetric			

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
5 / 20								

(Apo-B)载脂蛋白B			
Roche Hitachi/Hitachi Systems	g/L	1.01	0.97-1.05
Roche \ immunoturbidimetric			
TOSHIBA TBA	g/L	1.22	1.09-1.35
OTHER \ OTHER			
Dirui CS series	g/L	1.08	0.9-1.26
DIRUI \ immunoturbidimetric			
ALIEX CA400	g/L	1.13	1.09-1.17
SHLY \ immunoturbidimetric			
TOSHIBA TBA	g/L	1.18	1-1.36
Autobio \ immunoturbidimetric			
Mindray BS series	g/L	1.02	0.98-1.06
Mindray \ immunoturbidimetric			
Abbott ARCHITECT c series	g/L	1.11	1.03-1.19
BSBE \ immunoturbidimetric			
Beckman AU series	g/L	1.11	1.03-1.19
Reebio \ immunoturbidimetric			
Beckman AU series	g/L	1.02	0.9-1.14
chongqiong zybio \ immunoturbidimetric			
KHB ZY series	g/L	1.01	0.94-1.08
KHB \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	1.11	1.02-1.2
MAKER \ immunoturbidimetric			
Beckman AU series	g/L	1.08	0.93-1.23
BSBE \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	1.02	0.94-1.1
MAKER \ OTHER			
Roche cobas c 501/502/701/702	g/L	0.887	0.662-1.112
DENUO \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	g/L	0.961	0.901-1.021
BIOSINO \ immunoturbidimetric			
Mindray BS series	g/L	1.06	1.02-1.1
MAKER \ immunoturbidimetric			
Beckman AU series	g/L	1.03	0.77-1.29
Beckman Coulter \ immunoturbidimetric			
Roche cobas c 501/502/701/702	g/L	0.968	0.927-1.009
Dedicated \ immunoturbidimetric			
Roche Hitachi/Hitachi Systems	μmol/L	1.04	0.95-1.13
Roche \ immunoturbidimetric			
(ASO)抗链球菌溶血素 O 抗体			
Beckman AU series	IU/mL	103	98-108
BSBE \ OTHER			
Roche cobas c 501/502/701/702	IU/mL	230	69-391
DENUO \ Immunoturbidimetry (ITA)			
Roche Hitachi/Hitachi Systems	IU/mL	133	127-139
Roche \ Immunoturbidimetry (ITA)			
Beckman UniCel DxC 700AU	IU/mL	156	156-156
MedicalSystem \ OTHER			
(AST/ SGOT)天门冬氨酸氨基转移酶			
Beckman AU series	U/L	223	214-232
MAKER \ OTHER			
Beckman AU series	U/L	229	218-240
chongqiong zybio \ OTHER			
Beckman AU series	U/L	230	221-239
Dedicated \ UV-rate without P5P			
ALIEX CA400	U/L	231	224-238
SHLY \ OTHER			
TOSHIBA TBA	U/L	218	202-234
Autobio \ OTHER			
Mindray BS series	U/L	243	233-253
chongqiong zybio \ OTHER			
TOSHIBA TBA	U/L	225	208-242
chongqiong zybio \ OTHER			
Roche Hitachi/Hitachi Systems	U/L	225	206-244
MAKER \ OTHER			
TOSHIBA TBA	U/L	219	208-230
KHB \ OTHER			
TOSHIBA TBA	U/L	239	229-249
Mindray \ OTHER			
Abbott ARCHITECT c series	U/L	242	222-262
BSBE \ OTHER			
Beckman AU series	U/L	244	209-279
BSBE \ OTHER			
Beckman AU series	U/L	222	216-228
XJJC \ OTHER			
Beckman AU series	U/L	237	226-248
Beckman Coulter \ OTHER			
Roche cobas c 501/502/701/702	U/L	220	209-231
DENUO \ UV-rate without P5P			
Dirui CS series	U/L	219	198-240
DIRUI \ OTHER			
Mindray BS series	U/L	230	222-238
OTHER \ UV-rate without P5P			
TOSHIBA TBA	U/L	249	228-270
OTHER \ OTHER			
Beckman AU series	U/L	221	207-235
MedicalSystem \ OTHER			
(BUN)尿素氮			
Beckman AU series	mmol/L	17.4	16-18.8
Dedicated \ Urease-UV-rate			
Beckman AU series	mmol/L	17.4	15-19.8
Beckman Coulter \ Urease-UV-rate			
Roche Hitachi/Hitachi Systems	mmol/L	17.2	16.2-18.2
MAKER \ Dehydrogenase			

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
6 / 20								

Mindray BS series	mmol/L	15.9	14.8-17
KangHua(Weifang) \ Dehydrogenase	mmol/L	18.4	17.4-19.4
Beckman AU series	mmol/L	17	15.5-18.5
Beckman Coulter \ Urease colorimetry	mmol/L	18.5	17.3-19.7
Roche cobas c 501/502/701/702	mmol/L	18.6	17.2-20
DENUO \ Urease-UV-rate			
KHB ZY series	mmol/L		
KHB \ Dehydrogenase			
Roche Hitachi/Hitachi Systems			
Roche \ Urease-UV-rate			

(C3)补体 C3

Roche Hitachi/Hitachi Systems	mg/dL	153	146-160
Roche \ Immunoturbidimetry (ITA)			
Abbott ARCHITECT c series	mg/dL	173	155-191
BSBE \ OTHER			
Beckman AU series	mg/dL	157	150-164
BSBE \ OTHER			

(C4)补体 C4

Abbott ARCHITECT c series	mg/dL	28.1	25.3-30.9
BSBE \ OTHER			
Roche Hitachi/Hitachi Systems	mg/dL	24.7	22.1-27.3
Roche \ Immunoturbidimetry (ITA)			
Beckman AU series	mg/dL	23.9	22.3-25.5
BSBE \ OTHER			

(Ca)钙

Beckman AU series	mmol/L	2.8	2.72-2.88
Dedicated \ ISE (direct)			
CORNLEY Vitagas Series	mmol/L	2.1	2.07-2.13
CORNLEY \ ISE (indirect)			
Dirui CS series	mmol/L	2.67	1.72-3.62
DIRUI \ Arsenazo III			
Roche Hitachi/Hitachi Systems	mmol/L	2.87	2.7-3.04
MAKER \ Arsenazo III			
TOSHIBA TBA	mmol/L	2.95	2.89-3.01
Autobio \ Arsenazo III			
Mindray BS series	mmol/L	2.9	2.77-3.03
Mindray \ Arsenazo III			
Abbott ARCHITECT c series	mmol/L	2.88	2.79-2.97
BIOSINO \ Arsenazo III			
Abbott ARCHITECT c series	mmol/L	2.86	2.65-3.07
BSBE \ Arsenazo III			
Beckman AU series	mmol/L	2.96	2.7-3.22
OTHER \ Arsenazo III			
Roche Hitachi/Hitachi Systems	mmol/L	2.92	2.79-3.05
WanTai \ Arsenazo III			
Beckman AU series	mmol/L	2.96	2.79-3.13
Reebio \ Arsenazo III			
Beckman AU series	mmol/L	2.97	2.84-3.1
Beckman Coulter \ Arsenazo III			
Beckman AU series	mmol/L	2.84	2.57-3.11
BSBE \ Arsenazo III			
Roche cobas c 501/502/701/702	mmol/L	2.88	2.77-2.99
DENUO \ ISE (direct)			
Mindray BS series	mmol/L	2.69	2.14-3.24
MAKER \ Arsenazo III			
Roche Hitachi/Hitachi Systems	mmol/L	2.86	2.69-3.03
Hitachi \ Arsenazo III			
Roche cobas c 501/502/701/702	mmol/L	2.83	2.75-2.91
Dedicated \ BAPTA			
Beckman AU series	mmol/L	3.16	2.88-3.44
MedicalSystem \ Arsenazo III			
Abbott ARCHITECT c series	mmol/L	3.01	2.93-3.09
BIOSINO \ OCPC colorimetry			

(CG)甘胆酸

TOSHIBA TBA	mg/L	17.4	16.8-18
MedicalSystem \ OTHER			
Roche cobas 4000/6000/8000	mg/L	18.5	18.1-18.9
MedicalSystem \ OTHER			
Roche cobas c 501/502/701/702	mg/L	19	17.6-20.4
MedicalSystem \ Latex agglutination (LA)			

(CHE)胆碱酯酶

URIT-500B/330/180/55	U/L	7815	7534-8096
Uritest \ OTHER			
Beckman AU series	U/L	7033	6695-7371
MAKER \ Butyrylthiocholine Method			
TOSHIBA TBA	U/L	8004	7364-8644
Autobio \ Butyrylthiocholine Method			
Mindray BS series	U/L	7920	7540-8300
chongqiong zybio \ Butyrylthiocholine Method			
URIT-500B/330/180/55	U/L	7898	7693-8103
Uritest \ Butyrylthiocholine Method			
TOSHIBA TBA	U/L	7726	7278-8174
chongqiong zybio \ OTHER			
Mindray BS series	U/L	7560	7379-7741
Mindray \ Butyrylthiocholine Method			
TOSHIBA TBA	U/L	7867	7537-8197
Mindray \ Butyrylthiocholine Method			
Beckman AU series	U/L	7697	7266-8128
Beckman Coulter \ Butyrylthiocholine Method			
Roche Hitachi/Hitachi Systems	U/L	6380	6035-6725
MAKER \ Butyrylthiocholine Method			
Roche Hitachi/Hitachi Systems	U/L	7883	7505-8261
MAKER \ OTHER			

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
7 / 20								

Abbott ARCHITECT c series	U/L	8554	8263-8845
BSBE \ Butyrylthiocholine Method			
Roche Hitachi/Hitachi Systems	U/L	7974	7256-8692
BSBE \ OTHER			
Mindray BS series	U/L	6214	4089-8339
BSBE \ OTHER			
Beckman AU series	U/L	8398	7323-9473
BSBE \ Butyrylthiocholine Method			
Roche cobas c 501/502/701/702	U/L	7894	7468-8320
DENUO \ DTNB			
Beckman AU series	U/L	8232	7820-8644
Autobio \ Butyrylthiocholine Method			
TOSHIBA TBA	U/L	7279	7177-7381
OTHER \ Butyrylthiocholine Method			
Mindray CL-2000i/1000i/6000i series	U/L	5288	5087-5489
chongqiong zybio \ Butyrylthiocholine Method			

(CHOL/ TC)总胆固醇

Beckman AU series	mmol/L	5.42	5.09-5.75
MAKER \ OTHER			
Dirui CS series	mmol/L	5.24	5.04-5.44
DIRUI \ OTHER			
Roche Hitachi/Hitachi Systems	mmol/L	5.55	5.27-5.83
MAKER \ OTHER			
Beckman AU series	mmol/L	5	4.77-5.23
chongqiong zybio \ Cholesterol oxidase			
ALIEX CA400	mmol/L	5.25	5.14-5.36
SHLY \ Cholesterol oxidase			
Mindray BS series	mmol/L	4.62	4.51-4.73
chongqiong zybio \ OTHER			
TOSHIBA TBA	mmol/L	4.99	4.61-5.37
chongqiong zybio \ OTHER			
Roche Hitachi/Hitachi Systems	mmol/L	5.55	5.25-5.85
MAKER \ Cholesterol oxidase			
TOSHIBA TBA	mmol/L	5.32	5.14-5.5
Autobio \ OTHER			
TOSHIBA TBA	mmol/L	5.25	4.93-5.56
Mindray \ Cholesterol oxidase			
Abbott ARCHITECT c series	mmol/L	4.98	4.54-5.42
BSBE \ OTHER			
Beckman AU series	mmol/L	5.34	4.78-5.9
Beckman Coulter \ Cholesterol oxidase			
Roche Hitachi/Hitachi Systems	mmol/L	5.82	5.32-6.32
Roche \ Cholesterol oxidase			
Mindray BS series	mmol/L	5.42	5.17-5.67
BIOBASE \ Cholesterol oxidase			
Beckman AU series	mmol/L	5.17	4.86-5.48
BSBE \ OTHER			
Beckman AU series	mmol/L	5.24	5.03-5.45
Beckman Coulter \ OTHER			
Roche cobas c 501/502/701/702	mmol/L	5.38	4.73-6.03
DENUO \ Cholesterol oxidase			
TOSHIBA TBA	mmol/L	5.16	5.08-5.24
OTHER \ OTHER			
TOSHIBA TBA	mmol/L	5.08	3.72-6.44
Autobio \ Cholesterol oxidase			

(CK)肌酸激酶

Beckman AU series	U/L	682	630-734
chongqiong zybio \ Rate -AMP buffer			
Beckman AU series	U/L	700	659-741
Dedicated \ Rate -AMP buffer			
Roche Hitachi/Hitachi Systems	U/L	695	669-721
Roche \ Urease-UV-rate			
Dirui CS series	U/L	694	651-737
DIRUI \ OTHER			
Mindray BS series	U/L	836	809-863
chongqiong zybio \ Rate -AMP buffer			
Roche Hitachi/Hitachi Systems	U/L	684	646-722
MAKER \ OTHER			
Roche Hitachi/Hitachi Systems	U/L	748	705-791
Roche \ DGKC			
Mindray BS series	U/L	671	656-686
Mindray \ OTHER			
Abbott ARCHITECT c series	U/L	662	639-685
BSBE \ OTHER			
Beckman AU series	U/L	668	636-700
Reebio \ OTHER			
Beckman AU series	U/L	735	698-772
chongqiong zybio \ Rate -other buffers			
Beckman AU series	U/L	686	600-772
BSBE \ OTHER			
Beckman AU series	U/L	647	547-747
Beckman Coulter \ Rate -AMP buffer			
Mindray BS series	U/L	638	550-726
MedicalSystem \ Rate -other buffers			
Roche cobas c 501/502/701/702	U/L	673	627-719
DENUO \ Rate -AMP buffer			
Roche Hitachi/Hitachi Systems	U/L	663	614-712
BIOSINO \ OTHER			
Abbott ARCHITECT c series	U/L	710	676-744
BIOSINO \ DGKC			
Mindray BS series	U/L	658	629-687
OTHER \ Rate -other buffers			
Mindray BS series	U/L	670	645-695
MAKER \ OTHER			

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								8 / 20

(CK-MB mass)肌酸激酶-MB质量			
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	14.8	10.5-19.1
Mindray BS series Mindray \ immunoturbidimetric	ng/mL	21.7	16.6-26.8
Roche Hitachi/Hitachi Systems G-CELL \ OTHER	ng/mL	48	37.2-58.8
Beckman AU series BSBE \ OTHER	ng/mL	14.8	11.2-18.4
(CK-MB)肌酸激酶-MB (活性)			
URIT-500B/330/180/55 Uritest \ OTHER	U/L	53.4	19.1-87.7
Beckman AU series OTHER \ OTHER	U/L	72.6	66.9-78.3
Roche Hitachi/Hitachi Systems Roche \ DGKC	U/L	66.3	57.8-74.8
Mindray BS series chongqiong zybio \ Immune suppression	U/L	71.9	67.7-76.1
URIT-500B/330/180/55 Uritest \ Immune suppression	U/L	51.8	46.2-57.4
Mindray BS series Mindray \ Immune suppression	U/L	61.2	58.4-64
TOSHIBA TBA Autobio \ Immune suppression	U/L	61	42.2-79.8
Roche Hitachi/Hitachi Systems Bio-assay \ OTHER	ng/mL	12.7	11.4-14
Beckman AU series Reebio \ Immune suppression	U/L	77.8	71.4-84.2
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	62.1	55.4-68.8
Mindray BS series MAKER \ OTHER	U/L	47.5	0-97.7
Roche Hitachi/Hitachi Systems MAKER \ Immune suppression	U/L	73.5	65.7-81.3
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	U/L	65.5	47.8-83.2
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ Immune suppression	U/L	18.1	16.7-19.5
J&J Vitros 250/350 Ortho \ OTHER	U/L	106	102-110
Mindray BS series BSBE \ OTHER	ng/mL	16.2	14.2-18.2
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	U/L	85.6	73.4-97.8
KHB ZY series KHB \ Immune suppression	U/L	76	70.5-81.5
Beckman AU series BSBE \ OTHER	ng/mL	13.7	11.2-16.2
(Cl)氯			
Beckman AU series Dedicated \ ISE (direct)	mmol/L	101	97-105
Roche Hitachi/Hitachi Systems MAKER \ ISE (indirect)	mmol/L	107	100-114
Dirui CS series DIRUI \ ISE (indirect)	mmol/L	103	101-105
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	102	99-105
Mindray BS series Mindray \ ISE (indirect)	mmol/L	102	93-111
TOSHIBA TBA CANON MEDICAL \ ISE (direct)	mmol/L	103	102-104
Abbott ARCHITECT c series Abbott \ Coulometric titration	mmol/L	106	104-108
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	99.9	95.7-104.1
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	108	104-112
Beckman AU series OTHER \ ISE (direct)	mmol/L	97.9	70.5-125.3
Roche Hitachi/Hitachi Systems Hitachi \ OTHER	mmol/L	98.4	93.9-102.9
Mindray BS series Hitachi \ ISE (indirect)	mmol/L	95.4	76.9-113.9
Beckman AU series BSBE \ ISE (direct)	mmol/L	102	98-106
Roche cobas c 501/502/701/702 DENUO \ ISE (direct)	mmol/L	99	91.9-106.1
Roche Hitachi/Hitachi Systems Roche \ ISE (direct)	mmol/L	111	84-138
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	97.5	95-100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	103	96-110
Mindray BS series KangHua(Weifang) \ ISE (direct)	mmol/L	117	116-118
CORNLEY Vitagas Series CORNLEY \ ISE (indirect)	mmol/L	111	104-118
(CO2 (HCO3-))二氧化碳 (碳酸氢根)			
AB Sciex API Series ABX \ Colorimetry	mmol/L	8.72	7.62-9.82
Dirui CS series DIRUI \ OTHER	mmol/L	5.07	1.47-8.67
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	5.31	4.66-5.96

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
9 / 20								

Roche Hitachi/Hitachi Systems Hitachi \ OTHER	mmol/L	4.53	3.41-5.65
Beckman AU series BSBE \ OTHER	mmol/L	5.16	3.65-6.67
Beckman AU series Beckman Coulter \ Colorimetry	mmol/L	5.78	5.01-6.55
Beckman AU series Beckman Coulter \ Electrode	mmol/L	6.21	4.99-7.43
(Cr)肌酐			
Beckman AU series Dedicated \ Enzymatic	μmol/L	464	443-485
Dirui CS series DIRUI \ Enzymatic	μmol/L	459	443-475
Mindray BS series chongqiong zybio \ Enzymatic	μmol/L	469	459-479
Mindray BS series Mindray \ OTHER	μmol/L	470	461-479
Abbott ARCHITECT c series BIOSINO \ dry chemistry	μmol/L	488	480-496
Abbott ARCHITECT c series BSBE \ OTHER	μmol/L	477	457-497
Beckman AU series BSBE \ Enzymatic	μmol/L	456	422-490
Roche Hitachi/Hitachi Systems Autobio \ Enzymatic	μmol/L	77537	0-410481
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	5490	0-28350
Mindray BS series BIOBASE \ Enzymatic	μmol/L	456	428-484
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	480	460-500
Beckman AU series Beckman Coulter \ Picric acid-No blank Compensation	μmol/L	434	409-459
Roche Hitachi/Hitachi Systems Whitman \ Enzymatic	μmol/L	484	457-511
Beckman AU series BSBE \ OTHER	μmol/L	466	445-487
Mindray BS series Whitman \ Enzymatic	μmol/L	345	0-714
Beckman AU series Reebio \ OTHER	μmol/L	438	421-455
TOSHIBA TBA OTHER \ Enzymatic	μmol/L	449	428-470
Roche Hitachi/Hitachi Systems Roche \ Picric acid-No blank Compensation	μmol/L	17647	0-51247
Roche cobas c 501/502/701/702 DENUO \ Enzymatic	μmol/L	444	396-492
(CRP)C反应蛋白			
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	mg/dL	2638	1873-3403
(Cu)铜			
Beckman AU series Beckman Coulter \ Flame photometry	μmol/L	18.3	14.6-22
Beckman AU series MedicalSystem \ PAESA	μmol/L	13.9	13.5-14.3
Beckman UniCel Dx C 700AU MedicalSystem \ PAESA	μmol/L	13	13-13
(Cys C)半胱氨酸蛋白酶抑制剂C/ 胱抑素C			
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mg/L	2.81	1.66-3.96
Beckman AU series Autobio \ OTHER	mg/L	0.12	0.054-0.186
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	mg/L	0.14	0.091-0.189
Beckman AU series OTHER \ OTHER	mg/L	1.79	0-5.57
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	mg/L	1.48	1.2-1.76
Beckman UniCel Dx C 700AU MedicalSystem \ OTHER	mg/L	0.21	0.21-0.21
(DBIL)直接胆红素			
URIT-500B/330/180/55 Uritest \ OTHER	μmol/L	44.3	42.9-45.7
Beckman AU series MAKER \ OTHER	μmol/L	44.1	41-47.2
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	34.9	24.6-45.2
Beckman AU series Dedicated \ Vanadate	μmol/L	40.9	36.5-45.3
Beckman AU series Beckman Coulter \ Diazotization -No blank Compensation	μmol/L	29.9	28.2-31.6
TOSHIBA TBA KHB \ Vanadate	μmol/L	36.5	33.6-39.4
TOSHIBA TBA Autobio \ OTHER	μmol/L	39.6	33.7-45.5
Dirui CS series DIRUI \ OTHER	μmol/L	32.5	27.4-37.6
TOSHIBA TBA OTHER \ Vanadate	μmol/L	39.6	38-41.2
Mindray BS series chongqiong zybio \ Vanadate	μmol/L	44.4	42.9-45.9
URIT-500B/330/180/55 Uritest \ Diazotization	μmol/L	34.2	32.4-36

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								10 / 20

Beckman AU series Reebio \ OTHER	μmol/L	38.7	31-46.4
Mindray BS series Mindray \ Vanadate	μmol/L	38.9	38.4-39.4
TOSHIBA TBA Mindray \ Vanadate	μmol/L	40	38.9-41.1
Abbott ARCHITECT c series HuaChenBio \ Diazotization	μmol/L	38.6	36.8-40.4
Abbott ARCHITECT c series BSBE \ Vanadate	μmol/L	37.2	30.7-43.7
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	44.9	40-49.8
Beckman AU series XJKC \ Vanadate	μmol/L	40	38.5-41.5
Roche Hitachi/Hitachi Systems Beckman Coulter \ Vanadate	μmol/L	26.7	25.3-28.1
(E2)雌二醇			
Keysmile Smart series KangHua(Weifang) \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	775	772-778
(Fe)铁			
Beckman AU series BSBE \ OTHER	μmol/L	28.2	26.4-30
TOSHIBA TBA XJKC \ Ferrozine colorimetry	μmol/L	33.5	31.9-35.1
Beckman AU series Beckman Coulter \ TPTZ	μmol/L	27.4	25.1-29.7
Mindray BS series Mindray \ Ferrozine colorimetry	μmol/L	30.3	29.5-31.1
Beckman AU series LEADMAN \ Ferrozine colorimetry	μmol/L	29.3	27-31.6
Beckman AU series Beckman Coulter \ OTHER	μmol/L	29.5	28.4-30.6
Beckman AU series Reebio \ Ferrozine colorimetry	μmol/L	31.3	30-32.6
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	28.4	25.4-31.4
Beckman UniCel DxC 700AU MedicalSystem \ Ferrozine colorimetry	μmol/L	25.6	25.6-25.6
(FSH)促卵泡刺激素			
Keysmile Smart series KangHua(Weifang) \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	25.7	24.1-27.3
(FT3)游离T3			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	4.69	4.52-4.86
(FT4)游离T4			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21.6	20.2-23
(GA)糖化血清白蛋白			
Abbott ARCHITECT c series chongqiong zybio \ OTHER	%	18.4	17.8-19
(GGT)L-γ-谷氨酰基转移酶			
Beckman AU series Dedicated \ γ-glutamyl -4-nitroaniline	U/L	195	188-202
ALIEX CA400 SHLY \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	198	193-203
TOSHIBA TBA Mindray \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	194	168-220
Mindray BS series chongqiong zybio \ OTHER	U/L	171	166-176
TOSHIBA TBA chongqiong zybio \ OTHER	U/L	171	162-180
TOSHIBA TBA Autobio \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	188	183-193
Abbott ARCHITECT c series BIOSINO \ dry chemistry	U/L	207	200-214
Roche Hitachi/Hitachi Systems MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	196	181-211
Abbott ARCHITECT c series BSBE \ OTHER	U/L	194	188-200
Beckman AU series OTHER \ OTHER	U/L	191	158-224
Beckman AU series chongqiong zybio \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	178	174-182
Beckman AU series Beckman Coulter \ OTHER	U/L	198	185-211
Beckman AU series MedicalSystem \ OTHER	U/L	180	172-188
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	199	191-207
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	203	175-231
Beckman AU series XJKC \ OTHER	U/L	182	165-199
Mindray BS series OTHER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	189	177-201
Beckman AU series BSBE \ OTHER	U/L	200	185-215
Roche Hitachi/Hitachi Systems Roche \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	183	125-241

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
11 / 20								

(GLDH)谷氨酸脱氢酶			
Beckman AU series DiaSys \ DGKC	U/L	4.56	3.23-5.89
(GLU)葡萄糖			
Beckman AU series Dedicated \ Hexokinase	mmol/L	21.7	20.9-22.5
Roche Hitachi/Hitachi Systems Roche \ Hexokinase	mmol/L	17.2	11.2-23.2
TOSHIBA TBA chongqiong zybio \ Hexokinase	mmol/L	20.9	19.4-22.4
Dirui CS series DIRUI \ Glucose oxidase (GOD)	mmol/L	21.7	20.9-22.5
Mindray BS series chongqiong zybio \ Hexokinase	mmol/L	20.2	19.8-20.6
Mindray BS series Mindray \ Hexokinase	mmol/L	21.4	21.1-21.7
Abbott ARCHITECT c series BIOSINO \ Dehydrogenase	mmol/L	22	21.6-22.4
Roche Hitachi/Hitachi Systems MAKER \ Hexokinase	mmol/L	20.7	17.6-23.8
Abbott ARCHITECT c series BSBE \ Glucose oxidase (GOD)	mmol/L	19.9	16.6-23.2
Beckman AU series OTHER \ Hexokinase	mmol/L	17.8	11.8-23.8
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	19.8	14.7-24.9
Beckman AU series XJJC \ Hexokinase	mmol/L	21.1	20.6-21.6
Mindray BS series BIOBASE \ Glucose oxidase (GOD)	mmol/L	19.7	18.2-21.2
Roche Hitachi/Hitachi Systems MAKER \ Glucose oxidase (GOD)	mmol/L	21.3	20.7-21.9
Roche Hitachi/Hitachi Systems WanTai DRD \ Hexokinase	mmol/L	21.7	20.2-23.2
Beckman AU series BSBE \ Glucose oxidase (GOD)	mmol/L	21.1	20-22.2
Beckman AU series Reebio \ OTHER	mmol/L	22	21.1-22.9
Roche cobas c 501/502/701/702 DENUO \ Hexokinase	mmol/L	19.2	14.3-24.1
Roche Hitachi/Hitachi Systems BIOSINO \ Hexokinase	mmol/L	21.2	20-22.4
(GR)谷胱甘肽还原酶			
Beckman AU series BSBE \ OTHER	U/L	51.7	27.5-75.9
(Hap (HPT))触珠蛋白 (结合珠蛋白)			
Beckman AU series chongqiong zybio \ OTHER	mg/L	1010	966-1054
(HCY)同型半胱氨酸			
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	13	9.5-16.5
Beckman AU series Reebio \ Enzymatic Cycling Assays Method	μmol/L	14	12.6-15.4
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	17.6	15.2-20
TOSHIBA TBA Autobio \ Enzymatic Cycling Assays Method	μmol/L	17.5	15.4-19.6
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	μmol/L	20	19.3-20.7
Abbott ARCHITECT c series BSBE \ Enzymatic Cycling Assays Method	μmol/L	18.6	11.8-25.4
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	18.7	14.8-22.6
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	μmol/L	18	15.3-20.7
Beckman AU series BSBE \ Enzymatic Cycling Assays Method	μmol/L	16.3	13.4-19.2
Mindray BS series MAKER \ Enzymatic Cycling Assays Method	μmol/L	17.4	8.9-25.9
Roche cobas c 501/502/701/702 DENUO \ Enzymatic Cycling Assays Method	μmol/L	14.8	11.1-18.5
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	18.4	16.3-20.5
Beckman AU series Beckman Coulter \ Enzyme immunoassay (EIA)	μmol/L	15.4	11.9-18.9
TOSHIBA TBA WanTai \ OTHER	μmol/L	16.9	15.4-18.4
KHB ZY series KHB \ Enzymatic Cycling Assays Method	μmol/L	16.2	14.4-18
Abbott ARCHITECT c series BJBJ \ Enzymatic Cycling Assays Method	μmol/L	12.4	11.3-13.5
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	μmol/L	15.7	14.3-17.1
URIT 8000 series Uritest \ Enzymatic Cycling Assays Method	μmol/L	22.6	17.8-27.4
TOSHIBA TBA BSBE \ Enzymatic Cycling Assays Method	μmol/L	16.4	15.2-17.6
(HDL-C)高密度脂蛋白胆固醇			
TOSHIBA TBA KHB \ OTHER	mmol/L	1.52	1.44-1.6
Roche Hitachi/Hitachi Systems Roche \ Direct Measure	mmol/L	1.19	0.9-1.48

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
12 / 20								

Beckman AU series MAKER \ OTHER	mmol/L	1.39	1.3-1.48
Dirui CS series DIRUI \ selected suppression	mmol/L	1.23	1.11-1.35
Beckman AU series chongqiong zybïo \ Enzymatic	mmol/L	1.42	1.35-1.49
TOSHIBA TBA Autobio \ Direct Measure	mmol/L	1.23	1.07-1.39
ALIEX CA400 SHLY \ Direct Measure	mmol/L	1.13	1.06-1.2
Mindray BS series Mindray \ Direct Measure	mmol/L	1.12	1.08-1.16
Mindray BS series chongqiong zybïo \ Enzymatic	mmol/L	1.55	1.47-1.63
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	1.31	1.04-1.58
TOSHIBA TBA Mindray \ Direct Measure	mmol/L	1.26	1.21-1.31
Abbott ARCHITECT c series BIOSINO \ Direct Measure	mmol/L	1.6	1.54-1.66
Abbott ARCHITECT c series BSBE \ selected suppression	mmol/L	1.34	1.26-1.42
Beckman AU series MedicalSystem \ Direct Measure	mmol/L	1.3	1.22-1.38
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	1.2	0.93-1.47
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.45	1.17-1.73
Mindray BS series MAKER \ OTHER	mmol/L	0.971	0.519-1.423
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.27	1.19-1.35
Roche cobas c 501/502/701/702 DENUO \ Direct Measure	mmol/L	1.13	0.74-1.52
(H-FABP)心型脂肪酸结合蛋白			
Roche Hitachi/Hitachi Systems chongqiong zybïo \ immunoturbidimetric	ng/mL	16.9	15.2-18.6
(hs-CRP)超敏C反应蛋白			
Beckman AU series MAKER \ OTHER	mg/L	1.85	1.67-2.03
Beckman AU series BSBE \ OTHER	mg/L	1.74	1.41-2.07
Abbott ARCHITECT c series BSBE \ OTHER	mg/L	1.69	1.41-1.97
Beckman AU series Beckman Coulter \ OTHER	mg/L	1.74	1.61-1.87
Mindray BS series DiaSys \ Immunonephelometry (INA)	mg/L	1.96	1.84-2.08
Beckman AU series Reebio \ Immunonephelometry (INA)	mg/L	1.79	1.77-1.81
(IgA)免疫球蛋白 IgA			
TOSHIBA TBA Autobio \ Immunoturbidimetry (ITA)	mg/dL	2.92	2.27-3.57
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	277	241-313
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	215	201-229
Mindray BS series DiaSys \ Immunoturbidimetry (ITA)	mg/dL	227	215-239
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	mg/dL	234	212-256
Beckman AU series BSBE \ OTHER	mg/dL	286	272-300
(IgE)免疫球蛋白 IgE			
Mindray BS series MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	88.9	80.2-97.6
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	95.8	84.3-107.3
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	IU/mL	74.1	61.8-86.4
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	IU/mL	100	95-105
(IgG)免疫球蛋白 IgG			
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	mg/dL	1181	997-1365
TOSHIBA TBA Autobio \ Immunoturbidimetry (ITA)	mg/dL	11.2	10.4-12
Mindray BS series DiaSys \ Immunoturbidimetry (ITA)	mg/dL	1243	1198-1288
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	1166	1042-1290
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	978	915-1041
(IgM)免疫球蛋白 IgM			
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	mg/dL	94.1	86.6-101.6
TOSHIBA TBA Autobio \ Immunoturbidimetry (ITA)	mg/dL	0.983	0.863-1.103
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	103	88-118
Beckman AU series BSBE \ OTHER	mg/dL	121	112-130

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								13 / 20

TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	89.5	83.8-95.2
(IMA)缺血修饰白蛋白			
Roche Hitachi/Hitachi Systems OTHER \ OTHER	IU/mL	84.5	70.5-98.5
Mindray BS series OTHER \ OTHER	IU/mL	94.2	56.5-131.9
Beckman AU series OTHER \ OTHER	IU/mL	63.6	58-69.2
(IV-Col)IV型胶原			
Beckman UniCel DxC 700AU Bio-assay \ immunoturbidimetric	µg/L	37.7	37.7-37.7
(K)钾			
Beckman AU series Dedicated \ ISE (direct)	mmol/L	6.82	6.52-7.12
Dirui CS series DIRUI \ ISE (indirect)	mmol/L	6.79	6.08-7.5
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	6.8	6.53-7.07
TOSHIBA TBA CANON MEDICAL \ ISE (direct)	mmol/L	6.77	6.65-6.89
Abbott ARCHITECT c series Abbott \ dry chemistry	mmol/L	6.82	6.7-6.94
Beckman AU series OTHER \ ISE (direct)	mmol/L	6.31	5.74-6.88
Roche Hitachi/Hitachi Systems Hitachi \ OTHER	mmol/L	6.73	6.03-7.43
Mindray BS series Mindray \ ISE (indirect)	mmol/L	6.92	6.73-7.11
Roche Hitachi/Hitachi Systems Roche \ ISE (direct)	mmol/L	6.56	5.73-7.39
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	6.99	6.75-7.23
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	6.51	5.99-7.03
Beckman AU series BSBE \ ISE (direct)	mmol/L	6.66	6.34-6.98
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	6.91	6.72-7.1
Roche cobas c 501/502/701/702 DENUO \ ISE (direct)	mmol/L	6.71	6.2-7.22
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	6.93	6.74-7.12
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	6.76	6.64-6.88
Mindray BS series Hitachi \ ISE (indirect)	mmol/L	6.84	6.73-6.95
Mindray BS series KangHua(Weifang) \ ISE (direct)	mmol/L	6.93	6.43-7.43
CORNLEY Vitagas Series CORNLEY \ ISE (indirect)	mmol/L	7.01	6.73-7.29
(LA)乳酸			
Beckman AU series Dedicated \ Lactate to pyruvate (L-P)	mmol/L	8.59	8.3-8.88
TOSHIBA TBA DiaSys \ OTHER	mmol/L	6.04	5.16-6.92
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	7.86	7.09-8.63
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	6.56	5.23-7.89
Kangrun Kaeser Series Kangrun \ Lactate to pyruvate (L-P)	mmol/L	6.03	5.79-6.27
(LAP)亮氨酸氨基肽酶			
Roche Hitachi/Hitachi Systems BJLP \ OTHER	U/L	21	19.4-22.6
Beckman AU series BSBE \ OTHER	U/L	25.7	13-38.4
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	26.5	24.2-28.8
(LDH)乳酸脱氢酶			
URIT-500B/330/180/55 Uritest \ OTHER	U/L	470	450-490
Beckman AU series MAKER \ OTHER	U/L	476	453-499
Beckman AU series Dedicated \ Lactate to pyruvate (L-P)	U/L	448	429-467
Beckman AU series MedicalSystem \ OTHER	U/L	464	413-515
Dirui CS series DIRUI \ OTHER	U/L	443	413-473
Mindray BS series chongqiong zyb10 \ OTHER	U/L	453	440-466
URIT-500B/330/180/55 Uritest \ Lactate to pyruvate (L-P)	U/L	460	426-494
Mindray BS series Mindray \ OTHER	U/L	439	429-449
Abbott ARCHITECT c series BIOSINO \ dry chemistry	U/L	460	446-474
Abbott ARCHITECT c series BSBE \ OTHER	U/L	441	414-468
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	478	396-560



REF 581-2 6×5mL EXP 2026-03-29 LOT 376E040 更新日期：2025-0614 / 20			
Beckman AU series chongqiong zybio \ Lactate to pyruvate (L-P)	U/L	459	448-470
Mindray BS series MedicalSystem \ Lactate to pyruvate (L-P)	U/L	452	428-476
Roche Hitachi/Hitachi Systems MAKER \ Lactate to pyruvate (L-P)	U/L	469	458-480
Roche Hitachi/Hitachi Systems Whitman \ Lactate to pyruvate (L-P)	U/L	427	410-444
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	472	447-497
Beckman AU series BSBE \ OTHER	U/L	434	399-469
Roche Hitachi/Hitachi Systems Roche \ Lactate to pyruvate (L-P)	U/L	494	299-689
Beckman AU series Beckman Coulter \ OTHER	U/L	452	425-479
(LDH-I)乳酸脱氢酶同工酶1			
Beckman AU series Reebio \ OTHER	U/L	405	389-421
Roche Hitachi/Hitachi Systems WanTai DRD \ OTHER	U/L	282	263-301
(LDL-C)低密度脂蛋白胆固醇			
Beckman AU series MAKER \ OTHER	mmol/L	3.14	2.77-3.51
Dirui CS series DIRUI \ SUR	mmol/L	3.07	2.67-3.47
TOSHIBA TBA KHB \ SUR	mmol/L	3.11	2.87-3.35
Beckman AU series chongqiong zybio \ Enzymatic	mmol/L	3.64	3.49-3.79
Beckman AU series MedicalSystem \ SUR	mmol/L	2.95	2.31-3.59
TOSHIBA TBA Autobio \ SUR	mmol/L	3.09	2.63-3.55
ALIEX CA400 SHLY \ Direct Measure	mmol/L	3.23	3.07-3.39
Mindray BS series Mindray \ Direct Measure	mmol/L	3.58	3.29-3.87
Beckman AU series Reebio \ OTHER	mmol/L	3.03	2.81-3.25
Mindray BS series chongqiong zybio \ Enzymatic	mmol/L	3.64	3.4-3.88
TOSHIBA TBA Mindray \ Direct Measure	mmol/L	3.55	3.32-3.78
Abbott ARCHITECT c series BSBE \ SUR	mmol/L	2.83	2.56-3.1
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	2.99	2.37-3.61
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	3.15	2.77-3.53
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	3.31	2.69-3.93
Roche Hitachi/Hitachi Systems Roche \ Direct Measure	mmol/L	2.64	2.09-3.19
Roche cobas c 501/502/701/702 DENUO \ Direct Measure	mmol/L	3.03	2.68-3.38
Beckman AU series KHB \ OTHER	mmol/L	2.89	2.77-3.01
Abbott ARCHITECT c series BIOSINO \ SUR	mmol/L	3.05	2.85-3.25
(LH)促黄体生成素			
Keysmile Smart series KangHua(Weifang) \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	28.3	27.4-29.2
(Li)锂			
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	2.17	1.75-2.59
(LIP)脂肪酶			
Beckman AU series Autobio \ OTHER	U/L	96.2	44.6-147.8
Beckman AU series Reebio \ OTHER	U/L	23.4	0-731.3
Roche Hitachi/Hitachi Systems Roche \ Glucose oxidase (GOD)	U/L	88.2	67.4-109
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	103	95-111
TOSHIBA TBA Autobio \ Colorimetry	U/L	134	124-144
Beckman AU series BSBE \ Colorimetry	U/L	67.4	58.4-76.4
Roche Hitachi/Hitachi Systems BSBE \ OTHER	U/L	68.2	58.7-77.7
Mindray BS series BSBE \ OTHER	U/L	51.2	23.6-78.8
Beckman AU series Beckman Coulter \ OTHER	U/L	382	192-572
Abbott ARCHITECT c series BSBE \ OTHER	U/L	264	179-349
Beckman AU series BSBE \ OTHER	U/L	228	197-259
Beckman AU series LEADMAN \ Colorimetry	U/L	115	100-130
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	62.3	54.3-70.3



REF 581-2 6×5mL EXP 2026-03-29 LOT 376E040 更新日期：2025-0615 / 20			
TOSHIBA TBA MedicalSystem \ OTHER	U/L	57.9	54.8-61
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	U/L	58.4	46.7-70.1
(Lp(a))脂蛋白a			
Dirui CS series DIRUI \ OTHER	mg/L	75	41-109
Roche Hitachi/Hitachi Systems MAKER \ Immunonephelometry (INA)	mg/L	113	106-120
Beckman AU series BSBE \ OTHER	mmol/L	0.002	0.002-0.002
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	mmol/L	84.9	64.7-105.1
Mindray BS series chongqiong zybio \ OTHER	mg/L	79.1	76.1-82.1
Beckman AU series Reebio \ OTHER	mg/L	76.8	60.1-93.5
URIT 8000 series Uritest \ Immunoturbidimetry (ITA)	mg/L	98.2	79.1-117.3
TOSHIBA TBA Autobio \ OTHER	mg/L	103	86-120
Roche Hitachi/Hitachi Systems Roche \ Immunonephelometry (INA)	mg/L	114	89-139
Roche Hitachi/Hitachi Systems Roche \ Immunonephelometry (INA)	mmol/L	0	0-0
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mg/L	102	99-105
Roche Hitachi/Hitachi Systems chongqiong zybio \ OTHER	mg/L	322	306-338
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ OTHER	mg/L	43.4	40.9-45.9
TOSHIBA TBA Autobio \ Immunonephelometry (INA)	mg/L	80.3	76.6-84
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	mg/L	105	100-110
Beckman UniCel DxC 700AU MedicalSystem \ OTHER	mmol/L	0	0-0
Beckman UniCel DxC 700AU MedicalSystem \ OTHER	mg/L	94.5	94.5-94.5
(Lp-PLA2)脂蛋白磷脂酶A2			
TOSHIBA TBA MedicalSystem \ OTHER	ng/mL	101	96-106
(Mg)镁			
Mindray BS series Mindray \ Xylidyl Blue	mmol/L	1.55	1.42-1.68
Beckman AU series Dedicated \ Xylidyl Blue	mmol/L	1.41	1.36-1.46
Dirui CS series DIRUI \ Xylidyl Blue	mmol/L	1.51	1.27-1.75
TOSHIBA TBA Autobio \ Xylidyl Blue	mmol/L	1.46	1.41-1.51
Beckman AU series OTHER \ Xylidyl Blue	mmol/L	1.59	1.12-2.06
Beckman AU series Reebio \ Arsenazo □	mmol/L	1.3	1.14-1.46
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	1.39	1.25-1.53
Roche Hitachi/Hitachi Systems MAKER \ Calmagite colorimetry	mmol/L	1.59	1.54-1.64
Beckman AU series BSBE \ Xylidyl Blue	mmol/L	1.49	1.3-1.68
Roche cobas c 501/502/701/702 DENUO \ Calmagite colorimetry	mmol/L	1.42	1.34-1.5
Roche Hitachi/Hitachi Systems Roche \ Xylidyl Blue	mmol/L	1.41	1.25-1.57
Abbott ARCHITECT c series BIOSINO \ Arsenazo □	mmol/L	1.46	1.4-1.52
Mindray BS series MAKER \ Calmagite colorimetry	mmol/L	1.5	1.45-1.55
Roche Hitachi/Hitachi Systems MedicalSystem \ Arsenazo □	mmol/L	1.43	1.3-1.56
Roche cobas c 501/502/701/702 Dedicated \ Calmagite colorimetry	mmol/L	1.38	1.33-1.43
Roche Hitachi/Hitachi Systems MAKER \ Xylidyl Blue	mmol/L	1.39	1.29-1.49
TOSHIBA TBA OTHER \ Xylidyl Blue	mmol/L	1.38	1.22-1.54
TOSHIBA TBA MedicalSystem \ Arsenazo □	mmol/L	1.24	1.07-1.41
Beckman AU series MedicalSystem \ Arsenazo □	mmol/L	1.41	1.05-1.77
(Myo / Mb)肌红蛋白			
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μg/L	242	220-264
Roche Hitachi/Hitachi Systems OTHER \ Immunoturbidimetry (ITA)	μg/L	77.3	63.1-91.5
Mindray BS series MedicalSystem \ Immunoturbidimetry (ITA)	μg/L	39.3	33.9-44.7
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	μg/L	40.9	25.2-56.6
TOSHIBA TBA BSBE \ OTHER	μg/L	60.4	55.3-65.5

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
16 / 20								

(Na)钠			
Beckman AU series Dedicated \ ISE (direct)	mmol/L	143	137-149
Dirui CS series DIRUI \ ISE (indirect)	mmol/L	144	138-150
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	147	141-153
TOSHIBA TBA CANON MEDICAL \ ISE (direct)	mmol/L	147	145-149
Mindray BS series Mindray \ ISE (indirect)	mmol/L	146	128-164
Abbott ARCHITECT c series Abbott \ dry chemistry	mmol/L	146	144-148
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	149	144-154
Beckman AU series OTHER \ ISE (direct)	mmol/L	136	118-154
Roche Hitachi/Hitachi Systems Hitachi \ OTHER	mmol/L	143	138-148
Roche Hitachi/Hitachi Systems Roche \ ISE (direct)	mmol/L	153	143-163
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	145	127-163
Beckman AU series BSBE \ ISE (direct)	mmol/L	145	139-151
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	151	147-155
Roche cobas c 501/502/701/702 DENUO \ ISE (direct)	mmol/L	147	136-158
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	150	146-154
Mindray BS series Hitachi \ ISE (indirect)	mmol/L	143	121-165
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	146	143-149
Mindray BS series KangHua(Weifang) \ ISE (direct)	mmol/L	164	163-165
CORNLEY Vitagas Series CORNLEY \ ISE (indirect)	mmol/L	146	144-148
(NEFA)游离脂肪酸			
Beckman AU series Beckman Coulter \ ACS - ACOD	mmol/L	4.84	4.75-4.93
(P)磷			
Beckman AU series Reebio \ OTHER	mmol/L	2.9	2.78-3.02
Beckman AU series Dedicated \ Phosphomolybdate-UV-end point	mmol/L	2.88	2.75-3.01
Beckman AU series Reebio \ Phosphomolybdate-UV-end point	mmol/L	2.8	2.6-3
Mindray BS series Mindray \ OTHER	mmol/L	2.89	2.79-2.99
Abbott ARCHITECT c series BIOSINO \ dry chemistry	mmol/L	2.9	2.8-3
Beckman AU series BSBE \ Phosphomolybdate-UV-end point	mmol/L	2.64	2.01-3.27
Roche Hitachi/Hitachi Systems MAKER \ Phosphomolybdate-UV-end point	mmol/L	2.99	2.68-3.3
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	2.83	2.46-3.2
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	2.84	2.67-3.01
Beckman AU series Beckman Coulter \ Phosphomolybdate-UV-end point	mmol/L	2.88	2.4-3.36
Beckman AU series BSBE \ OTHER	mmol/L	2.95	2.76-3.14
Beckman AU series Beckman Coulter \ OTHER	mmol/L	2.95	2.84-3.06
Mindray BS series Mindray \ Phosphomolybdate UV-rate	mmol/L	3.12	2.91-3.33
Mindray BS series MAKER \ Phosphomolybdate-UV-end point	mmol/L	2.82	2.77-2.87
Roche Hitachi/Hitachi Systems MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	2.98	2.87-3.09
Roche cobas c 501/502/701/702 Dedicated \ Phosphomolybdate-UV-end point	mmol/L	2.87	2.77-2.97
TOSHIBA TBA MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	2.8	2.65-2.95
TOSHIBA TBA SHLY \ OTHER	mmol/L	2.93	2.85-3.01
Beckman AU series MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	2.77	2.34-3.2
(PAB)前白蛋白			
Roche Hitachi/Hitachi Systems Roche \ Immunoturbidimetry (ITA)	g/L	0.294	0.285-0.303
URIT 8000 series Uritest \ Immunoturbidimetry (ITA)	g/L	0.265	0.24-0.29
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	g/L	0.299	0.283-0.315
Roche Hitachi/Hitachi Systems MAKER \ Immunonephelometry (INA)	g/L	0.324	0.303-0.345
Beckman AU series Reebio \ OTHER	g/L	0.249	0.233-0.265
Roche Hitachi/Hitachi Systems MAKER \ OTHER	g/L	0.27	0.252-0.288

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								17 / 20

Beckman AU series KHB \ OTHER	g/L	0.278	0.266-0.29
Beckman AU series OTHER \ OTHER	g/L	0.269	0.255-0.283
Roche Hitachi/Hitachi Systems BIOSINO \ Immunonephelometry (INA)	g/L	0.259	0.236-0.282
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	g/L	0.299	0.284-0.314
TOSHIBA TBA BSBE \ Immunoturbidimetry (ITA)	g/L	0.296	0.289-0.303
Beckman UniCel Dx C 700AU MedicalSystem \ OTHER	g/L	0.268	0.268-0.268
(PAMY)胰淀粉酶			
Beckman AU series Reebio \ OTHER	U/L	409	390-428
Beckman AU series Beckman Coulter \ Enzymatic	U/L	465	461-469
(PRL)催/ 泌乳素			
Keysmile Smart series KangHua(Weifang) \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	643	631-655
(PRO/ TP)(总) 蛋白			
URIT-500B/330/180/55 Uritest \ OTHER	g/L	68.7	66.5-70.9
TOSHIBA TBA MedicalSystem \ Biuret-endpoint	g/L	70	64.7-75.3
Beckman AU series MAKER \ Biuret-rate	g/L	70.3	67.1-73.5
TOSHIBA TBA Autobio \ Biuret-endpoint	g/L	70.6	66.4-74.8
Beckman AU series Dedicated \ Biuret-endpoint	g/L	64.9	62.4-67.4
ALIE X CA400 SHLY \ Biuret-endpoint	g/L	72.4	71.1-73.7
Dirui CS series DIRUI \ Biuret-rate	g/L	66.3	63-69.6
Mindray BS series chongqiong zybio \ Biuret-rate	g/L	69.1	68-70.2
URIT-500B/330/180/55 Uritest \ Biuret-endpoint	g/L	67.6	62.5-72.7
TOSHIBA TBA chongqiong zybio \ OTHER	g/L	69.2	64.1-74.3
TOSHIBA TBA Mindray \ Biuret-endpoint	g/L	71.4	70-72.8
Roche Hitachi/Hitachi Systems MAKER \ Biuret-endpoint	g/L	72.6	70.6-74.6
Abbott ARCHITECT c series BSBE \ Biuret-endpoint	g/L	72.4	70.8-74
Roche Hitachi/Hitachi Systems MAKER \ Biuret-rate	g/L	71.3	67.2-75.4
TOSHIBA TBA OTHER \ Biuret-endpoint	g/L	68.5	66.9-70.1
Beckman AU series chongqiong zybio \ Biuret-endpoint	g/L	68.4	65.4-71.4
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	67.6	60.8-74.4
Beckman AU series XJKC \ Biuret-endpoint	g/L	69.6	67.5-71.7
Beckman AU series BSBE \ Biuret-endpoint	g/L	71.2	58.8-83.6
(PROG)孕酮			
Keysmile Smart series KangHua(Weifang) \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	10.4	10.2-10.6
(RF)类风湿因子			
Beckman AU series BSBE \ OTHER	IU/mL	10.8	7.9-13.7
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	IU/mL	48	9.5-86.5
Beckman UniCel Dx C 700AU MedicalSystem \ OTHER	IU/mL	6.6	6.6-6.6
(sd LDL-C)小而密低密度脂蛋白胆固醇			
TOSHIBA TBA Reebio \ Peroxidase	mg/L	16.6	15.2-18
(SOD)超氧化物歧化酶			
TOSHIBA TBA OTHER \ OTHER	U/L	46.1	37.4-54.8
(TBA)总胆汁酸			
Beckman AU series Reebio \ Enzymatic Cycling Assays Method	μmol/L	76.4	65.1-87.7
URIT-500B/330/180/55 Uritest \ OTHER	μmol/L	57.8	52.5-63.1
Beckman AU series sysmex \ Enzymatic Cycling Assays Method	μmol/L	53.6	49.5-57.7
Beckman AU series MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	64.8	57.5-72.1
Dirui CS series DIRUI \ Enzymatic Cycling Assays Method	μmol/L	75.8	56.2-95.4
URIT-500B/330/180/55 Uritest \ Enzymatic Cycling Assays Method	μmol/L	56.4	49.3-63.5

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								18 / 20

TOSHIBA TBA	μmol/L	64.1	52.3-75.9
Autobio \ Enzymatic Cycling Assays Method			
Roche Hitachi/Hitachi Systems	μmol/L	76	68.7-83.3
MAKER \ Enzymatic Cycling Assays Method			
TOSHIBA TBA	μmol/L	51.2	47.8-54.6
Mindray \ Enzymatic Cycling Assays Method			
Abbott ARCHITECT c series	μmol/L	52.5	50-55
BSBE \ Enzymatic Cycling Assays Method			
Beckman AU series	μmol/L	54.7	43.7-65.7
XJJC \ Enzymatic Cycling Assays Method			
Roche Hitachi/Hitachi Systems	μmol/L	59.9	53.6-66.2
BSBE \ Enzymatic Cycling Assays Method			
Beckman AU series	μmol/L	55.7	52.6-58.8
BSBE \ Enzymatic Cycling Assays Method			
Mindray BS series	μmol/L	39.9	0-82.3
BSBE \ Enzymatic Cycling Assays Method			
Beckman AU series	μmol/L	63.6	60.8-66.4
KHB \ Enzymatic Cycling Assays Method			
TOSHIBA TBA	μmol/L	60.3	53.4-67.2
chongqiong zybio \ Enzymatic Cycling Assays Method			
Beckman AU series	μmol/L	58.6	39.1-78.1
OTHER \ Enzymatic Cycling Assays Method			
Beckman AU series	μmol/L	53.8	37.6-70
Beckman Coulter \ Enzymatic Cycling Assays Method			
Roche cobas c 501/502/701/702	μmol/L	51.7	38.4-65
DENUO \ Enzymatic Cycling Assays Method			

(TBIL)总胆红素

URIT-500B/330/180/55	μmol/L	132	124-140
Uritest \ OTHER			
Beckman AU series	μmol/L	113	105-121
MAKER \ OTHER			
TOSHIBA TBA	μmol/L	121	115-127
OTHER \ Vanadate			
Beckman AU series	μmol/L	109	100-118
Dedicated \ Vanadate			
Beckman AU series	μmol/L	118	115-121
Beckman Coulter \ Diazotization -No blank Compensation			
TOSHIBA TBA	μmol/L	107	89-125
Autobio \ OTHER			
ALIEX CA400	μmol/L	119	115-123
SHLY \ OTHER			
Dirui CS series	μmol/L	96.5	82.8-110.2
DIRUI \ OTHER			
Mindray BS series	μmol/L	105	101-109
chongqiong zybio \ Vanadate			
URIT-500B/330/180/55	μmol/L	112	105-119
Uritest \ Diazotization			
Beckman AU series	μmol/L	102	89-115
Reebio \ OTHER			
TOSHIBA TBA	μmol/L	113	92-134
KHB \ Vanadate			
TOSHIBA TBA	μmol/L	103	97-109
Mindray \ Vanadate			
Abbott ARCHITECT c series	μmol/L	98.2	96.2-100.2
HuaChenBio \ Diazotization			
Abbott ARCHITECT c series	μmol/L	105	90-120
BSBE \ Vanadate			
Beckman AU series	μmol/L	109	107-111
chongqiong zybio \ Vanadate			
Roche Hitachi/Hitachi Systems	μmol/L	126	101-151
MAKER \ OTHER			
Roche Hitachi/Hitachi Systems	μmol/L	105	75-135
Beckman Coulter \ Vanadate			
TOSHIBA TBA	μmol/L	104	98-110
chongqiong zybio \ OTHER			

(TES)睾酮

Keysmile Smart series	ng/mL	7.03	6.95-7.11
KangHua(Weifang) \ Chemiluminescence Immunoassay (CLIA)			

(TG)甘油三酯

URIT-500B/330/180/55	mmol/L	2.48	2.4-2.56
Uritest \ OTHER			
TOSHIBA TBA	mmol/L	2.39	2.29-2.49
KHB \ GPO-POD			
Beckman AU series	mmol/L	2.39	2.21-2.57
MAKER \ OTHER			
Dirui CS series	mmol/L	2.23	2.1-2.36
DIRUI \ OTHER			
ALIEX CA400	mmol/L	1.99	1.9-2.08
SHLY \ OTHER			
Mindray BS series	mmol/L	2.3	2.23-2.37
Mindray \ GPO-POD			
Mindray BS series	mmol/L	2.34	2.25-2.43
chongqiong zybio \ OTHER			
URIT-500B/330/180/55	mmol/L	2.46	2.36-2.56
Uritest \ GPO-POD			
TOSHIBA TBA	mmol/L	2.27	2.1-2.44
chongqiong zybio \ OTHER			
Roche Hitachi/Hitachi Systems	mmol/L	2.46	2.22-2.7
MAKER \ OTHER			
Abbott ARCHITECT c series	mmol/L	2.35	2.28-2.42
BIOSINO \ Acetyl - acetone photometry			
Abbott ARCHITECT c series	mmol/L	2.29	2.06-2.52
BSBE \ OTHER			
Beckman AU series	mmol/L	2.3	2.23-2.37
chongqiong zybio \ GPO-POD			

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
19 / 20								

Beckman AU series chongqiong zybio \ GK-PK-LDH	mmol/L	2.35	2.09-2.61
Mindray BS series BIOBASE \ GPO-POD	mmol/L	2.23	2.08-2.38
Beckman AU series BSBE \ OTHER	mmol/L	2.32	2.13-2.51
Beckman AU series Beckman Coulter \ OTHER	mmol/L	2.35	2.25-2.45
Roche cobas c 501/502/701/702 DENUO \ Free glycerol removed	mmol/L	2.36	1.99-2.73
Beckman AU series MedicalSystem \ OTHER	mmol/L	2.46	2.04-2.88
(TIBC)总铁结合力			
Beckman AU series LEADMAN \ Ferrozine colorimetry	μmol/L	57.6	54.6-60.6
Abbott ARCHITECT c series LEADMAN \ Ferene	μmol/L	60.8	55.8-65.8
(Tnl)肌钙蛋白I			
Beckman AU series Reebio \ OTHER	ng/mL	0.405	0.213-0.597
(TRF)转铁蛋白			
Abbott ARCHITECT c series BSBE \ Immunoturbidimetry (ITA)	mg/dL	281	266-296
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	mg/dL	272	251-293
Beckman AU series BSBE \ OTHER	mg/dL	290	270-310
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	281	269-293
Abbott ARCHITECT c series chongqiong zybio \ Immunoturbidimetry (ITA)	mg/dL	268	262-274
Abbott ARCHITECT c series chongqiong zybio \ OTHER	mg/dL	262	257-267
TOSHIBA TBA XJKC \ Immunoturbidimetry (ITA)	mg/dL	263	255-271
(TSH)促甲状腺素			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	1.97	1.84-2.1
(TT3)总三碘甲状原氨酸			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.89	1.78-2
(TT4)总T4			
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	82.7	81.2-84.2
(UA)尿酸			
Beckman AU series Reebio \ Uricase colorimetry	μmol/L	465	452-478
Beckman AU series Dedicated \ Uricase-UV	μmol/L	461	433-489
Dirui CS series DIRUI \ OTHER	μmol/L	458	443-473
Mindray BS series chongqiong zybio \ OTHER	μmol/L	476	465-487
TOSHIBA TBA chongqiong zybio \ OTHER	μmol/L	453	419-487
Mindray BS series Mindray \ OTHER	μmol/L	460	454-466
Abbott ARCHITECT c series BIOSINO \ dry chemistry	μmol/L	489	480-498
Abbott ARCHITECT c series BSBE \ OTHER	μmol/L	454	429-479
Beckman AU series OTHER \ OTHER	μmol/L	501	390-612
Roche Hitachi/Hitachi Systems MAKER \ Uricase colorimetry	μmol/L	443	414-472
Roche Hitachi/Hitachi Systems Wako \ OTHER	μmol/L	473	421-525
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	452	330-574
Beckman AU series Reebio \ OTHER	μmol/L	470	454-486
Beckman AU series chongqiong zybio \ Uricase colorimetry	μmol/L	452	427-477
Mindray BS series BIOBASE \ Uricase colorimetry	μmol/L	453	430-476
Roche Hitachi/Hitachi Systems Whitman \ Uricase colorimetry	μmol/L	449	398-500
Beckman AU series BSBE \ OTHER	μmol/L	464	402-526
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	474	460-488
Mindray BS series Whitman \ Uricase colorimetry	μmol/L	376	99-653
(UIBC)不饱和和铁结合力			
Beckman AU series LEADMAN \ OTHER	μmol/L	53.9	48.1-59.7
Beckman AU series LEADMAN \ Ferene	μmol/L	30.2	22-38.4
TOSHIBA TBA XJKC \ Ferrozine colorimetry	μmol/L	38.8	30.3-47.3

REF	581-2	6×5mL	EXP	2026-03-29	LOT	376E040	更新日期：	2025-06
								20 / 20

(Urea)尿素			
URIT-500B/330/180/55 Uritest \ OTHER	mmol/L	16.2	15-17.4
Roche cobas c 501/502/701/702 DENUO \ Urease-UV-rate	mmol/L	17.2	14.3-20.1
Beckman AU series chongqiong zybio \ OTHER	mmol/L	17.6	16.8-18.4
Dirui CS series DIRUI \ OTHER	mmol/L	16.4	14.3-18.5
TOSHIBA TBA Autobio \ Urease electrode	mmol/L	16.9	13.5-20.3
Mindray BS series chongqiong zybio \ OTHER	mmol/L	16.4	16-16.8
URIT-500B/330/180/55 Uritest \ Diacetyl monoxime	mmol/L	15.8	14.1-17.5
TOSHIBA TBA chongqiong zybio \ OTHER	mmol/L	16.5	15.5-17.5
TOSHIBA TBA Autobio \ OTHER	mmol/L	18.4	17.1-19.7
Mindray BS series Mindray \ OTHER	mmol/L	18	17.4-18.6
Abbott ARCHITECT c series BIOSINO \ Diacetyl monoxime	mmol/L	17.4	16.7-18.1
Abbott ARCHITECT c series BSBE \ OTHER	mmol/L	17.5	15.6-19.4
Beckman AU series Reebio \ OTHER	mmol/L	17.9	16.8-19
Beckman AU series OTHER \ OTHER	mmol/L	17.1	14.3-19.9
Beckman AU series chongqiong zybio \ Urease colorimetry	mmol/L	17.7	16.4-19
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	17.7	16.6-18.8
Roche Hitachi/Hitachi Systems Whitman \ Urease-UV-rate	mmol/L	17	14.6-19.4
Beckman AU series BSBE \ OTHER	mmol/L	17.8	16.2-19.4
Beckman AU series Beckman Coulter \ Urease colorimetry	mmol/L	17.4	15-19.8
(Zn)锌			
Beckman AU series WanTai DRD \ Colorimetry	μmol/L	30.4	28.9-31.9
Beckman UniCel DxC 700AU MedicalSystem \ OTHER	μmol/L	31.1	31.1-31.1
(α1M)α1微球蛋白			
Roche Hitachi/Hitachi Systems chongqiong zybio \ immunoturbidimetric	mg/L	141	139-143
(β2M)β-2-微球蛋白			
Beckman AU series Autobio \ OTHER	ng/mL	1091	943-1239
Mindray BS series chongqiong zybio \ OTHER	ng/mL	1728	1624-1832
URIT 8000 series Uritest \ Immunoturbidimetry (ITA)	ng/mL	1424	1205-1643
Beckman AU series OTHER \ OTHER	ng/mL	971	522-1420
URIT-500B/330/180/55 Uritest \ OTHER	ng/mL	1987	1884-2090
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	ng/mL	890	797-983
Beckman AU series Reebio \ OTHER	ng/mL	1280	1144-1416
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ OTHER	ng/mL	1129	1097-1161

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。