

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
1 / 16								

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(ACE)血管紧张素转化酶					
Beckman AU series BSBE \ OTHER	U/L	24.7	15.4%	<5	<100
Beckman AU series Iprocom \ Enzymatic	U/L	22	10%	<5	<100
(a-HBDH)α-羟丁酸脱氢酶					
URIT-500B/330/180/55 Uritest \ OTHER	U/L	146	3.8%	9	276
Beckman AU series LEADMAN \ OTHER	U/L	140	2.3%	9	269
Roche Hitachi/Hitachi Systems Reebio \ Enzymatic Cycling Assays Method	U/L	148	4.3%	9	267
Mindray BS series MAKER \ Enzymatic Cycling Assays Method	U/L	147	2.9%	9	255
Uritest 8000 series Uritest \ OTHER	U/L	129	6.6%	9	243
Mindray BS series BSBE \ OTHER	U/L	145	5.5%	9	237
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	U/L	147	2.9%	9	215
OTHER SHLY \ DGKC	U/L	154	8.5%	7	199
Beckman AU series KeFang(Guangzhou) \ OTHER	U/L	168	2.6%	7	171
Beckman AU series Reebio \ Enzymatic Cycling Assays Method	U/L	151	2%	6	106
Beckman AU series Dedicated \ immunoturbidimetric	U/L	122	2.7%	5	228
Beckman AU series MedicalSystem \ OTHER	U/L	163	5.5%	5	108
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	143	4.7%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ DGKC	U/L	146	2.7%	<5	<100
TOSHIBA TBA Autobio \ OTHER	U/L	146	3.3%	<5	111
Beckman AU series OTHER \ OTHER	U/L	144	2.1%	<5	106
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	148	1.8%	<5	<100
DiruiCS series DIRUI \ OTHER	U/L	148	4.7%	<5	<100
Canon TBA Series chongqiong zybio \ DGKC	U/L	138	2.4%	<5	<100
(Alb)白蛋白					
Roche Hitachi/Hitachi Systems BIOSINO \ Bromocresol green	g/L	27	2.7%	9	282
Beckman AU series LEADMAN \ Bromocresol green	g/L	27.3	1.1%	9	270
Mindray BS series MAKER \ Bromocresol green	g/L	28.3	0.8%	9	254
Uritest 8000 series Uritest \ Bromocresol green	g/L	27.1	0.9%	9	243
Mindray BS series BSBE \ Bromocresol green	g/L	29.1	2.9%	9	240
Roche Hitachi/Hitachi Systems Reebio \ Bromocresol green	g/L	26.4	7.9%	8	246
Roche Hitachi/Hitachi Systems MedicalSystem \ Bromocresol green	g/L	28.1	1.2%	7	191
OTHER SHLY \ Bromocresol green	g/L	30.4	1.8%	7	177
Mindray BS series Mindray \ Bromocresol green	g/L	27.7	2.5%	7	139
URIT-500B/330/180/55 Uritest \ Bromocresol green	g/L	28.3	2.4%	6	143
Beckman AU series MedicalSystem \ Bromocresol green	g/L	27	5.7%	6	122
TOSHIBA TBA Mindray \ Bromocresol green	g/L	26.5	4.1%	5	138
URIT-500B/330/180/55 Uritest \ OTHER	g/L	29.9	3%	<5	132
Siemens ADVIA 1800/2400 BIOSINO \ Bromocresol green	g/L	27.8	1.6%	36	2015
DiruiCS series DIRUI \ Bromocresol green	g/L	27.8	1.5%	<5	<100
Canon TBA Series chongqiong zybio \ Bromocresol green	g/L	29	2.6%	<5	<100
Beckman AU series BSBE \ Bromocresol green	g/L	28.8	3.5%	<5	<100
Beckman AU series XJKC \ Bromocresol green	g/L	29.2	0.8%	<5	<100
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	26.8	2.1%	29	1036
(ALP)碱性磷酸酶					
Roche Hitachi/Hitachi Systems BIOSINO \ Rate -AMP buffer	U/L	33.3	6.8%	9	281
Beckman AU series LEADMAN \ Rate -AMP buffer	U/L	35.3	4.4%	9	270
Uritest 8000 series Uritest \ Rate -AMP buffer	U/L	29.7	10.3%	9	247



REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
2 / 16								

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series	U/L	36.2	7.4%	9	244
BSBE \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	37.2	3.5%	9	220
MedicalSystem \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	35.7	7.3%	8	243
Reebio \ Rate -AMP buffer					
Beckman AU series	U/L	38.3	10.1%	8	186
MedicalSystem \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	35.8	4.3%	7	202
MAKER \ Rate -AMP buffer					
TOSHIBA TBA	U/L	32.7	5.9%	7	182
OTHER \ Rate -AMP buffer					
OTHER	U/L	27.1	13.1%	6	158
SHLY \ OTHER					
Beckman AU series	U/L	33.2	11.9%	6	153
Beckman Coulter \ Rate -AMP buffer					
URIT-500B/330/180/55	U/L	33	6.6%	6	144
Uritest \ Rate -AMP buffer					
Beckman AU series	U/L	36	3.8%	5	<100
Reebio \ Rate -AMP buffer					
TOSHIBA TBA	U/L	33.4	2.5%	5	142
Mindray \ Rate -AMP buffer					
TOSHIBA TBA	U/L	34.3	7.7%	5	104
MedicalSystem \ Rate -AMP buffer					
Beckman AU series	U/L	35.7	3.7%	<5	<100
MAKER \ Rate -AMP buffer					
URIT-500B/330/180/55	U/L	29.3	6.4%	<5	133
Uritest \ OTHER					
Siemens ADVIA 1800/2400	U/L	36.5	3.2%	36	2001
BIOSINO \ Rate -AMP buffer					
DiruiCS series	U/L	35.3	4.1%	<5	<100
DIRUI \ OTHER					
(ALT/ SGPT)丙氨酸氨基转移酶					
URIT-500B/330/180/55	U/L	40	4%	9	273
Uritest \ OTHER					
Beckman AU series	U/L	35.1	3.1%	9	270
LEADMAN \ OTHER					
Beckman AU series	U/L	35.2	2.4%	9	265
Dedicated \ UV-rate with P5P					
Roche Hitachi/Hitachi Systems	U/L	35.3	11.1%	9	257
BSBE \ OTHER					
Uritest 8000 series	U/L	45	4.3%	9	247
Uritest \ OTHER					
Mindray BS series	U/L	37.6	10.5%	9	221
BSBE \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	38.2	3.4%	9	211
MedicalSystem \ UV-rate without P5P					
Mindray BS series	U/L	33.8	2.8%	8	491
Dedicated \ UV-rate without P5P					
Beckman AU series	U/L	36.5	3.6%	6	167
chongqiong zybio \ OTHER					
OTHER	U/L	34.6	8.8%	6	167
SHLY \ OTHER					
Mindray BS series	U/L	35.8	4.6%	6	103
Mindray \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	37.2	3.5%	5	213
BIOSINO \ UV-rate without P5P					
TOSHIBA TBA	U/L	36.5	2.5%	5	141
Mindray \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	36.1	4.4%	<5	<100
MAKER \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	38.4	3.9%	<5	136
BIOSINO \ OTHER					
DiruiCS series	U/L	35.7	3.7%	<5	<100
DIRUI \ OTHER					
Canon TBA Series	U/L	29.5	5.2%	<5	<100
chongqiong zybio \ UV-rate without P5P					
Beckman AU series	U/L	34.7	5.6%	<5	<100
BSBE \ OTHER					
Beckman AU series	U/L	32.4	6%	<5	<100
KHB \ OTHER					
(AMY)淀粉酶					
Beckman AU series	U/L	67	2.3%	9	483
Beckman Coulter \ OTHER					
Beckman AU series	U/L	63	2.1%	9	269
LEADMAN \ PNP-G7					
Mindray BS series	U/L	65.7	2.4%	9	253
MAKER \ PNP-G7					
Beckman AU series	U/L	63.6	1.5%	9	252
Dedicated \ Iodine colorimetry					
Roche Hitachi/Hitachi Systems	U/L	67.3	11.9%	9	248
Reebio \ PNP-G7					
Mindray BS series	U/L	65.2	4.8%	9	239
BSBE \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	62	3.4%	9	219
MedicalSystem \ PNP-G7					
Mindray BS series	U/L	56.1	1.9%	8	491
Dedicated \ Iodine colorimetry					
TOSHIBA TBA	U/L	65.4	1.6%	8	216
OTHER \ OTHER					
Beckman AU series	U/L	63.2	2.1%	7	<100
Reebio \ OTHER					

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
3 / 16								

批号：376B8D4					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
OTHER	U/L	56.7	9%	7	200
SHLY \ OTHER					
Beckman AU series	U/L	58.3	2.4%	5	128
Reebio \ PNP-G7					
Roche Hitachi/Hitachi Systems	U/L	57.2	1.5%	<5	<100
Wako \ CNP_G3					
Beckman AU series	U/L	66.5	1.9%	<5	106
Dedicated \ PNP-G7					
DiruiCS series	U/L	65.1	3.1%	<5	<100
DIRUI \ OTHER					
TOSHIBA TBA	U/L	63	3.2%	<5	<100
WanTai \ PNP-G7					
Canon TBA Series	U/L	60.7	3.4%	<5	<100
chongqiong zybio \ PNP-G7					
Roche Hitachi/Hitachi Systems	U/L	64.6	3%	<5	<100
WanTai DRD \ PNP-G7					
Beckman AU series	U/L	62.6	1.7%	14	391
OTHER \ OTHER					
(Apo-A I)载脂蛋白A I					
Beckman AU series	g/L	0.921	2.8%	9	269
LEADMAN \ immunoturbidimetric					
Beckman AU series	g/L	0.894	2.1%	9	253
Dedicated \ immunoturbidimetric					
Uritest 8000 series	g/L	1.27	1.8%	9	244
Uritest \ immunoturbidimetric					
TOSHIBA TBA	g/L	0.9	1.5%	8	233
OTHER \ immunoturbidimetric					
OTHER	g/L	0.734	4.9%	7	199
SHLY \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.978	1.5%	7	187
MedicalSystem \ immunoturbidimetric					
DiruiCS series	g/L	0.989	5%	5	152
DIRUI \ immunoturbidimetric					
Mindray BS series	g/L	0.947	1.7%	<5	<100
Mindray \ OTHER					
Beckman AU series	g/L	0.928	1.8%	<5	<100
MAKER \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.848	4.6%	<5	<100
BIOSINO \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.921	1.3%	<5	199
MAKER \ OTHER					
Siemens ADVIA 1800/2400	g/L	1	2.6%	36	1997
BIOSINO \ immunoturbidimetric					
TOSHIBA TBA	g/L	1.03	2.2%	<5	<100
OTHER \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	0.913	3%	<5	<100
MAKER \ immunoturbidimetric					
Beckman AU series	g/L	0.864	11.2%	21	668
Reebio \ immunoturbidimetric					
Beckman AU series	g/L	0.941	1.1%	<5	<100
chongqiong zybio \ immunoturbidimetric					
Mindray BS series	g/L	0.984	11.7%	<5	<100
Mindray \ immunoturbidimetric					
Beckman AU series	g/L	0.923	4.2%	14	388
KINGSBIO \ OTHER					
Beckman AU series	g/L	0.897	2.9%	10	456
DongOu \ immunoturbidimetric					
(Apo-B)载脂蛋白B					
Beckman AU series	g/L	0.45	6.8%	9	270
LEADMAN \ immunoturbidimetric					
Beckman AU series	g/L	0.463	2.3%	9	253
Dedicated \ immunoturbidimetric					
Uritest 8000 series	g/L	0.623	5.3%	9	247
Uritest \ immunoturbidimetric					
TOSHIBA TBA	g/L	0.433	2.2%	8	227
OTHER \ immunoturbidimetric					
OTHER	g/L	0.451	10.2%	7	199
SHLY \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.515	1.8%	7	188
MedicalSystem \ immunoturbidimetric					
DiruiCS series	g/L	0.444	8.2%	5	152
DIRUI \ immunoturbidimetric					
Mindray BS series	g/L	0.401	1.9%	<5	<100
Mindray \ OTHER					
Beckman AU series	g/L	0.43	3.1%	<5	<100
MAKER \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.456	4.1%	<5	<100
BIOSINO \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.47	1.7%	<5	194
MAKER \ OTHER					
Siemens ADVIA 1800/2400	g/L	0.506	3.8%	36	2001
BIOSINO \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.575	2.2%	<5	<100
MAKER \ immunoturbidimetric					
Beckman AU series	g/L	0.446	7.2%	<5	<100
BSBE \ immunoturbidimetric					
TOSHIBA TBA	g/L	0.416	19.3%	<5	<100
OTHER \ OTHER					
Beckman AU series	g/L	0.428	4.9%	21	682
Reebio \ immunoturbidimetric					
Beckman AU series	g/L	0.433	6.5%	<5	<100
chongqiong zybio \ immunoturbidimetric					

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
4 / 16								

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series Mindray \ immunoturbidimetric	g/L	0.383	4.6%	<5	<100
Beckman AU series KINGSBIO \ OTHER	g/L	0.384	2.7%	14	392
(AST/ SGOT)天门冬氨酸氨基转移酶					
URIT-500B/330/180/55 Uritest \ OTHER	U/L	40.5	4.6%	9	273
Beckman AU series LEADMAN \ OTHER	U/L	40.7	4.1%	9	270
Beckman AU series Dedicated \ UV-rate with P5P	U/L	38.1	2.4%	9	262
Mindray BS series MAKER \ OTHER	U/L	38.9	3.1%	9	256
Roche Hitachi/Hitachi Systems BSBE \ UV-rate without P5P	U/L	44.3	3.6%	9	250
Uritest 8000 series Uritest \ OTHER	U/L	43.5	3.2%	9	246
Mindray BS series BSBE \ UV-rate without P5P	U/L	43.6	3.9%	9	245
Roche Hitachi/Hitachi Systems MedicalSystem \ UV-rate without P5P	U/L	42.2	3.3%	9	213
Mindray BS series Dedicated \ UV-rate without P5P	U/L	37.8	2.9%	8	493
Beckman AU series chongqiong zybio \ OTHER	U/L	39.1	4.7%	6	176
OTHER SHLY \ OTHER	U/L	34	6.2%	6	166
Mindray BS series Mindray \ OTHER	U/L	37.8	3.6%	6	103
Roche Hitachi/Hitachi Systems BIOSINO \ UV-rate with P5P	U/L	39.6	3.4%	5	215
TOSHIBA TBA Mindray \ OTHER	U/L	40.9	2%	5	140
Roche Hitachi/Hitachi Systems MAKER \ UV-rate without P5P	U/L	38.2	3.3%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	38.1	5.5%	<5	137
Siemens ADVIA 1800/2400 BIOSINO \ UV-rate without P5P	U/L	38.8	4.5%	36	2018
DiruiCS series DIRUI \ OTHER	U/L	37.2	3.9%	<5	<100
Canon TBA Series chongqiong zybio \ UV-rate without P5P	U/L	39.9	5%	<5	<100
(Ca)钙					
Beckman AU series LEADMAN \ Arsenazo III	mmol/L	1.83	1.7%	9	268
Mindray BS series MAKER \ Arsenazo III	mmol/L	1.87	2.5%	9	255
Mindray BS series BSBE \ ISE (direct)	mmol/L	1.84	3.6%	9	239
Roche Hitachi/Hitachi Systems MedicalSystem \ Arsenazo III	mmol/L	1.88	1.7%	7	189
Beckman AU series MedicalSystem \ Arsenazo III	mmol/L	1.93	5.7%	7	171
MZKL CORNLEY \ ISE (indirect)	mmol/L	2.01	0.8%	6	<100
Roche Hitachi/Hitachi Systems Health World \ Arsenazo III	mmol/L	1.84	1.7%	<5	<100
Beckman AU series MAKER \ Arsenazo III	mmol/L	1.96	8.6%	<5	<100
Beckman AU series Dedicated \ ISE (direct)	mmol/L	1.77	1.6%	<5	106
DiruiCS series DIRUI \ Arsenazo III	mmol/L	2.06	0.9%	<5	<100
TOSHIBA TBA MedicalSystem \ Arsenazo III	mmol/L	1.91	1.6%	<5	<100
Canon TBA Series chongqiong zybio \ Arsenazo III	mmol/L	2.02	2.8%	<5	<100
Beckman AU series BSBE \ Arsenazo III	mmol/L	1.84	9.9%	<5	<100
Beckman AU series Dedicated \ Arsenazo III	mmol/L	1.85	1.6%	24	873
Beckman AU series Reebio \ Arsenazo III	mmol/L	1.85	3.1%	21	659
Mindray BS series Mindray \ Arsenazo III	mmol/L	1.77	2.2%	<5	<100
Beckman AU series Beckman Coulter \ Arsenazo III	mmol/L	1.84	2.4%	16	442
Beckman AU series DiaSys \ Arsenazo III	mmol/L	1.82	1.6%	13	324
Siemens ADVIA 1800/2400 BIOSINO \ Arsenazo III	mmol/L	1.8	2.2%	12	672
(CHE)胆碱酯酶					
Beckman AU series LEADMAN \ OTHER	U/L	4780	1.3%	9	269
Roche Hitachi/Hitachi Systems BSBE \ Butyrylthiocholine Method	U/L	5327	8.4%	9	269
Mindray BS series MAKER \ Butyrylthiocholine Method	U/L	4804	1.1%	9	257
Beckman AU series Dedicated \ Butyrylthiocholine Method	U/L	4637	1.7%	9	254

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
5 / 16								

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series BSBE \ Butyrylthiocholine Method	U/L	5443	2.9%	9	247
Beckman AU series MAKER \ Butyrylthiocholine Method	U/L	4646	2.4%	9	103
Mindray BS series Dedicated \ DTNB	U/L	5153	2.9%	8	492
Roche Hitachi/Hitachi Systems MedicalSystem \ Butyrylthiocholine Method	U/L	5035	2.2%	7	193
TOSHIBA TBA KHB \ Butyrylthiocholine Method	U/L	4896	0.9%	7	184
URIT-500B/330/180/55 Uritest \ Butyrylthiocholine Method	U/L	5168	1.5%	6	143
TOSHIBA TBA Mindray \ Butyrylthiocholine Method	U/L	5244	2.2%	5	121
Roche Hitachi/Hitachi Systems BIOSINO \ Butyrylthiocholine Method	U/L	5210	4%	<5	<100
URIT-500B/330/180/55 Uritest \ OTHER	U/L	5073	1.7%	<5	133
TOSHIBA TBA OTHER \ OTHER	U/L	5189	1.2%	<5	<100
Canon TBA Series chongqiong zybio \ Butyrylthiocholine Method	U/L	4890	2.7%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ OTHER	U/L	5073	2.2%	<5	<100
Beckman AU series BSBE \ Butyrylthiocholine Method	U/L	5187	7.2%	<5	<100
Beckman AU series Beckman Coulter \ Butyrylthiocholine Method	U/L	4796	3.5%	28	996
Mindray BS series chongqiong zybio \ Butyrylthiocholine Method	U/L	5200	2.1%	<5	<100
(CHOL/ TC)总胆固醇					
Beckman AU series Beckman Coulter \ OTHER	mmol/L	2.29	2.2%	9	488
URIT-500B/330/180/55 Uritest \ Cholesterol oxidase	mmol/L	2.24	4.3%	9	276
Beckman AU series LEADMAN \ Cholesterol oxidase	mmol/L	2.37	1.7%	9	268
Mindray BS series BSBE \ OTHER	mmol/L	2.36	4.5%	9	256
Uritest 8000 series Uritest \ OTHER	mmol/L	2.46	2%	9	247
Roche Hitachi/Hitachi Systems BSBE \ Cholesterol oxidase	mmol/L	2.2	4.5%	8	243
Beckman AU series MedicalSystem \ OTHER	mmol/L	2.33	2.3%	8	184
Roche Hitachi/Hitachi Systems MedicalSystem \ Cholesterol oxidase	mmol/L	2.27	1.4%	7	193
TOSHIBA TBA KHB \ OTHER	mmol/L	2.28	2.2%	7	180
Roche Hitachi/Hitachi Systems MAKER \ Cholesterol oxidase	mmol/L	2.34	2.9%	7	136
OTHER SHLY \ Cholesterol oxidase	mmol/L	2.19	3.6%	6	161
Mindray BS series Mindray \ Cholesterol oxidase	mmol/L	2.27	1.9%	6	103
Beckman AU series chongqiong zybio \ Cholesterol oxidase	mmol/L	2.18	2.6%	5	<100
DiruiCS series DIRUI \ OTHER	mmol/L	2.26	2.1%	5	153
TOSHIBA TBA Mindray \ Cholesterol oxidase	mmol/L	2.36	5.4%	5	141
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	2.2	4.6%	<5	135
Siemens ADVIA 1800/2400 BIOSINO \ Cholesterol oxidase	mmol/L	2.27	2.9%	36	2015
Canon TBA Series chongqiong zybio \ Cholesterol oxidase	mmol/L	2.26	2.3%	<5	<100
TOSHIBA TBA OTHER \ Cholesterol oxidase	mmol/L	2.34	1.9%	<5	<100
(CK)肌酸激酶					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	147	3.3%	9	284
URIT-500B/330/180/55 Uritest \ OTHER	U/L	89.9	4.1%	9	277
Roche Hitachi/Hitachi Systems BSBE \ Rate -AMP buffer	U/L	158	8.8%	9	270
Beckman AU series LEADMAN \ Rate -AMP buffer	U/L	166	4.4%	9	269
Mindray BS series MAKER \ OTHER	U/L	143	1.2%	9	255
Uritest 8000 series Uritest \ OTHER	U/L	95.3	27.4%	9	245
Mindray BS series BSBE \ Rate -AMP buffer	U/L	166	7.9%	9	222
Roche Hitachi/Hitachi Systems MedicalSystem \ Urease-UV-rate	U/L	159	1.2%	9	220
Mindray BS series Dedicated \ Rate -other buffers	U/L	158	1.5%	8	494
TOSHIBA TBA OTHER \ OTHER	U/L	164	1.1%	8	218
OTHER SHLY \ OTHER	U/L	164	2.9%	7	198

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
6 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series KeFang(Guangzhou) \ Rate -other buffers	U/L	164	2.8%	7	164
Beckman AU series chongqiong zybio \ Rate -AMP buffer	U/L	156	2.3%	6	164
TOSHIBA TBA Autobio \ OTHER	U/L	151	4.8%	5	134
Beckman AU series MAKER \ OTHER	U/L	142	2.1%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Rate -AMP buffer	U/L	150	2.2%	<5	<100
DiruiCS series DIRUI \ OTHER	U/L	153	3.1%	<5	<100
Canon TBA Series chongqiong zybio \ Rate -AMP buffer	U/L	165	2.3%	<5	<100
TOSHIBA TBA OTHER \ Rate -other buffers	U/L	159	3.9%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Beckman AU series LEADMAN \ Immune suppression	U/L	20.5	15.2%	9	270
Mindray BS series MAKER \ Enzyme immunoassay (EIA)	U/L	18.6	11.8%	9	255
Beckman AU series Dedicated \ Immune suppression	U/L	18.1	6.2%	9	255
Roche Hitachi/Hitachi Systems Reebio \ DGKC	U/L	20.9	8.7%	9	251
Mindray BS series BSBE \ DGKC	U/L	20	8.6%	9	231
Roche Hitachi/Hitachi Systems MedicalSystem \ Immune suppression	U/L	20.3	8.5%	7	181
Uritest 8000 series Uritest \ Immune suppression	U/L	6.61	17.7%	6	175
URIT-500B/330/180/55 Uritest \ Immune suppression	U/L	11.1	18.7%	6	144
OTHER SHLY \ Immune suppression	U/L	27.1	24.8%	5	142
Roche Hitachi/Hitachi Systems MAKER \ Immune suppression	U/L	17.7	10.3%	<5	203
URIT-500B/330/180/55 Uritest \ OTHER	U/L	14.3	33.6%	<5	127
Beckman AU series OTHER \ OTHER	U/L	22.9	8.4%	<5	106
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	12.7	12.4%	<5	<100
TOSHIBA TBA Autobio \ Immune suppression	U/L	15.4	8.8%	<5	<100
Mindray BS series chongqiong zybio \ Immune suppression	U/L	18	2.7%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Chemiluminescence enzyme immunoassay (CLEIA)	U/L	24.9	4.5%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ DGKC	U/L	7.29	24.9%	16	597
Beckman AU series Reebio \ Immune suppression	U/L	24	8.1%	14	614
Beckman AU series chongqiong zybio \ OTHER	U/L	22.8	6.5%	14	392
(Cl)氯					
Roche Hitachi/Hitachi Systems Reebio \ ISE (indirect)	mmol/L	77.7	8.9%	9	271
Mindray BS series MAKER \ ISE (indirect)	mmol/L	82.9	0.8%	9	257
Mindray BS series BSBE \ ICP-MS	mmol/L	84.3	2.3%	9	236
Mindray BS series Dedicated \ ISE (indirect)	mmol/L	82.3	1.2%	8	507
Beckman AU series Reebio \ ISE (direct)	mmol/L	84.3	1.4%	7	<100
Roche Hitachi/Hitachi Systems Hitachi \ ISE (direct)	mmol/L	78.3	1.8%	7	188
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	76.5	1.6%	6	206
OTHER OTHER \ ISE (direct)	mmol/L	79.6	1.1%	<5	<100
TOSHIBA TBA OTHER \ ISE (indirect)	mmol/L	85.7	1%	<5	<100
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	75.5	1.3%	<5	108
Beckman AU series Dedicated \ ISE (direct)	mmol/L	80.3	1.9%	<5	106
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	81.4	1.1%	37	1246
DiruiCS series DIRUI \ ISE (indirect)	mmol/L	92.4	0.8%	<5	<100
Canon TBA Series chongqiong zybio \ ISE (direct)	mmol/L	79.7	4.2%	<5	<100
Beckman AU series BSBE \ ISE (direct)	mmol/L	80.6	3.6%	<5	<100
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	81.2	0.9%	24	880
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	82.1	1.5%	24	1356
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	79.5	8%	<5	<100

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
7 / 16								

批号：376B8D4					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems MedicalSystem \ ISE (direct)	mmol/L	80.6	1.2%	<5	<100
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Roche Hitachi/Hitachi Systems MedicalSystem \ PEPC	mmol/L	18.1	7.5%	9	162
Beckman AU series HuiLi \ Colorimetry	mmol/L	25.4	2.3%	5	<100
DiruiCS series DIRUI \ OTHER	mmol/L	8.35	4.1%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	9.82	5.8%	<5	<100
TOSHIBA TBA OTHER \ OTHER	mmol/L	4.3	2.3%	<5	<100
Beckman AU series BIOSINO \ PEPC	mmol/L	11.5	10%	<5	<100
Beckman AU series HuiLi \ OTHER	mmol/L	27.1	0.4%	<5	<100
Beckman AU series BSBE \ OTHER	mmol/L	7.05	17.7%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ OTHER	mmol/L	6.96	9.1%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	5.45	7.9%	<5	<100
(Cr)肌酐					
URIT-500B/330/180/55 Uritest \ Enzymatic	μmol/L	68.8	4.8%	9	274
Beckman AU series LEADMAN \ Enzymatic	μmol/L	69.8	2%	9	270
Mindray BS series MAKER \ Enzymatic	μmol/L	68.6	2.4%	9	253
Uritest 8000 series Uritest \ Enzymatic	μmol/L	68.5	2.2%	9	247
Mindray BS series BSBE \ OTHER	μmol/L	72.7	3.1%	9	229
Beckman AU series chongqiong zybjo \ Enzymatic	μmol/L	69	4.7%	9	194
Mindray BS series Dedicated \ Enzymatic	μmol/L	71.8	2%	8	495
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	μmol/L	73.2	3.8%	7	190
Beckman AU series MedicalSystem \ Enzymatic	μmol/L	73.1	8%	6	130
Beckman AU series Dedicated \ OTHER	μmol/L	67.5	4%	5	231
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic	μmol/L	74.3	2.1%	5	215
TOSHIBA TBA Autobio \ Enzymatic	μmol/L	72.2	3.7%	5	142
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	66.3	3.6%	<5	<100
Beckman AU series MAKER \ Enzymatic	μmol/L	69.9	5.6%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic	μmol/L	69.5	3%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μmol/L	72.7	4.9%	<5	136
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic	μmol/L	72.6	3%	36	2005
DiruiCS series DIRUI \ Enzymatic	μmol/L	68.9	2.4%	<5	<100
OTHER SHLY \ Enzymatic	μmol/L	55.3	18.6%	<5	<100
(DBIL)直接胆红素					
Beckman AU series LEADMAN \ Vanadate	μmol/L	7.74	7.5%	9	270
Roche Hitachi/Hitachi Systems BSBE \ Vanadate	μmol/L	9.07	7.5%	9	267
Beckman AU series Dedicated \ sodium nitrite oxidase method	μmol/L	7.94	4.1%	9	258
Uritest 8000 series Uritest \ OTHER	μmol/L	7.82	3.8%	9	247
Mindray BS series BSBE \ Vanadate	μmol/L	9.28	10.9%	9	240
Mindray BS series Dedicated \ Vanadate	μmol/L	7.51	6.4%	8	494
Roche Hitachi/Hitachi Systems MedicalSystem \ sodium nitrite oxidase method	μmol/L	9.25	1.7%	7	188
Beckman AU series MedicalSystem \ sodium nitrite oxidase method	μmol/L	7.44	3.1%	7	152
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	7.39	2.5%	6	347
TOSHIBA TBA OTHER \ Vanadate	μmol/L	6.73	2.9%	6	171
OTHER SHLY \ sodium nitrite oxidase method	μmol/L	7.54	15.5%	6	167
URIT-500B/330/180/55 Uritest \ Diazotization	μmol/L	7.5	3.9%	6	142
Roche Hitachi/Hitachi Systems DiaSys \ Diazotization -Blank Compensation	μmol/L	7.4	5.9%	5	215
Beckman AU series Reebio \ OTHER	μmol/L	7.66	3.6%	5	144

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
								8 / 16

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
TOSHIBA TBA	μmol/L	8	3.1%	5	142
Mindray \ Vanadate					
Beckman AU series	μmol/L	7.29	3.1%	5	128
Beckman Coulter \ Diazotization -No blank Compensation					
Roche Hitachi/Hitachi Systems	μmol/L	7.3	6.2%	<5	<100
BIOSINO \ OTHER					
Roche Hitachi/Hitachi Systems	μmol/L	8.99	3.9%	<5	<100
MAKER \ sodium nitrite oxidase method					
URIT-500B/330/180/55	μmol/L	9.16	3.7%	<5	133
Uritest \ OTHER					
(Fe)铁					
Beckman AU series	μmol/L	17.5	2.2%	9	485
Beckman Coulter \ OTHER					
Beckman AU series	μmol/L	17.2	3.1%	5	228
Dedicated \ Ferrozine colorimetry					
TOSHIBA TBA	μmol/L	16.8	4.2%	<5	<100
XJJC \ Ferrozine colorimetry					
Roche Hitachi/Hitachi Systems	μmol/L	18.7	2.9%	<5	<100
Wako \ Bathophenanthroline					
Siemens ADVIA 1800/2400	μmol/L	16.6	4.4%	36	2008
BIOSINO \ Ferrozine colorimetry					
Beckman AU series	μmol/L	16.7	2.4%	12	229
Reebio \ Ferrozine colorimetry					
Beckman AU series	μmol/L	17	2.4%	11	297
LEADMAN \ Ferrozine colorimetry					
Siemens ADVIA 1800/2400	μmol/L	16.6	1.6%	<5	<100
Siemens \ Ferrozine colorimetry					
Roche cobas 4000/6000/8000	μmol/L	17	10%	<5	<100
Roche \ Ferrozine colorimetry					
Beckman AU series	μmol/L	14.7	10%	<5	<100
MedicalSystem \ Ferrozine colorimetry					
Mindray BS series	μmol/L	16.4	0.4%	<5	<100
Mindray \ Ferrozine colorimetry					
(GGT)L-γ-谷氨酰基转移酶					
URIT-500B/330/180/55	U/L	31.2	5.5%	9	276
Uritest \ OTHER					
Beckman AU series	U/L	31	2.8%	9	270
LEADMAN \ γ-glutamyl -3-carboxy-4-nitroaniline					
Roche Hitachi/Hitachi Systems	U/L	32	4.6%	9	267
BSBE \ γ-glutamyl -3-carboxy-4-nitroaniline					
Mindray BS series	U/L	32.1	6.4%	9	252
BSBE \ OTHER					
Uritest 8000 series	U/L	29	2.2%	9	247
Uritest \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	30.8	2.3%	9	217
MedicalSystem \ γ-glutamyl -3-carboxy-4-nitroaniline					
Mindray BS series	U/L	32.8	2.9%	8	493
Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline					
TOSHIBA TBA	U/L	30.9	2.6%	8	211
OTHER \ OTHER					
Beckman AU series	U/L	32.2	2.4%	7	<100
Reebio \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	31.7	2.1%	7	219
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	30.3	3.8%	7	132
MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline					
OTHER	U/L	26.3	5.6%	6	164
SHLY \ OTHER					
Mindray BS series	U/L	31.8	5.5%	6	104
Mindray \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	32.5	2.1%	5	217
BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline					
TOSHIBA TBA	U/L	33.2	2.1%	5	146
Mindray \ γ-glutamyl -3-carboxy-4-nitroaniline					
Beckman AU series	U/L	33.5	1.5%	<5	<100
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	31.9	4.4%	<5	132
BIOSINO \ OTHER					
Beckman AU series	U/L	33.2	2.8%	<5	106
Dedicated \ γ-glutamyl -4-nitroaniline					
Siemens ADVIA 1800/2400	U/L	31.8	3.3%	36	2009
BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline					
(GLU)葡萄糖					
Beckman AU series	mmol/L	4.19	1.8%	9	485
Reebio \ OTHER					
URIT-500B/330/180/55	mmol/L	4.27	5.6%	9	275
Uritest \ Glucose oxidase (GOD)					
Beckman AU series	mmol/L	4.02	2.7%	9	270
LEADMAN \ Glucose oxidase (GOD)					
Mindray BS series	mmol/L	4	0.9%	9	257
MAKER \ Hexokinase					
Roche Hitachi/Hitachi Systems	mmol/L	3.86	4%	9	256
BSBE \ Hexokinase					
Uritest 8000 series	mmol/L	3.86	2.1%	9	247
Uritest \ Glucose oxidase (GOD)					
Mindray BS series	mmol/L	4.04	3.5%	9	236
BSBE \ Hexokinase					
TOSHIBA TBA	mmol/L	4.41	1.7%	8	237
OTHER \ Glucose oxidase (GOD)					
Beckman AU series	mmol/L	3.78	3.9%	8	168
MedicalSystem \ Glucose oxidase (GOD)					

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
9 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems MAKER \ Hexokinase	mmol/L	4.01	1.6%	8	158
OTHER SHLY \ Glucose oxidase (GOD)	mmol/L	4.06	3.3%	7	200
Roche Hitachi/Hitachi Systems MedicalSystem \ Glucose oxidase (GOD)	mmol/L	4.07	1.2%	7	193
TOSHIBA TBA Autobio \ Hexokinase	mmol/L	4.03	1.9%	5	146
Beckman AU series MAKER \ Hexokinase	mmol/L	3.97	2.1%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Hexokinase	mmol/L	3.96	5.3%	<5	142
Siemens ADVIA 1800/2400 BIOSINO \ Hexokinase	mmol/L	4.13	1.9%	36	2019
DiruiCS series DIRUI \ Glucose oxidase (GOD)	mmol/L	4.05	2%	<5	<100
TOSHIBA TBA OTHER \ Hexokinase	mmol/L	3.99	1.4%	<5	<100
Canon TBA Series chongqiong zybïo \ Hexokinase	mmol/L	4.02	2.2%	<5	<100
(GPDA)甘氨酸脯氨酸二肽氨基肽酶					
Beckman AU series Iprocom \ Enzymatic	U/L	45	10%	<5	<100
(HCY)同型半胱氨酸					
Uritest 8000 series Uritest \ Enzymatic Cycling Assays Method	μmol/L	9.84	14.6%	9	242
Roche Hitachi/Hitachi Systems Reebio \ Enzymatic Cycling Assays Method	μmol/L	10.7	9.2%	9	229
Mindray BS series BSBE \ Enzymatic Cycling Assays Method	μmol/L	9.95	18%	9	222
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	9.84	4.5%	7	<100
Beckman AU series Reebio \ Enzymatic Cycling Assays Method	μmol/L	8.78	6.5%	6	<100
TOSHIBA TBA Autobio \ Enzymatic Cycling Assays Method	μmol/L	10.1	10%	6	157
TOSHIBA TBA WanTai \ OTHER	μmol/L	9.12	3.7%	<5	<100
Beckman AU series MAKER \ Enzymatic Cycling Assays Method	μmol/L	10.8	6.7%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	11.3	5.2%	<5	125
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	μmol/L	8.77	6.7%	<5	<100
Beckman AU series BSBE \ Enzymatic Cycling Assays Method	μmol/L	9.1	7.4%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzyme linked immunosorbent assays (ELISA)	μmol/L	10.1	1.8%	<5	<100
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	μmol/L	8.08	8.8%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	8.51	5.2%	14	613
Beckman AU series OTHER \ OTHER	μmol/L	7.84	7%	14	378
Siemens ADVIA 1800/2400 BIOSINO \ Immunoturbidimetry (ITA)	μmol/L	8.27	5%	13	674
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	8.8	5%	12	675
Beckman AU series DongOu \ OTHER	μmol/L	8.77	3.9%	10	457
Siemens ADVIA 1800/2400 Siemens \ Immunonephelometry (INA)	μmol/L	8.6	5.7%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Beckman AU series Beckman Coulter \ OTHER	mmol/L	0.618	3.1%	9	484
URIT-500B/330/180/55 Uritest \ selected suppression	mmol/L	0.727	9.5%	9	272
Beckman AU series LEADMAN \ Direct Measure	mmol/L	0.712	7.5%	9	269
Beckman AU series Dedicated \ Enzymatic	mmol/L	0.703	2.1%	9	254
Mindray BS series BSBE \ selected suppression	mmol/L	0.697	6.1%	9	250
Uritest 8000 series Uritest \ selected suppression	mmol/L	0.793	3.2%	9	248
Roche Hitachi/Hitachi Systems Reebio \ selected suppression	mmol/L	0.682	2.6%	8	239
Beckman AU series KeFang(Guangzhou) \ Direct Measure	mmol/L	0.624	3.1%	8	156
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	mmol/L	0.79	2.3%	7	193
OTHER SHLY \ Direct Measure	mmol/L	0.692	7.4%	6	170
TOSHIBA TBA OTHER \ selected suppression	mmol/L	0.72	12.3%	6	162
Beckman AU series chongqiong zybïo \ Enzymatic	mmol/L	0.723	2.8%	5	<100
DiruiCS series DIRUI \ selected suppression	mmol/L	0.641	5.6%	5	152
TOSHIBA TBA Mindray \ Direct Measure	mmol/L	0.703	2.4%	5	140



REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
10 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
TOSHIBA TBA	mmol/L	0.708	2.6%	5	139
KHB \ OTHER					
Beckman AU series	mmol/L	0.715	5%	<5	<100
MedicalSystem \ Direct Measure					
Mindray BS series	mmol/L	0.724	1.9%	<5	<100
Mindray \ Direct Measure					
Roche Hitachi/Hitachi Systems	mmol/L	0.749	1.8%	<5	<100
MAKER \ Direct Measure					
Siemens ADVIA 1800/2400	mmol/L	0.816	3%	36	2015
BIOSINO \ Direct Measure					
(K)钾					
Roche Hitachi/Hitachi Systems	mmol/L	2.99	6.3%	9	274
BSBE \ ISE (indirect)					
Mindray BS series	mmol/L	2.92	0.8%	9	257
MAKER \ ISE (indirect)					
Mindray BS series	mmol/L	2.94	1.5%	8	501
Dedicated \ ISE (indirect)					
Beckman AU series	mmol/L	2.85	1.6%	7	<100
Reebio \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	2.9	1.9%	7	191
Hitachi \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	2.95	1.4%	5	215
Dedicated \ ICP-MS					
OTHER	mmol/L	2.77	1.4%	<5	<100
OTHER \ ISE (direct)					
TOSHIBA TBA	mmol/L	2.84	1.6%	<5	<100
OTHER \ ISE (indirect)					
Beckman AU series	mmol/L	2.89	2.2%	<5	108
Dedicated \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	2.8	1.2%	<5	105
OTHER \ OTHER					
Beckman AU series	mmol/L	2.95	3.8%	37	1277
Beckman Coulter \ ISE (indirect)					
DiruiCS series	mmol/L	3.17	1.3%	<5	<100
DIRUI \ ISE (indirect)					
Canon TBA Series	mmol/L	2.82	3.7%	<5	<100
chongqiong zybio \ ISE (direct)					
Beckman AU series	mmol/L	2.85	3.5%	<5	<100
BSBE \ ISE (direct)					
Beckman AU series	mmol/L	2.87	1.2%	24	876
Dedicated \ ISE (indirect)					
Siemens ADVIA 1800/2400	mmol/L	2.94	1.8%	24	1342
Dedicated \ ISE (indirect)					
Beckman AU series	mmol/L	3.32	18.9%	<5	<100
Beckman Coulter \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	2.91	1.7%	<5	<100
MedicalSystem \ ISE (direct)					
MZKL	mmol/L	2.87	1.7%	14	288
CORNLEY \ ISE (indirect)					
(LA)乳酸					
Beckman AU series	mmol/L	0.997	4.4%	<5	106
Dedicated \ Lactate to pyruvate (L-P)					
Roche Hitachi/Hitachi Systems	mmol/L	0.972	2.4%	<5	<100
LEADMAN \ OTHER					
Roche cobas E601/602	mmol/L	1.04	10%	<5	<100
Roche \ Lactate to pyruvate (L-P)					
Beckman AU series	mmol/L	0.72	9.7%	<5	<100
MedicalSystem \ Lactate to pyruvate (L-P)					
(LDH)乳酸脱氢酶					
Beckman AU series	U/L	143	2.9%	9	270
LEADMAN \ Lactate to pyruvate (L-P)					
Roche Hitachi/Hitachi Systems	U/L	146	2.3%	9	257
BSBE \ Lactate to pyruvate (L-P)					
Mindray BS series	U/L	148	2.7%	9	255
MAKER \ Lactate to pyruvate (L-P)					
Uritest 8000 series	U/L	157	2.3%	9	248
Uritest \ OTHER					
Mindray BS series	U/L	148	3.7%	9	247
BSBE \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	155	2.1%	9	217
MedicalSystem \ Lactate to pyruvate (L-P)					
Mindray BS series	U/L	153	2.8%	8	489
Dedicated \ Lactate to pyruvate (L-P)					
TOSHIBA TBA	U/L	151	4.6%	8	238
OTHER \ Lactate to pyruvate (L-P)					
Beckman AU series	U/L	154	4.9%	8	201
MedicalSystem \ Lactate to pyruvate (L-P)					
OTHER	U/L	133	7.4%	7	197
SHLY \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	152	1.9%	7	170
MAKER \ Lactate to pyruvate (L-P)					
URIT-500B/330/180/55	U/L	157	3.3%	6	143
Uritest \ Lactate to pyruvate (L-P)					
Beckman AU series	U/L	152	3.1%	6	108
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	149	4.3%	5	174
BIOSINO \ Lactate to pyruvate (L-P)					
TOSHIBA TBA	U/L	146	2.2%	5	143
Autobio \ Lactate to pyruvate (L-P)					
Beckman AU series	U/L	153	5.6%	5	136
MedicalSystem \ OTHER					

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
11 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series MAKER \ Lactate to pyruvate (L-P)	U/L	149	2.6%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	153	3.7%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	153	1.5%	<5	195
(LDL-C)低密度脂蛋白胆固醇					
URIT-500B/330/180/55 Uritest \ SUR	mmol/L	1.19	4.2%	9	276
Beckman AU series LEADMAN \ SUR	mmol/L	1.19	2.4%	9	269
Roche Hitachi/Hitachi Systems Reebio \ SUR	mmol/L	1.17	11.6%	9	268
Beckman AU series Dedicated \ Enzymatic	mmol/L	1.3	2.4%	9	253
Mindray BS series BSBE \ SUR	mmol/L	1.18	4.6%	9	248
Uritest 8000 series Uritest \ SUR	mmol/L	1.17	2.2%	9	246
TOSHIBA TBA OTHER \ SUR	mmol/L	1.08	5.2%	8	176
Beckman AU series KeFang(Guangzhou) \ Direct Measure	mmol/L	1.19	3.8%	8	168
Roche Hitachi/Hitachi Systems MedicalSystem \ SUR	mmol/L	1.24	5.8%	7	191
Beckman AU series Reebio \ OTHER	mmol/L	1.21	5.5%	6	216
OTHER SHLY \ Direct Measure	mmol/L	1.34	3.7%	6	158
Beckman AU series chongqiong zybio \ Enzymatic	mmol/L	1.43	3.9%	5	<100
Beckman AU series KHB \ OTHER	mmol/L	1.06	1.9%	5	268
DiruiCS series DIRUI \ SUR	mmol/L	1.14	4.4%	5	152
TOSHIBA TBA Mindray \ Direct Measure	mmol/L	1.43	3.2%	5	138
Beckman AU series MedicalSystem \ SUR	mmol/L	1.13	1.7%	5	131
TOSHIBA TBA KHB \ SUR	mmol/L	1.17	3.3%	<5	<100
Mindray BS series Mindray \ Direct Measure	mmol/L	1.48	1.6%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Direct Measure	mmol/L	1.36	2.5%	<5	<100
(LIP)脂肪酶					
Beckman AU series Beckman Coulter \ OTHER	U/L	31.8	3.7%	9	440
Beckman AU series LEADMAN \ Colorimetry	U/L	22.4	19.1%	9	267
Roche Hitachi/Hitachi Systems Reebio \ Glucose oxidase (GOD)	U/L	43.3	12%	9	253
Beckman AU series Reebio \ OTHER	U/L	24.2	19.1%	7	131
Roche Hitachi/Hitachi Systems BSBE \ OTHER	U/L	31.5	8%	<5	<100
Mindray BS series BSBE \ OTHER	U/L	62.4	1.5%	<5	<100
Beckman AU series BSBE \ OTHER	U/L	42.4	3%	<5	<100
Beckman AU series KINGSBIO \ OTHER	U/L	27.3	9.7%	14	378
Beckman AU series Iprocom \ Colorimetry	U/L	21	10%	<5	<100
Beckman AU series MedicalSystem \ Enzymatic Cycling Assays Method	U/L	30.7	10%	<5	<100
Roche cobas E601/602 Roche \ Enzymatic Cycling Assays Method	U/L	31.6	10%	<5	<100
Beckman AU series Autobio \ OTHER	U/L	30.8	8.9%	<5	<100
(Lp(a))脂蛋白a					
DiruiCS series DIRUI \ OTHER	mg/L	45.1	5.3%	5	152
TOSHIBA TBA Autobio \ OTHER	mg/L	78.3	20.6%	5	117
Roche cobas 4000/6000/8000 OTHER \ Immunoturbidimetry (ITA)	mg/L	45	85.1%	<5	<100
Uritest 8000 series Uritest \ Immunoturbidimetry (ITA)	mg/L	68	8.3%	<5	116
Mindray BS series chongqiong zybio \ OTHER	mg/L	43.6	8.6%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mg/L	62.9	3.7%	<5	149
Roche Hitachi/Hitachi Systems chongqiong zybio \ OTHER	mg/L	134	34.1%	<5	<100
Mindray CL-2000i/1000i/6000i series chongqiong zybio \ OTHER	mg/L	79.6	1.6%	<5	<100
Beckman AU series Iprocom \ Immunoturbidimetry (ITA)	mg/L	50	10%	<5	<100
Beckman AU series Reebio \ OTHER	mg/L	58.3	1.9%	<5	<100

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
12 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray CL-2000i/1000i/6000i series Mindray \ OTHER	mg/L	53.6	4.5%	<5	<100
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	mg/L	78.3	3.4%	<5	<100
(MAO)单胺氧化酶					
Beckman AU series Iprocom \ Colorimetry	U/L	3	10%	<5	<100
Mindray BS series chongqiong zybio \ OTHER	U/L	4.05	28.6%	<5	<100
(Mg)镁					
Beckman AU series LEADMAN \ Xylidyl Blue	mmol/L	0.66	2.2%	9	270
Roche Hitachi/Hitachi Systems Reebio \ Xylidyl Blue	mmol/L	0.781	9%	9	251
Mindray BS series BSBE \ ICP-MS	mmol/L	0.709	6.7%	9	246
TOSHIBA TBA OTHER \ Xylidyl Blue	mmol/L	0.635	4.4%	8	213
Roche Hitachi/Hitachi Systems MedicalSystem \ Xylidyl Blue	mmol/L	0.683	4.9%	7	188
Beckman AU series MedicalSystem \ Arsenazo □	mmol/L	0.684	5.2%	7	151
Roche Hitachi/Hitachi Systems Health World \ Arsenazo □	mmol/L	0.694	4.8%	<5	<100
Beckman AU series Dedicated \ Xylidyl Blue	mmol/L	0.649	2%	<5	106
DiruiCS series DIRUI \ Xylidyl Blue	mmol/L	0.711	11.5%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Calmagite colorimetry	mmol/L	0.748	3%	<5	<100
Canon TBA Series chongqiong zybio \ Xylidyl Blue	mmol/L	0.705	6.5%	<5	<100
TOSHIBA TBA MedicalSystem \ Arsenazo □	mmol/L	0.619	6.5%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Xylidyl Blue	mmol/L	0.629	3.8%	<5	115
Siemens ADVIA 1800/2400 BIOSINO \ Xylidyl Blue	mmol/L	0.668	3.2%	24	1347
Beckman AU series BSBE \ Xylidyl Blue	mmol/L	0.635	9.5%	<5	<100
Beckman AU series Reebio \ Arsenazo □	mmol/L	0.69	6.1%	16	584
Beckman AU series Dedicated \ Calmagite colorimetry	mmol/L	0.668	1.9%	15	616
Beckman AU series OTHER \ Xylidyl Blue	mmol/L	0.654	4.1%	14	387
Beckman AU series DiaSys \ Xylidyl Blue	mmol/L	0.672	3.9%	13	325
(Na)钠					
Mindray BS series MAKER \ ISE (indirect)	mmol/L	110	0.8%	9	256
Roche Hitachi/Hitachi Systems Reebio \ ISE (indirect)	mmol/L	111	4.9%	9	250
Mindray BS series BSBE \ ICP-MS	mmol/L	112	1.7%	9	233
Mindray BS series Dedicated \ ISE (indirect)	mmol/L	110	1.2%	8	502
Beckman AU series Reebio \ ISE (direct)	mmol/L	113	1%	7	<100
Roche Hitachi/Hitachi Systems Hitachi \ ISE (direct)	mmol/L	111	1.7%	7	191
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	110	1.4%	5	214
OTHER OTHER \ ISE (direct)	mmol/L	112	0.8%	<5	<100
TOSHIBA TBA OTHER \ ISE (indirect)	mmol/L	108	1.2%	<5	<100
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	109	0.9%	<5	107
Beckman AU series Dedicated \ ISE (direct)	mmol/L	107	2.3%	<5	106
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	109	1.1%	37	1279
DiruiCS series DIRUI \ ISE (indirect)	mmol/L	116	0.8%	<5	<100
Canon TBA Series chongqiong zybio \ ISE (direct)	mmol/L	114	3.7%	<5	<100
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	109	1.1%	24	875
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	111	1.6%	24	1352
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	126	14.9%	<5	<100
Beckman AU series BSBE \ ISE (direct)	mmol/L	109	4%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ ISE (direct)	mmol/L	111	1.5%	<5	<100
(NEFA)游离脂肪酸					
Siemens ADVIA 1800/2400 OTHER \ ACS - ACOD	mmol/L	0.665	3.7%	36	2011

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
13 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series KINGSBIO \ OTHER	mmol/L	0.671	2.2%	14	392
Beckman AU series DongOu \ ACS - ACOD	mmol/L	0.607	2.1%	10	458
Siemens ADVIA 1800/2400 Siemens \ ACS - ACOD	mmol/L	0.664	0.7%	<5	<100
Beckman AU series Iprocom \ ACS - ACOD	mmol/L	0.54	10%	<5	<100
Roche cobas E601/602 Roche \ ACS - ACOD	mmol/L	0.626	10%	<5	<100
(P)磷					
Beckman AU series Beckman Coulter \ OTHER	mmol/L	0.695	1.9%	9	482
Beckman AU series LEADMAN \ Phosphomolybdate-UV-end point	mmol/L	0.667	1.6%	9	270
Roche Hitachi/Hitachi Systems Reebio \ Phosphomolybdate-UV-end point	mmol/L	0.741	12.6%	9	266
Mindray BS series BSBE \ Phosphomolybdate UV-rate	mmol/L	0.691	11.7%	9	240
TOSHIBA TBA SHLY \ OTHER	mmol/L	0.731	2.2%	8	230
Roche Hitachi/Hitachi Systems MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	0.672	4.9%	7	184
Beckman AU series Reebio \ OTHER	mmol/L	0.676	2.9%	6	<100
Roche Hitachi/Hitachi Systems MAKER \ Phosphomolybdate-UV-end point	mmol/L	0.697	7.1%	6	186
TOSHIBA TBA Autobio \ Phosphomolybdate UV-rate	mmol/L	0.801	4.2%	5	143
Beckman AU series Reebio \ Phosphomolybdate-UV-end point	mmol/L	0.692	8.7%	5	116
Beckman AU series MAKER \ OTHER	mmol/L	0.74	4.8%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Phosphomolybdate-UV-end point	mmol/L	0.694	2.5%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	0.696	1.6%	<5	202
Beckman AU series Dedicated \ Phosphomolybdate-UV-end point	mmol/L	0.646	5.6%	<5	106
Canon TBA Series chongqiong zybjo \ Phosphomolybdate-UV-end point	mmol/L	0.74	1.7%	<5	<100
TOSHIBA TBA MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	0.665	5.1%	<5	<100
Beckman AU series BSBE \ OTHER	mmol/L	0.642	4.9%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Phosphomolybdate-UV-end point	mmol/L	0.708	4%	24	1351
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	0.674	8.2%	<5	<100
(PAB)前白蛋白					
Beckman AU series KHB \ OTHER	g/L	0.182	4.5%	6	343
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	g/L	0.174	1.9%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Immunonephelometry (INA)	g/L	0.153	3.8%	<5	<100
Beckman AU series Reebio \ OTHER	g/L	0.142	3.2%	<5	137
Uritest 8000 series Uritest \ Immunoturbidimetry (ITA)	g/L	0.158	8.1%	<5	116
Roche Hitachi/Hitachi Systems MAKER \ OTHER	g/L	0.164	4.3%	<5	115
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	g/L	0.166	0.5%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Immunonephelometry (INA)	g/L	0.181	0.7%	<5	<100
Mindray BS series Mindray \ Immunonephelometry (INA)	g/L	0.169	1.4%	<5	<100
(PRO/ TP)(总) 蛋白					
Beckman AU series Beckman Coulter \ OTHER	g/L	40.3	2.6%	9	484
Roche Hitachi/Hitachi Systems BSBE \ Biuret-endpoint	g/L	43.9	10.6%	9	272
Beckman AU series LEADMAN \ Biuret-endpoint	g/L	42.5	3.1%	9	270
Mindray BS series MAKER \ Biuret-endpoint	g/L	41.7	6.5%	9	255
Uritest 8000 series Uritest \ Biuret-endpoint	g/L	41.2	1.3%	9	245
Mindray BS series BSBE \ Biuret-rate	g/L	43.7	3.4%	9	223
Roche Hitachi/Hitachi Systems MedicalSystem \ Biuret-endpoint	g/L	42.8	1.5%	9	219
Mindray BS series Dedicated \ Biuret-endpoint	g/L	42.4	1.4%	8	491
Beckman AU series MedicalSystem \ Biuret-endpoint	g/L	40.6	4.1%	8	186
OTHER SHLY \ OTHER	g/L	43	9.3%	7	169
Beckman AU series MAKER \ Biuret-rate	g/L	41.5	2.2%	6	<100

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
14 / 16								

		批号：376B8D4			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
URIT-500B/330/180/55	g/L	40.7	3.8%	6	142
Uritest \ Biuret-endpoint					
Mindray BS series	g/L	42.5	1.6%	6	114
Mindray \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	42.7	1.6%	5	215
BIOSINO \ Biuret-endpoint					
TOSHIBA TBA	g/L	42	1%	5	137
Mindray \ Biuret-endpoint					
TOSHIBA TBA	g/L	41.5	1.1%	5	121
MedicalSystem \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	43.4	0.9%	<5	<100
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	42.8	4.7%	<5	<100
BIOSINO \ OTHER					
Beckman AU series	g/L	42	4.9%	<5	<100
MAKER \ OTHER					
(RBP)视黄醇结合蛋白					
Roche Hitachi/Hitachi Systems	mg/dL	2.97	4.7%	7	<100
MedicalSystem \ Immunoturbidimetry (ITA)					
Beckman AU series	mg/dL	2.6	10%	<5	<100
Iprocom \ Immunoturbidimetry (ITA)					
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems	μmol/L	12	11.4%	9	267
Reebio \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	12	4.5%	9	256
LEADMAN \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	13.5	3.5%	9	252
Dedicated \ Enzymatic Cycling Assays Method					
Uritest 8000 series	μmol/L	12.4	7.5%	9	246
Uritest \ Enzymatic Cycling Assays Method					
Mindray BS series	μmol/L	10.4	8.7%	9	232
BSBE \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	11.8	2.8%	7	345
KHB \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	11	9%	7	185
MedicalSystem \ Enzymatic Cycling Assays Method					
TOSHIBA TBA	μmol/L	11.8	3.8%	7	171
KHB \ Enzymatic Cycling Assays Method					
URIT-500B/330/180/55	μmol/L	11.9	6.8%	6	144
Uritest \ Enzymatic Cycling Assays Method					
TOSHIBA TBA	μmol/L	8.88	5.1%	5	137
Mindray \ Enzymatic Cycling Assays Method					
Roche Hitachi/Hitachi Systems	μmol/L	11.2	6.7%	<5	<100
BIOSINO \ Enzymatic Cycling Assays Method					
Roche Hitachi/Hitachi Systems	μmol/L	13.7	4.6%	<5	<100
MAKER \ Enzymatic Cycling Assays Method					
Roche cobas 4000/6000/8000	μmol/L	12.4	3.1%	<5	<100
WeiRiKang \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	13.1	5.2%	<5	<100
Reebio \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	10.2	4%	<5	135
Beckman Coulter \ Enzymatic Cycling Assays Method					
URIT-500B/330/180/55	μmol/L	11.2	5.1%	<5	133
Uritest \ OTHER					
Beckman AU series	μmol/L	10.7	5.7%	<5	106
sysmex \ Enzymatic Cycling Assays Method					
Siemens ADVIA 1800/2400	μmol/L	13.5	4.5%	36	1996
BIOSINO \ Enzymatic Cycling Assays Method					
DiruiCS series	μmol/L	13.8	15.4%	<5	<100
DIRUI \ Enzymatic Cycling Assays Method					
(TBIL)总胆红素					
Beckman AU series	μmol/L	15.9	3.2%	9	267
LEADMAN \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	13.7	6.6%	9	266
BSBE \ Vanadate					
Uritest 8000 series	μmol/L	15.2	2.8%	9	247
Uritest \ OTHER					
Mindray BS series	μmol/L	14.5	6.2%	9	209
BSBE \ Vanadate					
TOSHIBA TBA	μmol/L	16.5	3%	7	199
OTHER \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	15.4	1.6%	7	193
MedicalSystem \ sodium nitrite oxidase method					
Beckman AU series	μmol/L	17.6	2%	7	150
MedicalSystem \ sodium nitrite oxidase method					
Beckman AU series	μmol/L	15.6	1.8%	6	346
Beckman Coulter \ Diazotization -Blank Compensation					
OTHER	μmol/L	13.8	6.1%	6	159
SHLY \ sodium nitrite oxidase method					
URIT-500B/330/180/55	μmol/L	13.7	4.9%	6	144
Uritest \ Diazotization					
Mindray BS series	μmol/L	14.2	7%	6	132
Mindray \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	14.2	3.9%	5	215
DiaSys \ Diazotization -Blank Compensation					
Beckman AU series	μmol/L	14	3%	5	141
Reebio \ OTHER					
TOSHIBA TBA	μmol/L	13.7	3.1%	5	140
Mindray \ Vanadate					
Beckman AU series	μmol/L	15.4	2.1%	5	129
Beckman Coulter \ Diazotization -No blank Compensation					

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
15 / 16								

		批号：376B8D4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μmol/L	13.9	4.2%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ sodium nitrite oxidase method	μmol/L	15.4	4.9%	<5	<100
URIT-500B/330/180/55 Uritest \ OTHER	μmol/L	15.3	3.1%	<5	133
Siemens ADVIA 1800/2400 BIOSINO \ Vanadate	μmol/L	15.2	4.6%	36	2021
(TG)甘油三酯					
Beckman AU series Beckman Coulter \ OTHER	mmol/L	0.834	1.7%	9	482
Beckman AU series LEADMAN \ OTHER	mmol/L	0.816	5.2%	9	270
Beckman AU series Dedicated \ GPO-POD	mmol/L	0.876	6.8%	9	257
Uritest 8000 series Uritest \ OTHER	mmol/L	0.88	2.8%	9	246
Roche Hitachi/Hitachi Systems BSBE \ GPO-POD	mmol/L	0.907	12.3%	8	245
Roche Hitachi/Hitachi Systems MedicalSystem \ GPO-POD	mmol/L	0.931	2.9%	7	193
Mindray BS series BSBE \ OTHER	mmol/L	0.882	17.3%	7	192
TOSHIBA TBA OTHER \ OTHER	mmol/L	0.879	6.6%	6	163
OTHER SHLY \ OTHER	mmol/L	0.871	10.3%	6	160
URIT-500B/330/180/55 Uritest \ GPO-POD	mmol/L	0.865	2.7%	6	142
DiruiCS series DIRUI \ OTHER	mmol/L	0.81	2.7%	5	152
TOSHIBA TBA KHB \ GPO-POD	mmol/L	0.811	2.5%	5	144
TOSHIBA TBA Autobio \ GPO-POD	mmol/L	0.819	16.1%	5	123
Mindray BS series Mindray \ GPO-POD	mmol/L	0.876	1.7%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ GPO-POD	mmol/L	0.863	1.9%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	0.77	7.2%	<5	139
URIT-500B/330/180/55 Uritest \ OTHER	mmol/L	0.853	3%	<5	133
Siemens ADVIA 1800/2400 BIOSINO \ GPO-POD	mmol/L	0.843	2%	36	2014
Beckman AU series BSBE \ OTHER	mmol/L	0.883	5%	<5	<100
(TRF)转铁蛋白					
Beckman AU series BSBE \ OTHER	mg/dL	166	2.5%	<5	<100
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	mg/dL	151	2.4%	<5	<100
Roche cobas E601/602 Roche \ Immunonephelometry (INA)	mg/dL	167	10%	<5	<100
(UA)尿酸					
URIT-500B/330/180/55 Uritest \ OTHER	μmol/L	165	7.8%	9	274
Beckman AU series LEADMAN \ Uricense colorimetry	μmol/L	167	1%	9	270
Uritest 8000 series Uritest \ OTHER	μmol/L	162	1.3%	9	248
Mindray BS series BSBE \ Uricense colorimetry	μmol/L	161	2.7%	9	233
Mindray BS series Dedicated \ Uricense colorimetry	μmol/L	156	1.8%	8	494
Roche Hitachi/Hitachi Systems BSBE \ Uricense colorimetry	μmol/L	161	5%	8	244
Beckman AU series MedicalSystem \ Uricense colorimetry	μmol/L	170	4.8%	8	179
Roche Hitachi/Hitachi Systems MedicalSystem \ Uricense colorimetry	μmol/L	167	1.8%	7	193
Beckman AU series Reebio \ Uricense colorimetry	μmol/L	160	1.8%	6	135
Roche Hitachi/Hitachi Systems BIOSINO \ Uricense-UV	μmol/L	152	2.3%	5	216
TOSHIBA TBA Autobio \ Uricense colorimetry	μmol/L	169	2.4%	5	142
Beckman AU series MAKER \ OTHER	μmol/L	156	3.2%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Uricense-UV	μmol/L	159	1.7%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Uricense colorimetry	μmol/L	169	4%	<5	136
Beckman AU series Dedicated \ Uricense-UV	μmol/L	163	3.9%	<5	106
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	155	1.1%	<5	103
Siemens ADVIA 1800/2400 BIOSINO \ Uricense colorimetry	μmol/L	165	3.4%	36	2010
DiruiCS series DIRUI \ OTHER	μmol/L	157	2.7%	<5	<100

REF	581-1	6×5mL	EXP	2025-11-30	LOT	376B8D4	更新日期：	2024-04
16 / 16								

批号：376B8D4					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Canon TBA Series chongqiong zybjo \ Uricase colorimetry	μmol/L	157	1.6%	<5	<100
(UIBC)不饱和铁结合力					
Beckman AU series Dedicated \ Ferrozine colorimetry	μmol/L	18.8	3.7%	5	220
TOSHIBA TBA XJKC \ Ferrozine colorimetry	μmol/L	26.2	6.4%	<5	<100
Siemens ADVIA 1800/2400 DiaSys \ Ferrozine colorimetry	μmol/L	20	4.8%	12	655
Beckman AU series LEADMAN \ Ferene	μmol/L	19.4	13.6%	<5	<100
Roche cobas E601/602 Roche \ Ferrozine colorimetry	μmol/L	18.5	10%	<5	<100
(Urea)尿素					
Beckman AU series LEADMAN \ Urease-UV-rate	mmol/L	4.89	3.7%	9	269
Roche Hitachi/Hitachi Systems BSBE \ Urease electrode	mmol/L	4.55	5.4%	9	267
Beckman AU series Dedicated \ Urease colorimetry	mmol/L	4.78	3.1%	9	260
Uritest 8000 series Uritest \ OTHER	mmol/L	5.1	3.7%	9	246
Mindray BS series BSBE \ Urease electrode	mmol/L	4.79	3.5%	9	229
TOSHIBA TBA OTHER \ OTHER	mmol/L	4.7	2.4%	8	232
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	4.97	3.6%	7	294
Beckman AU series MedicalSystem \ OTHER	mmol/L	4.46	1.7%	7	152
Beckman AU series chongqiong zybjo \ OTHER	mmol/L	4.83	2.9%	6	163
URIT-500B/330/180/55 Uritest \ Diacetyl monoxime	mmol/L	4.53	5.9%	6	144
Roche Hitachi/Hitachi Systems MAKER \ Urease-UV-rate	mmol/L	4.95	2.8%	<5	<100
Beckman AU series MAKER \ OTHER	mmol/L	4.77	3%	<5	<100
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	4.81	3.8%	<5	138
URIT-500B/330/180/55 Uritest \ OTHER	mmol/L	4.57	2.9%	<5	133
TOSHIBA TBA OTHER \ Urease-UV-rate	mmol/L	4.43	3.6%	<5	123
TOSHIBA TBA Autobio \ Urease electrode	mmol/L	4.68	5.6%	<5	106
DiruiCS series DIRUI \ OTHER	mmol/L	4.73	2.7%	<5	<100
Roche Hitachi/Hitachi Systems Whitman \ Urease-UV-rate	mmol/L	4.62	3%	<5	<100
Beckman AU series chongqiong zybjo \ Urease colorimetry	mmol/L	5.24	4.1%	<5	<100
(β2M)β-2-微球蛋白					
Roche Hitachi/Hitachi Systems MedicalSystem \ Immunoturbidimetry (ITA)	ng/mL	1777	7.7%	7	<100
Roche cobas 4000/6000/8000 WeiRiKang \ Immunoturbidimetry (ITA)	ng/mL	834	8.8%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	ng/mL	611	4.4%	<5	202
Uritest 8000 series Uritest \ Immunoturbidimetry (ITA)	ng/mL	968	8%	<5	116
Mindray BS series chongqiong zybjo \ OTHER	ng/mL	1128	1.6%	<5	<100
Mindray CL-2000i/1000i/6000i series chongqiong zybjo \ OTHER	ng/mL	1742	3.2%	<5	<100
Beckman AU series Reebio \ OTHER	ng/mL	859	9.8%	15	592
Beckman AU series Iprocom \ Immunonephelometry (INA)	ng/mL	900	10%	<5	<100
Beckman AU series Autobio \ OTHER	ng/mL	758	11.5%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ OTHER	ng/mL	513	5.7%	<5	<100
Mindray BS series Mindray \ OTHER	ng/mL	926	7.8%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。