

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
1 / 6								

		批号：376DFDE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(a-HBDH)α-羟丁酸脱氢酶					
Roche cobas c 501/502/701/702 Roche \ DGKC	U/L	483	1.2%	8	243
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	U/L	504	5.9%	7	304
Beckman AU series MAKER \ DGKC	U/L	421	0.5%	16	448
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	U/L	463	7.3%	11	310
Beckman AU series Beckman Coulter \ DGKC	U/L	480	12%	<5	<100
(Alb)白蛋白					
Roche cobas c 501/502/701/702 Roche \ Bromocresol green	g/L	45.6	2%	8	244
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	46.9	2.3%	7	193
Abbott ARCHITECT ci series BIOSINO \ Bromocresol green	g/L	44.3	4.8%	27	693
Roche Hitachi/Hitachi Systems MedicalSystem \ Bromocresol green	g/L	44.7	1.4%	21	767
Beckman AU series MAKER \ Bromocresol green	g/L	46.5	0.8%	16	441
Siemens Dimension series Siemens \ Bromocresol green	g/L	48.7	1.1%	<5	<100
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	44.8	12%	<5	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	g/L	41	20.4%	<5	<100
(ALP)碱性磷酸酶					
Roche cobas c 501/502/701/702 Roche \ Rate -AMP buffer	U/L	410	2.7%	8	243
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	420	2.3%	7	191
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	358	19.6%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Rate -AMP buffer	U/L	487	6.4%	27	781
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	520	3.9%	21	769
Beckman AU series MAKER \ Rate -AMP buffer	U/L	487	1.7%	16	451
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	555	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	U/L	428	1.5%	<5	<100
(ALT/ SGPT)丙氨酸氨基转移酶					
Roche cobas c 501/502/701/702 Roche \ UV-rate without P5P	U/L	167	2.5%	8	245
Roche cobas 4000/6000/8000 Dedicated \ UV-rate with P5P	U/L	169	3.7%	7	193
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	179	3.5%	27	708
Roche Hitachi/Hitachi Systems MedicalSystem \ UV-rate without P5P	U/L	203	3.4%	21	765
Beckman AU series MAKER \ OTHER	U/L	195	1.5%	16	450
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	213	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	U/L	203	1.2%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ UV-rate without P5P	U/L	192	3.4%	<5	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	129	20.9%	<5	<100
(ammo)血氨					
Siemens Dimension series MedicalSystem \ Enzymatic	mmol/L	0.597	13.5%	<5	<100
(AMY)淀粉酶					
Roche cobas c 501/502/701/702 Roche \ PNP-G7	U/L	455	1.4%	8	242
Roche cobas 4000/6000/8000 Roche \ PNP-G5	U/L	456	7.8%	7	192
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	201	3.6%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ PNP-G7	U/L	386	3.7%	16	554
Beckman AU series MAKER \ PNP-G7	U/L	335	9.4%	15	423
Abbott ARCHITECT ci series BIOSINO \ PNP-G7	U/L	408	2.7%	11	311
Beckman AU series Beckman Coulter \ Iodine colorimetry	U/L	384	12%	<5	<100
Siemens Dimension series KUAKE \ OTHER	U/L	394	1.5%	<5	<100
(Apo-A I)载脂蛋白A I					
Roche cobas c 501/502/701/702 Roche \ immunoturbidimetric	g/L	1.84	1.9%	9	256

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
2 / 6								

		批号：376DFDE			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Dedicated \ immunoturbidimetric	g/L	2.07	3%	7	191
Roche Hitachi/Hitachi Systems MedicalSystem \ immunoturbidimetric	g/L	2.04	11.8%	18	632
(Apo-B)载脂蛋白B					
Roche cobas c 501/502/701/702 Roche \ immunoturbidimetric	g/L	0.946	5.1%	8	242
Roche cobas 4000/6000/8000 Dedicated \ immunoturbidimetric	g/L	1.02	4.1%	7	189
Roche Hitachi/Hitachi Systems MedicalSystem \ immunoturbidimetric	g/L	0.966	3.8%	18	633
(AST/ SGOT)天门冬氨酸氨基转移酶					
Roche cobas c 501/502/701/702 Roche \ UV-rate without P5P	U/L	216	2.2%	8	243
Roche cobas 4000/6000/8000 Dedicated \ UV-rate with P5P	U/L	222	2.8%	7	188
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	191	4.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	222	8.8%	27	601
Roche Hitachi/Hitachi Systems MedicalSystem \ UV-rate without P5P	U/L	241	4.2%	21	765
Beckman AU series MAKER \ OTHER	U/L	228	1.2%	16	449
Roche Hitachi/Hitachi Systems Wako \ UV-rate without P5P	U/L	237	1%	<5	<100
Siemens Dimension series Siemens \ OTHER	U/L	214	2.3%	<5	<100
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	230	12%	<5	<100
(BUN)尿素氮					
Roche cobas 4000/6000/8000 Roche \ Urease-UV-rate	mmol/L	18.3	3.4%	7	202
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	16.2	8.5%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Urease-UV-rate	mmol/L	17.4	3.9%	21	766
Siemens Dimension series Siemens \ OTHER	mmol/L	18.3	1.9%	<5	<100
(C4)补体 C4					
Abbott ARCHITECT ci series BIOSINO \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	22.6	7.3%	8	189
(Ca)钙					
Roche cobas c 501/502/701/702 Roche \ BAPTA	mmol/L	2.92	2.1%	8	242
Roche cobas 4000/6000/8000 Dedicated \ BAPTA	mmol/L	2.92	2.5%	7	189
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	2.82	3.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Arsenazo III	mmol/L	2.74	2.9%	27	736
Roche Hitachi/Hitachi Systems Wako \ MTB colorimetry	mmol/L	2.72	2%	19	682
Beckman AU series MAKER \ Arsenazo III	mmol/L	2.79	2.8%	16	448
Beckman AU series Beckman Coulter \ BAPTA	mmol/L	2.81	12%	<5	<100
(CHE)胆碱酯酶					
Roche Hitachi/Hitachi Systems MedicalSystem \ DTNB	U/L	8871	2.1%	7	297
Roche cobas 4000/6000/8000 Roche \ DTNB	U/L	7997	2.7%	7	190
Beckman AU series MAKER \ Butyrylthiocholine Method	U/L	7463	0.1%	16	452
Roche cobas c 501/502/701/702 Roche \ Butyrylthiocholine Method	U/L	7080	17.3%	12	340
Abbott ARCHITECT ci series BIOSINO \ DTNB	U/L	7968	2.4%	11	312
Beckman AU series Beckman Coulter \ Butyrylthiocholine Method	U/L	8867	12%	<5	<100
Siemens Dimension series KUAKE \ OTHER	U/L	8816	3%	<5	<100
(CHOL/ TC)总胆固醇					
Roche cobas c 501/502/701/702 Roche \ Cholesterol oxidase	mmol/L	5.14	1.5%	8	245
Roche cobas 4000/6000/8000 Dedicated \ Cholesterol oxidase	mmol/L	5.28	2.3%	7	193
Abbott ARCHITECT ci series BIOSINO \ Cholesterol oxidase	mmol/L	5.16	2.2%	27	764
Roche Hitachi/Hitachi Systems MedicalSystem \ Cholesterol oxidase	mmol/L	5.4	2.2%	21	767
Beckman AU series MAKER \ Cholesterol oxidase	mmol/L	5.83	3%	16	455
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	5.22	12%	<5	<100

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
3 / 6								

批号：376DFDE					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CK)肌酸激酶					
Roche cobas c 501/502/701/702 Roche \ Urease-UV-rate	U/L	673	1.9%	9	251
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	683	2.1%	7	191
Abbott ARCHITECT ci series BIOSINO \ Urease-UV-rate	U/L	638	6.3%	27	740
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	661	3.7%	19	680
Beckman AU series MAKER \ DGKC	U/L	661	0.6%	16	450
Beckman AU series Beckman Coulter \ Urease-UV-rate	U/L	682	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	U/L	635	1.5%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Roche Hitachi/Hitachi Systems MedicalSystem \ DGKC	U/L	51.4	10.2%	7	285
Beckman AU series MAKER \ Immune suppression	U/L	42.5	3.5%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Immune suppression	U/L	28.3	5.5%	11	311
Roche cobas c 501/502/701/702 Roche \ Enzyme immunoassay (EIA)	U/L	70.3	20.1%	10	259
Roche Hitachi/Hitachi Systems Roche \ DGKC	U/L	50.9	10%	<5	<100
Abbott ARCHITECT ci series Abbott \ Immune suppression	U/L	28.5	10.2%	<5	<100
(Cl)氯					
Roche cobas c 501/502/701/702 Roche \ ISE (direct)	mmol/L	100	1.5%	9	254
Roche cobas 4000/6000/8000 Dedicated \ ISE (direct)	mmol/L	101	2.6%	7	192
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	108	1.5%	27	795
Roche Hitachi/Hitachi Systems Roche \ ISE (indirect)	mmol/L	102	1.4%	21	755
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	106	3.4%	17	489
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	113	14.5%	<5	<100
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Beckman AU series Beckman Coulter \ PEPC	mmol/L	7.09	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	mmol/L	4.21	27%	<5	<100
(Cr)肌酐					
Roche cobas c 501/502/701/702 Roche \ Enzymatic	μmol/L	460	1.9%	8	242
Roche cobas 4000/6000/8000 Dedicated \ Enzymatic	μmol/L	473	2%	7	188
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	526	7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Picric acid-Blank Compensation	μmol/L	465	2.5%	27	763
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	μmol/L	479	3.7%	21	769
Beckman AU series MAKER \ Enzymatic	μmol/L	496	1.1%	16	450
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	471	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	μmol/L	483	1.5%	<5	<100
(CRP)C反应蛋白					
Roche Hitachi/Hitachi Systems MedicalSystem \ Latex agglutination (LA)	mg/dL	151	11.1%	6	241
(Cu)铜					
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	19.4	12%	<5	<100
(DBIL)直接胆红素					
Roche cobas c 501/502/701/702 Roche \ Diazotization	μmol/L	38.6	6.1%	9	259
Roche cobas 4000/6000/8000 Dedicated \ Diazotization	μmol/L	41.2	5.2%	7	189
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	36.5	7.2%	21	761
Beckman AU series MAKER \ Vanadate	μmol/L	49.4	2.8%	16	451
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	46.1	4.4%	11	312
Siemens Dimension series Siemens \ Vanadate	μmol/L	28.5	3.3%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	44.7	12%	<5	<100

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
4 / 6								

		批号：376DFDE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Fe)铁					
Beckman AU series LEADMAN \ Ferrozine colorimetry	μmol/L	31.3	0.9%	16	440
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	28.5	12%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
Roche cobas c 501/502/701/702 Roche \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	189	1.9%	8	243
Roche cobas 4000/6000/8000 Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	196	3.4%	7	194
Abbott ARCHITECT ci series BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	201	4.4%	27	756
Roche Hitachi/Hitachi Systems MedicalSystem \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	190	4.1%	21	761
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	229	3.1%	<5	<100
Beckman AU series MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	202	1%	16	454
Beckman AU series Beckman Coulter \ γ-glutamyl -4-nitroaniline	U/L	186	12%	<5	<100
Siemens Dimension series Siemens \ OTHER	U/L	222	1.3%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	182	0.9%	<5	<100
(GLU)葡萄糖					
Roche cobas c 501/502/701/702 Roche \ Hexokinase	mmol/L	20.8	1.5%	8	242
Roche cobas 4000/6000/8000 Dedicated \ Hexokinase	mmol/L	21.3	9.2%	7	193
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	12.2	16.3%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Hexokinase	mmol/L	15.7	2.2%	27	768
Roche Hitachi/Hitachi Systems MedicalSystem \ Glucose oxidase (GOD)	mmol/L	14.3	3%	21	765
Beckman AU series MAKER \ Glucose oxidase (GOD)	mmol/L	16.4	1%	16	453
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	14.8	12%	<5	<100
Siemens Dimension series Siemens \ Glucose oxidase (GOD)	mmol/L	15.9	1.9%	<5	<100
(HCY)同型半胱氨酸					
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	15.5	5%	7	295
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	15	8.3%	7	176
Beckman AU series MAKER \ OTHER	μmol/L	19.4	2.3%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Roche cobas 4000/6000/8000 Dedicated \ Direct Measure	mmol/L	1.08	9.4%	7	193
Abbott ARCHITECT ci series BIOSINO \ selected suppression	mmol/L	1.64	4.7%	27	764
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	mmol/L	1.7	2.8%	18	637
Beckman AU series MAKER \ OTHER	mmol/L	1.53	5.2%	16	452
Roche cobas c 501/502/701/702 Roche \ Direct Measure	mmol/L	1.19	24%	11	309
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.08	12%	<5	<100
(IgG)免疫球蛋白 IgG					
Abbott ARCHITECT ci series BIOSINO \ Immunonephelometry (INA)	mg/dL	989	2.8%	8	193
(K)钾					
Roche cobas c 501/502/701/702 Roche \ ISE (direct)	mmol/L	6.95	0.9%	8	245
Roche cobas 4000/6000/8000 Dedicated \ ISE (direct)	mmol/L	6.84	3.3%	7	200
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	6.9	1.8%	27	765
Roche Hitachi/Hitachi Systems Roche \ ISE (indirect)	mmol/L	6.94	2.1%	21	765
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	6.73	7.6%	<5	<100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	6.85	4.3%	17	502
(LA)乳酸					
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	8.35	12%	<5	<100
(LDH)乳酸脱氢酶					
Roche cobas c 501/502/701/702 Roche \ Lactate to pyruvate (L-P)	U/L	452	1.4%	8	243
Roche cobas 4000/6000/8000 Dedicated \ Lactate to pyruvate (L-P)	U/L	460	3.4%	7	187

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
5 / 6								

		批号：376DFDE			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Ortho Vitros 3600/5600	U/L	441	3.2%	<5	<100
Ortho \ dry chemistry					
Roche Hitachi/Hitachi Systems	U/L	476	5.6%	21	751
MedicalSystem \ Lactate to pyruvate (L-P)					
Beckman AU series	U/L	441	1.1%	16	451
MAKER \ Lactate to pyruvate (L-P)					
Abbott ARCHITECT ci series	U/L	442	3%	11	313
BIOSINO \ Lactate to pyruvate (L-P)					
Siemens Dimension series	U/L	449	1.9%	<5	<100
Siemens \ OTHER					
Beckman AU series	U/L	468	12%	<5	<100
Beckman Coulter \ Lactate to pyruvate (L-P)					
(LDL-C)低密度脂蛋白胆固醇					
Roche cobas 4000/6000/8000	mmol/L	3.29	10.2%	7	191
Dedicated \ Direct Measure					
Abbott ARCHITECT ci series	mmol/L	2.39	2.8%	27	764
BIOSINO \ selected suppression					
Roche Hitachi/Hitachi Systems	mmol/L	3.06	8.9%	18	627
MedicalSystem \ Enzymatic					
Beckman AU series	mmol/L	3.01	15.3%	14	405
MAKER \ Enzymatic					
Beckman AU series	mmol/L	2.74	12%	<5	<100
Beckman Coulter \ Enzymatic					
(Li)锂					
Beckman AU series	mmol/L	1.67	12%	<5	<100
Beckman Coulter \ Colorimetry					
(LIP)脂肪酶					
Beckman AU series	U/L	28	12%	<5	<100
Beckman Coulter \ Glucose oxidase (GOD)					
(Mg)镁					
Roche cobas 4000/6000/8000	mmol/L	1.46	3.1%	7	185
Dedicated \ Calmagite colorimetry					
Ortho Vitros 3600/5600	mmol/L	1.5	14%	<5	<100
Ortho \ dry chemistry					
Roche Hitachi/Hitachi Systems	mmol/L	1.5	2.4%	19	631
Wako \ Xylidyl Blue					
Beckman AU series	mmol/L	1.44	1.2%	16	448
MAKER \ Xylidyl Blue					
Abbott ARCHITECT ci series	mmol/L	1.54	5.4%	11	311
BIOSINO \ Calmagite colorimetry					
Beckman AU series	mmol/L	1.16	12%	<5	<100
Beckman Coulter \ Xylidyl Blue					
(Na)钠					
Roche cobas c 501/502/701/702	mmol/L	146	1.2%	9	255
Roche \ ISE (direct)					
Roche cobas 4000/6000/8000	mmol/L	148	1.6%	7	193
Dedicated \ ISE (direct)					
Abbott ARCHITECT ci series	mmol/L	149	1.4%	27	766
BIOSINO \ ISE (indirect)					
Roche Hitachi/Hitachi Systems	mmol/L	149	1.2%	21	753
Roche \ ISE (indirect)					
Beckman AU series	mmol/L	153	1.7%	16	454
Beckman Coulter \ ISE (direct)					
Beckman AU series	mmol/L	153	12%	<5	<100
Beckman Coulter \ ISE (indirect)					
Siemens Dimension series	mmol/L	150	1.4%	<5	<100
Siemens \ ISE (indirect)					
Ortho Vitros 3600/5600	mmol/L	154	9.8%	<5	<100
Ortho \ dry chemistry					
(P)磷					
Roche cobas c 501/502/701/702	mmol/L	2.91	1.9%	8	244
Roche \ Phosphomolybdate-UV-end point					
Roche cobas 4000/6000/8000	mmol/L	2.91	3.7%	7	183
Dedicated \ Phosphomolybdate-UV-end point					
Ortho Vitros 3600/5600	mmol/L	2.91	16.6%	<5	<100
Ortho \ dry chemistry					
Abbott ARCHITECT ci series	mmol/L	2.78	1.7%	27	763
BIOSINO \ Phosphomolybdate UV-rate					
Roche Hitachi/Hitachi Systems	mmol/L	2.75	3%	19	689
Wako \ Phosphomolybdate-UV-end point					
Beckman AU series	mmol/L	2.78	2.4%	16	457
MAKER \ Phosphomolybdate-UV-end point					
Beckman AU series	mmol/L	2.71	12%	<5	<100
Beckman Coulter \ Phosphomolybdate UV-rate					
Siemens Dimension series	mmol/L	2.76	2%	<5	<100
Siemens \ OTHER					
(PRO/ TP)(总) 蛋白					
Roche cobas c 501/502/701/702	g/L	69.7	1.5%	8	244
Roche \ Biuret-endpoint					
Roche cobas 4000/6000/8000	g/L	71.6	2.4%	7	197
Dedicated \ Biuret-rate					
Abbott ARCHITECT ci series	g/L	70.6	2.6%	27	688
BIOSINO \ Biuret-rate					
Roche Hitachi/Hitachi Systems	g/L	71.3	1.4%	21	761
MedicalSystem \ Biuret-endpoint					
Beckman AU series	g/L	72.4	2.1%	16	452
MAKER \ Biuret-endpoint					

REF	581-2	6×5mL	EXP	2025-05-26	LOT	376DFDE	更新日期：	2024-04
6 / 6								

		批号：376DFDE			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	68.3	12%	<5	<100
Siemens Dimension series Siemens \ Biuret-rate	g/L	72	0.9%	<5	<100
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	60	8.6%	7	303
Roche cobas 4000/6000/8000 Dedicated \ Enzymatic Cycling Assays Method	μmol/L	53.1	6.2%	7	190
Roche cobas 4000/6000/8000 WeiRiKang \ Enzymatic Cycling Assays Method	μmol/L	64.7	2.7%	<5	<100
Beckman AU series MAKER \ Enzymatic Cycling Assays Method	μmol/L	58.3	6.1%	16	451
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	57.6	8.1%	11	309
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	69.3	12%	<5	<100
(TBIL)总胆红素					
Roche cobas c 501/502/701/702 Roche \ Diazotization -Blank Compensation	μmol/L	95.6	4.4%	8	242
Roche cobas 4000/6000/8000 Dedicated \ Diazotization -Blank Compensation	μmol/L	101	6.9%	7	191
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	136	4.8%	27	763
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	120	4.1%	21	767
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	79.4	1.5%	<5	<100
Beckman AU series MAKER \ Diazotization	μmol/L	132	4.7%	16	455
Siemens Dimension series Siemens \ Vanadate	μmol/L	137	1.3%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	134	12%	<5	<100
(TG)甘油三酯					
Roche cobas c 501/502/701/702 Roche \ Free glycerol removed	mmol/L	2.38	2.4%	9	259
Roche cobas 4000/6000/8000 Dedicated \ GPO-POD	mmol/L	2.32	4.7%	7	187
Abbott ARCHITECT ci series BIOSINO \ GK-PK-LDH	mmol/L	1.87	3.6%	27	794
Roche Hitachi/Hitachi Systems MedicalSystem \ GPO-POD	mmol/L	2.07	2.4%	21	767
Beckman AU series MAKER \ GK-PK-LDH	mmol/L	1.99	2.5%	16	454
Beckman UniCel Dxl 600/800 Beckman Coulter \ GPO-POD	mmol/L	1.77	12%	<5	<100
(TIBC)总铁结合力					
Beckman AU series LEADMAN \ Ferrozine colorimetry	μmol/L	61.3	3.2%	16	452
(UA)尿酸					
Roche cobas c 501/502/701/702 Roche \ Uricase colorimetry	μmol/L	449	1.8%	8	241
Roche cobas 4000/6000/8000 Dedicated \ Uricase colorimetry	μmol/L	474	2.2%	7	190
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	456	14.3%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Phosphotungstate colorimetry	μmol/L	483	2.3%	27	794
Roche Hitachi/Hitachi Systems MedicalSystem \ Uricase colorimetry	μmol/L	510	2%	21	767
Beckman AU series MAKER \ Uricase colorimetry	μmol/L	483	1.3%	16	443
Siemens Dimension series Siemens \ OTHER	μmol/L	502	1.5%	<5	<100
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	494	12%	<5	<100
(UIBC)不饱和铁结合力					
Beckman AU series LEADMAN \ Ferrozine colorimetry	μmol/L	31	4.2%	16	456
(Urea)尿素					
Abbott ARCHITECT ci series BIOSINO \ Diacetyl monoxime	mmol/L	18	4.4%	27	557
Beckman AU series MAKER \ OTHER	mmol/L	18.6	2.5%	16	453
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	17.7	12%	<5	<100

注：此报告所提供的数据均基于检测同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。