

REF	760-2 6x3mL	EXP 2026-10-29	LOT 48801C3	更新日期:	2025-02 1 / 3
------------	----------------	-----------------------	--------------------	-------	------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(K)钾	(Na)钠	(Ca)钙
(P)磷	(Fe)铁	(Mg)镁
(Li)锂	(Cu)铜	(Zn)锌
(Al)铝	(Sb)锑	(As)砷
(Be)铍	(Bi)铋	(Cd)镉
(Cr)铬	(Co)钴	(I)碘
(Pb)铅	(Mn)锰	(Hg)汞
(Mo)钼	(Ni)镍	(Se)硒
(Ti)钛	(Sn)锡	(V)钒
(Au)金	(B)硼	(Rb)铷
(Ag)银	(Sr)锶	(Ba)钡
(Br)溴	(Cs)铯	(Si)硅
(S)硫		

REF	760-2	6x3mL	EXP	2026-10-29	LOT	48801C3	更新日期: 2025-02	2 / 3
------------	-------	-------	------------	------------	------------	---------	---------------	-------

批号: 48801C3			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(Ag)银			
OTHER Dedicated \ ICP-MS	mg/L	0.013	0.01-0.016
(Al)铝			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	μmol/L	0.964	0.846-1.082
OTHER Dedicated \ ICP-MS	mg/L	0.062	0.05-0.074
(As)砷			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	μmol/L	1.06	0.92-1.2
OTHER Dedicated \ ICP-MS	mg/L	0.012	0.01-0.014
(Au)金			
OTHER Dedicated \ ICP-MS	μg/L	0.05	0.039-0.061
(B)硼			
OTHER Dedicated \ ICP-MS	mg/L	0.032	0.026-0.038
(Ba)钡			
OTHER Dedicated \ ICP-MS	mg/L	0.173	0.138-0.208
(Be)铍			
OTHER Dedicated \ ICP-MS	mg/L	0.001	0.001-0.001
(Bi)铋			
OTHER Dedicated \ ICP-MS	μg/L	5.72	4.58-6.86
(Br)溴			
OTHER Dedicated \ ICP-MS	mg/L	0.972	0.778-1.166
(Ca)钙			
OTHER OTHER \ OTHER	mmol/L	1.07	0.89-1.25
OTHER Dedicated \ ICP-MS	mmol/L	1.38	1.1-1.66
(Cd)镉			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	μmol/L	0.057	0.052-0.062
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	μg/L	6.38	5.84-6.92
OTHER Dedicated \ ICP-MS	μmol/L	0.054	0.043-0.065
OTHER Dedicated \ ICP-MS	μg/L	6.05	4.84-7.26
(Co)钴			
OTHER Dedicated \ ICP-MS	μmol/L	0.009	0.007-0.011
OTHER Dedicated \ ICP-MS	μg/L	0.54	0.432-0.648
(Cr)铬			
OTHER Dedicated \ ICP-MS	μg/L	10.8	8.6-13
(Cs)铯			
OTHER Dedicated \ ICP-MS	mg/L	0.003	0.002-0.004
(Cu)铜			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	μmol/L	14.8	11.4-18.2
OTHER Dedicated \ ICP-MS	μmol/L	16.5	13.2-19.8
(Fe)铁			
OTHER Dedicated \ ICP-MS	μmol/L	6186	4949-7423
(Hg)汞			
OTHER Dedicated \ ICP-MS	μmol/L	0.113	0.09-0.136
(I)碘			
OTHER Dedicated \ ICP-MS	μg/L	75.6	60.5-90.7
(K)钾			
OTHER Dedicated \ ICP-MS	mmol/L	30.5	24.4-36.6

REF	760-2	6x3mL	EXP	2026-10-29	LOT	48801C3	更新日期： 2025-02 3 / 3
------------	-------	-------	------------	------------	------------	---------	------------------------

(Li) 锂			
OTHER Dedicated \ ICP-MS	mmol/L	0.002	0.002-0.002
(Mg) 镁			
OTHER OTHER \ OTHER	mmol/L	1.11	0.91-1.31
OTHER Dedicated \ ICP-MS	mmol/L	1.82	1.46-2.18
(Mn) 锰			
OTHER Dedicated \ ICP-MS	µg/L	25.9	20.7-31.1
(Mo) 钼			
OTHER Dedicated \ ICP-MS	µg/L	5.4	4.32-6.48
(Na) 钠			
OTHER Dedicated \ ICP-MS	mmol/L	65.8	52.6-79
(Ni) 镍			
OTHER Dedicated \ ICP-MS	µg/L	9.72	7.78-11.66
(P) 磷			
OTHER Dedicated \ ICP-MS	mmol/L	6.28	5.02-7.54
(Pb) 铅			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	µmol/L	1.47	1.28-1.66
OTHER Dedicated \ ICP-MS	µmol/L	1.72	1.38-2.06
(Rb) 铷			
OTHER Dedicated \ ICP-MS	mg/L	1.62	1.3-1.94
(S) 硫			
OTHER Dedicated \ ICP-MS	mg/L	1080	864-1296
(Sb) 锑			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	µmol/L	6.25	5.94-6.56
OTHER Dedicated \ ICP-MS	mg/L	0.026	0.021-0.031
(Se) 硒			
OTHER Dedicated \ ICP-MS	µmol/L	0.192	0.154-0.23
(Si) 硅			
OTHER Dedicated \ ICP-MS	mg/L	0.648	0.518-0.778
(Sn) 锡			
OTHER Dedicated \ ICP-MS	µg/L	4.32	3.46-5.18
(Sr) 锶			
OTHER Dedicated \ ICP-MS	mg/L	0.081	0.065-0.097
(Ti) 钛			
OTHER Dedicated \ ICP-MS	µg/L	3.24	2.59-3.89
(V) 钒			
OTHER Dedicated \ ICP-MS	µg/L	3.24	2.59-3.89
(Zn) 锌			
OTHER EWAI \ Atomic absorption spectroscopy (AAS)	µmol/L	97.4	87.7-107.1
OTHER Dedicated \ ICP-MS	µmol/L	102	82-122

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。