报告类别： **汇总实验室数据报告**

质控品名： **多专业免疫分析非定值质控品**

REF	539-2	6×3mL	EXP	2023-04-12	LOT	336C8D2	更新日期：	2024-03

		批号：336C8D2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	132	5%	20	570
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	119	6.1%	10	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	112	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	126	9.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	127	10.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	117	10.3%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	68	6.3%	20	524
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	22.7	4.5%	13	378
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	30.2	10.3%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	66.2	10.4%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	6.15	12%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	35.2	14.5%	20	515
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.9	3.9%	13	379
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.8	12%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.1	11%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21.7	11.1%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	29.3	9.9%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.3	4.7%	20	576
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	78.6	3.8%	10	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	59.6	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	54.3	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	61.2	10.3%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	80.2	11.5%	<5	<100
(CA 15-3)糖类抗原 15-3					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	54.3	4.7%	20	570
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.8	3.6%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.3	10.3%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.6	9.8%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.2	11.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	24.9	10.4%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	51.2	3.5%	20	571
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	40.8	5.6%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	51.7	10.3%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	92.4	10.4%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.7	10.1%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **多专业免疫分析非定值质控品**

REF	539-2	6×3mL	EXP	2023-04-12	LOT	336C8D2	更新日期：	2024-03

		批号：336C8D2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CEA)癌胚抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	2.6%	20	574
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.2	6.5%	10	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.2	11.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.88	12.1%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.92	9.9%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	11.7%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Roche cobas c 501/502/701/702 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.86	12%	<5	<100
(Cortisol)皮质醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	206	13.9%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	246	10.2%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	247	10.1%	<5	<100
(C-P)C-肽					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.46	5.2%	20	181
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.96	3.6%	12	353
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.62	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.55	9.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.55	10.2%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6338	10.5%	<5	<100
(E2)雌二醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	986	3.6%	20	176
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	825	3.4%	13	373
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	947	10.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	929	10.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1098	9.8%	<5	<100
(FA)叶酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.25	10.4%	20	179
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.26	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.17	10.3%	<5	<100
(Ferritin)铁蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	226	11.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	201	10.4%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	143	9.8%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.96	3%	20	180
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.74	6.2%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.04	10.3%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.96	9.9%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **多专业免疫分析非定值质控品**

REF	539-2	6×3mL	EXP	2023-04-12	LOT	336C8D2	更新日期：	2024-03

		批号：336C8D2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.82	10%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.21	11.9%	<5	<100
(FSH)促卵泡刺激素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	30	4.1%	20	182
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	32.8	3.7%	13	378
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	36.6	6.8%	10	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	32.6	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	34.2	10.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	33.9	12.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	32.7	10.1%	<5	<100
(FT3)游离T3					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	14.5	3.3%	20	568
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.6	11.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.1	3.8%	13	380
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	11.7	8.9%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	13.5	9.6%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	10.4	11.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.5	10.4%	<5	<100
(FT4)游离T4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	33.7	4%	20	566
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	29.8	11.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	23.3	3.2%	13	376
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	33.5	5.5%	10	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	30.4	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	33.2	10.8%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	29.6	11.5%	<5	<100
(Fβ-hCG)游离β-绒毛膜促性腺激素					
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	108	11.1%	<5	<100
(HCY)同型半胱氨酸					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μmol/L	24.2	10.3%	<5	<100
(HGH)人生长激素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	4.38	12.1%	<5	<100
(INS)胰岛素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	32.5	5%	20	179
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	23.6	2.2%	13	381
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	35.2	10.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	35.8	10.1%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	34.5	10.4%	<5	<100
(LH)促黄体生成素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	22.3	5.2%	20	183
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	16.3	3.4%	13	376

报告类别： **汇总实验室数据报告**

质控品名： **多专业免疫分析非定值质控品**

REF	539-2	6×3mL	EXP	2023-04-12	LOT	336C8D2	更新日期：	2024-03

		批号：336C8D2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	22.3	6.3%	10	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	22.8	11.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	20.4	10.3%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	19.8	10.6%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	23.1	10.4%	<5	<100
(PRL)催/ 泌乳素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.4	4.6%	20	181
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.8	2.2%	13	372
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.5	7.7%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.9	10.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.9	11.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.9	9.5%	<5	<100
(PROG)孕酮					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.6	3.3%	20	435
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.13	3.2%	13	376
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.4	5.2%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.7	10.3%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.5	11.9%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.1	10.2%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	11%	<5	<100
(PSA)前列腺特异性抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.9	3.4%	20	170
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.9	4.3%	10	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.5	11.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.5	11.9%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.6	11.1%	<5	<100
(PTH)甲状旁腺素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	8.6	5.9%	20	181
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	8.25	10.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	8.69	10%	<5	<100
(TES)睾酮					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.48	6%	20	183
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.407	7.1%	13	375
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.56	9.8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.43	11.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.531	10.3%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.2	3.8%	19	197
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.26	10.7%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.86	9.9%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **多专业免疫分析非定值质控品**

REF	539-2	6×3mL	EXP	2023-04-12	LOT	336C8D2	更新日期：	2024-03

		批号：336C8D2			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TSH)促甲状腺素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.1	3.4%	20	574
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	11.8	13.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	9.09	2.9%	13	377
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	12.1	5.1%	10	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	10.1	11.9%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	9.68	10.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.2	11.6%	<5	<100
(TT3)总三碘甲状原氨酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.64	4%	20	212
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.18	14.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.38	3.9%	13	377
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.38	10.1%	10	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.67	12%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.38	10.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.19	9.8%	<5	<100
(TT4)总T4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	125	3.9%	20	213
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	132	13.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	119	3%	13	377
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	138	7.8%	10	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	103	9.8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	111	10.5%	<5	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	110	9.2%	<5	<100
(VB12)维生素B12					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	659	5.2%	20	180
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	711	10.7%	<5	<100
(β2M)β-2-微球蛋白					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1340	10.4%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.112	5.7%	10	<100
Abbott ARCHITECT i series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.128	9.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.135	10.4%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。