



REF	780-C	6×1mL	EXP	2023-10-03	LOT	4A6ACED	更新日期：	2023-02
								1 / 2

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(HBsAg)乙肝表面抗原		



REF	780-C	6×1mL	EXP	2023-10-03	LOT	4A6ACED	更新日期：	2023-02
2 / 2								

		批号：4A6ACED	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(HBsAg)乙肝表面抗原			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	Sample/Cu toff	20.3	15.4-25.2
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	Sample/Cu toff	14.2	10.8-17.6
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	Sample/Cu toff	13.6	10.3-16.9
Manual KHB \ Chemiluminescence Immunoassay (CLIA)	Sample/Cu toff	4.65	3.53-5.77
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	(Index)	18.7	14.2-23.2