

REF 510-1	6x3mL	EXP 2025-09-26	LOT 30A61AA	更新日期： 2020-12 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A61AA				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
(AFP)甲胎蛋白				
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.76	2.22-3.3	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.68	2.97-4.39	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.23	3.41-5.05	
(CA 125)糖类抗原 125				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.6	26-39.2	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	28.3	22.7-33.9	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.6	26.3-38.9	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.4	23.5-35.3	
(CA 15-3)糖类抗原 15-3				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.3	12.4-18.2	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	14.9	11.9-17.9	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.8	11.9-17.7	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.9	12-17.8	
(CA 19-9)糖类抗原 19-9				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	45.3	36.5-54.1	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	26.2	20.9-31.5	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	27.6	22.6-32.6	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.2	11.5-16.9	
(CA242)糖类抗原 242				
Manual GLD \ Enzyme linked immunosorbent assays (ELISA)	U/mL	13.7	11.1-16.3	
(CEA)癌胚抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.04	2.42-3.66	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.54	1.24-1.84	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.73	2.17-3.29	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.25	1.83-2.67	
(Ferritin)铁蛋白				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	µg/L	62.8	50.3-75.3	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	µg/L	53.7	43.3-64.1	
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	µg/L	72.6	58.8-86.4	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	µg/L	49.7	40.2-59.2	
(F-PSA)游离前列腺特异性抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.95	0.784-1.12	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.94	0.734-1.15	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.11	0.912-1.31	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.58	1.27-1.89	
(hCG)人绒毛膜促性腺激素				
Siemens ADVIA Centaur XP Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	10.6	8.44-12.8	

REF 510-1	6x3mL	EXP 2025-09-26	LOT 30A61AA	更新日期： 2020-12 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A61AA				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
(PSA)前列腺特异性抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.37	1.91-2.83	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.26	1.81-2.71	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.24	1.76-2.72	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.28	1.73-2.83	
(SCC)鳞状细胞癌抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.69	1.37-2.01	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.48	1.24-1.72	
(TPA)组织多肽抗原				
DiaSorin Liaison /XL DiaSorin \ Chemiluminescence Immunoassay (CLIA)	U/L	392	314-470	
(TPS)组织多肽特异抗原				
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	16.4	13.2-19.6	
(β-hCG)人绒毛膜促性腺激素β亚基				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	13.6	11.1-16.1	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。