

REF	581-1	6×5mL	EXP	2025-05-26	LOT	376B8CE	更新日期：	2022-06
1 / 2								

批号：376B8CE			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(a-HBDH)α-羟丁酸脱氢酶			
Beckman AU series Beckman Coulter \ DGKC	U/L	145	110-180
(Alb)白蛋白			
Beckman AU series Beckman Coulter \ Bromocresol green	μmol/L	411	312-510
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	27.1	20.6-33.6
(ALP)碱性磷酸酶			
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	45.1	34.3-55.9
(ALT/ SGPT)丙氨酸氨基转移酶			
Beckman AU series Beckman Coulter \ UV with P5P	U/L	39.6	30.1-49.1
(AMY)淀粉酶			
Beckman AU series Beckman Coulter \ Iodine colorimetry	U/L	62.5	47.5-77.5
(AST/ SGOT)天门冬氨酸氨基转移酶			
Beckman AU series Beckman Coulter \ UV without P5P	U/L	38.2	29-47.4
(Ca)钙			
Beckman AU series Beckman Coulter \ BAPTA	mmol/L	1.9	1.44-2.36
(CHE)胆碱酯酶			
Beckman AU series Beckman Coulter \ Butyrylthiocholine Method	U/L	5410	4110-6700
(CHOL/ TC)总胆固醇			
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	2.21	1.68-2.74
(CK)肌酸激酶			
Beckman AU series Beckman Coulter \ Urease-UV-rate	U/L	149	113-185
(Cl)氯			
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	76.8	58.4-95.2
(CO2 (HCO3-))二氧化碳 (碳酸氢根)			
Beckman AU series Beckman Coulter \ PEPC	mmol/L	10.6	8.06-13.1
(Cr)肌酐			
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	74.1	56.3-91.9
(Cu)铜			
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	10.9	8.28-13.5
(DBIL)直接胆红素			
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	4.47	3.4-5.54
(Fe)铁			
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	10.9	8.28-13.5
(GGT)L-γ-谷氨酰基转移酶			
Beckman AU series Beckman Coulter \ γ-glutamyl -4-nitroaniline	U/L	31.1	23.6-38.6
(GLU)葡萄糖			
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	4.18	3.18-5.18
(HDL-C)高密度脂蛋白胆固醇			
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	0.54	0.41-0.67
(K)钾			
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	2.01	1.53-2.49
(LA)乳酸			
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	0.98	0.745-1.22
(LDH)乳酸脱氢酶			
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	146	111-181



REF	581-1	6×5mL	EXP	2025-05-26	LOT	376B8CE	更新日期：	2022-06
2 / 2								

批号：376B8CE			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(LDL-C)低密度脂蛋白胆固醇			
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.17	0.889-1.45
(Li)锂			
Beckman AU series Beckman Coulter \ Colorimetry	mmol/L	0.44	0.334-0.546
(LIP)脂肪酶			
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	23.4	17.8-29
(Mg)镁			
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.58	0.441-0.719
(Na)钠			
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	109	82.8-135
(P)磷			
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	0.68	0.517-0.843
(PRO/ TP)(总) 蛋白			
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	40.9	31.1-50.7
(TBA)总胆汁酸			
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	10.9	8.28-13.5
(TBIL)总胆红素			
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	14.2	10.8-17.6
(TG)甘油三酯			
Beckman UniCel Dxl 600/800 Beckman Coulter \ GPO-POD	mmol/L	0.7	0.532-0.868
(UA)尿酸			
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	169	128-210
(Urea)尿素			
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	4.36	3.31-5.41

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。