

|                  |       |                       |                    |               |
|------------------|-------|-----------------------|--------------------|---------------|
| <b>REF</b> 510-2 | 6x3mL | <b>EXP</b> 2025-09-26 | <b>LOT</b> 30A88BA | 更新日期： 2024-06 |
|                  |       |                       |                    | 1 / 3         |

本报告含以下项目的示范实验室数据

| 分析项目(Analyte)         | 分析项目(Analyte)      | 分析项目(Analyte)       |
|-----------------------|--------------------|---------------------|
| (AFP)甲胎蛋白             | (CA 125)糖类抗原 125   | (CA 15-3)糖类抗原 15-3  |
| (CA 19-9)糖类抗原 19-9    | (CA 72-4)糖类抗原 72-4 | (CEA)癌胚抗原           |
| (Cyfra 21-1)细胞角蛋白19片段 | (Ferritin)铁蛋白      | (F-PSA)游离前列腺特异性抗原   |
| (hCG)人绒毛膜促性腺激素        | (PSA)前列腺特异性抗原      | (SCC)鳞状细胞癌抗原        |
| (TPA)组织多肽抗原           | (TPS)组织多肽特异抗原      | (β-hCG)人绒毛膜促性腺激素β亚基 |

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

|                  |       |                       |                    |               |
|------------------|-------|-----------------------|--------------------|---------------|
| <b>REF</b> 510-2 | 6x3mL | <b>EXP</b> 2025-09-26 | <b>LOT</b> 30A88BA | 更新日期： 2024-06 |
|                  |       |                       |                    | 2 / 3         |

| 批号： 30A88BA  |       |      |            |  |
|--|-------|------|------------|--|
| 项目\仪器\试剂方法   | 单位    | 均值   | + / - 2 SD |  |
| <b>(AFP)甲胎蛋白</b>   |       |      |            |  |
| Beckman UniCel Dxl 600/800<br>Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )  | ng/mL | 83.2 | 67.1-99.3  |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | ng/mL | 70.5 | 56.1-84.9  |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | ng/mL | 60.3 | 48.1-72.5  |  |
| <b>(CA 125)糖类抗原 125</b>  |       |      |            |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 229  | 184-274    |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | IU/mL | 179  | 138-220    |  |
| DiaSorin Liaison /XL<br>Diasorin \ Chemiluminescence Immunoassay (CLIA )               | IU/mL | 276  | 221-331    |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | IU/mL | 288  | 232-344    |  |
| <b>(CA 15-3)糖类抗原 15-3</b>  |       |      |            |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | IU/mL | 104  | 80.4-128   |  |
| DiaSorin Liaison /XL<br>Diasorin \ Chemiluminescence Immunoassay (CLIA )               | IU/mL | 95.3 | 76.4-114   |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | IU/mL | 69.3 | 54.5-84.1  |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 96.3 | 77-116     |  |
| maglumi system<br>Snibe \ Chemiluminescence Immunoassay (CLIA )                        | IU/mL | 108  | 86.4-130   |  |
| <b>(CA 19-9)糖类抗原 19-9</b>  |       |      |            |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | IU/mL | 178  | 142-214    |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | IU/mL | 76.8 | 60.5-93.1  |  |
| DiaSorin Liaison /XL<br>Diasorin \ Chemiluminescence Immunoassay (CLIA )               | IU/mL | 187  | 152-222    |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | IU/mL | 398  | 318-478    |  |
| <b>(CA 72-4)糖类抗原 72-4</b>  |       |      |            |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | IU/mL | 8.23 | 6.21-10.3  |  |
| <b>(CEA)癌胚抗原</b>   |       |      |            |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | ng/mL | 46.3 | 37.3-55.3  |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | ng/mL | 35.7 | 27.9-43.5  |  |
| Beckman UniCel Dxl 600/800<br>Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )  | ng/mL | 31.3 | 25.3-37.3  |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | ng/mL | 26.5 | 21.7-31.3  |  |
| <b>(Cyfra 21-1)细胞角蛋白19片段</b>   |       |      |            |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | ng/mL | 10.3 | 8.08-12.5  |  |
| <b>(Ferritin)铁蛋白</b>   |       |      |            |  |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)          | µg/L  | 352  | 278-426    |  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme immunoassay (CLEIA) | µg/L  | 306  | 250-362    |  |
| Beckman Access<br>Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )              | µg/L  | 268  | 219-317    |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | µg/L  | 409  | 329-489    |  |
| <b>(F-PSA)游离前列腺特异性抗原</b>   |       |      |            |  |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )            | ng/mL | 4.23 | 3.36-5.1   |  |

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

|                  |       |                       |                    |               |
|------------------|-------|-----------------------|--------------------|---------------|
| <b>REF</b> 510-2 | 6x3mL | <b>EXP</b> 2025-09-26 | <b>LOT</b> 30A88BA | 更新日期： 2024-06 |
|                  |       |                       |                    | 3 / 3         |

|   |       |      |           |
|---|-------|------|-----------|
| Beckman Access<br>Beckman Coulter \ Chemiluminescence<br>Immunoassay (CLIA )              | ng/mL | 6.03 | 4.66-7.4  |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme<br>immunoassay (CLEIA) | ng/mL | 5.12 | 4.16-6.08 |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay<br>(ECLIA)          | ng/mL | 4.86 | 3.89-5.83 |
| <b>(hCG)人绒毛膜促性腺激素</b>   |       |      |           |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Enzyme linked immunosorbent assays<br>(ELISA)   | IU/L  | 98.6 | 78.8-118  |
| <b>(PSA)前列腺特异性抗原</b>  |       |      |           |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Chemiluminescence enzyme<br>immunoassay (CLEIA) | ng/mL | 16.2 | 13-19.4   |
| Beckman Access<br>Beckman Coulter \ Chemiluminescence<br>Immunoassay (CLIA )              | ng/mL | 19.3 | 15.4-23.2 |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )               | ng/mL | 17.9 | 14.6-21.2 |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay<br>(ECLIA)          | ng/mL | 18.3 | 14.7-21.9 |
| <b>(SCC)鳞状细胞癌抗原</b>   |       |      |           |
| Abbott ARCHITECT i series<br>Abbott \ Chemiluminescence Immunoassay (CLIA )               | ng/mL | 12.8 | 10.2-15.4 |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay<br>(ECLIA)          | ng/mL | 8.85 | 6.98-10.7 |
| <b>(TPA)组织多肽抗原</b>  |       |      |           |
| DiaSorin Liaison /XL<br>Diasorin \ Chemiluminescence Immunoassay (CLIA )                  | U/L   | 3820 | 3040-4590 |
| <b>(TPS)组织多肽特异抗原</b>  |       |      |           |
| Manual<br>OTHER \ Enzyme linked immunosorbent assays<br>(ELISA)                           | U/L   | 104  | 86-122    |
| <b>(β-hCG)人绒毛膜促性腺激素β亚基</b>  |       |      |           |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay<br>(ECLIA)          | IU/L  | 157  | 129-185   |

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。