

REF 539-1	6x3mL	EXP 2025-05-03	LOT 336A28B	更新日期： 2022-10 1 / 3
------------------	-------	-----------------------	--------------------	------------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(TT3)总三碘甲状腺原氨酸	(FT3)游离T3	(TT4)总T4
(FT4)游离T4	(TSH)促甲状腺素	(Cortisol)皮质醇
(FSH)促卵泡刺激素	(LH)促黄体生成素	(PROG)孕酮
(PRL)催/ 泌乳素	(TES)睾酮	(E2)雌二醇
(C-P)C-肽	(FA)叶酸	(INS)胰岛素
(VB12)维生素B12	(CEA)癌胚抗原	(hCG)人绒毛膜促性腺激素
(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3	(CA 19-9)糖类抗原 19-9
(HCY)同型半胱氨酸	(Anti-Tg)抗甲状腺球蛋白抗体	(Anti-TPO)抗甲状腺过氧化物酶抗体
(DHEA-S)硫酸脱氢表雄酮	(SHBG)性激素结合球蛋白	

REF 539-1	6x3mL	EXP 2025-05-03	LOT 336A28B	更新日期： 2022-10 2 / 3
------------------	-------	-----------------------	--------------------	------------------------

批号：336A28B			
项目\仪器\试剂方法	单位	均值	+ / - 2 SD
(Anti-Tg)抗甲状腺球蛋白抗体			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	24.4	18.6-30.2
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	55.1	41.9-68.3
(Anti-TPO)抗甲状腺过氧化物酶抗体			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21	16-26
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	70.7	53.7-87.7
(CA 125)糖类抗原 125			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21.7	16.5-26.9
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.2	10.8-17.6
(CA 15-3)糖类抗原 15-3			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21.8	16.6-27
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.3	17.7-28.9
(CA 19-9)糖类抗原 19-9			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	117	88.9-145
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	21.6	16.4-26.8
(CEA)癌胚抗原			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.97	3.02-4.92
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.02	2.3-3.74
(Cortisol)皮质醇			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31	23.6-38.4
(C-P)C-肽			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.923	0.701-1.14
(DHEA-S)硫酸脱氢表雄酮			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	287	218-356
(E2)雌二醇			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	256	195-317
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	407	309-505
(FA)叶酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.2	0.152-0.248
(FSH)促卵泡刺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.98	3.02-4.94
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.78	4.39-7.17
(FT3)游离T3			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	4.11	3.12-5.1
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3.82	2.9-4.74
(FT4)游离T4			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	7.76	5.91-9.61
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	9.73	7.39-12.1
(hCG)人绒毛膜促性腺激素			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.53	3.44-5.62
(HCY)同型半胱氨酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	9.84	7.48-12.2

REF 539-1	6x3mL	EXP 2025-05-03	LOT 336A28B	更新日期： 2022-10 3 / 3
------------------	-------	-----------------------	--------------------	------------------------

(INS)胰岛素				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	15		11.4-18.6
(LH)促黄体生成素				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	1.12		0.85-1.39
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.83		2.91-4.75
(PRL)催/泌乳素				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.77		2.1-3.44
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.53		2.68-4.38
(PROG)孕酮				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.6		1.22-1.98
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.21		0.919-1.5
(SHBG)性激素结合球蛋白				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	17.3		13.1-21.5
(TES)睾酮				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.661		0.503-0.819
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.773		0.587-0.959
(TSH)促甲状腺素				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	0.029		0.023-0.035
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	0.021		0.015-0.027
(TT3)总三碘甲状腺原氨酸				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.62		0.471-0.769
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.487		0.37-0.604
(TT4)总T4				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	24		18.2-29.8
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	nmol/L	36.9		28-45.8
(VB12)维生素B12				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	161		122-200

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。