

<b>REF</b> 600-2	30x1.7mL	<b>EXP</b> 2026-07-04	<b>LOT</b> 393DDBF	更新日期： 2022-10
				1 / 1

批号：393DDBF				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(Ca)钙</b>				
Radiometer ABL Series Dedicated \ ISE (indirect)	mmol/L	1.1	0.836-1.36	
Roche cobas b 121/123/221 Dedicated \ ISE (indirect)	mmol/L	1.13	0.859-1.4	
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	1.15	0.874-1.43	
<b>(Cl)氯</b>				
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	101	76.9-125	
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	95.2	72.3-118	
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	93.2	70.8-116	
<b>(K)钾</b>				
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	4.07	3.09-5.05	
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	4.15	3.15-5.15	
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	4.13	3.14-5.12	
<b>(Na)钠</b>				
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	142	108-176	
Radiometer AQT Series Dedicated \ ISE (direct)	mmol/L	140	106-174	
Siemens RAPIDPoint 400/500 Siemens \ ISE (indirect)	mmol/L	143	109-177	
IL Gem Premier 3000 series Dedicated \ ICP-MS	mmol/L	146	111-181	
<b>(PCO2)二氧化碳分压</b>				
Siemens RAPIDPoint 400/500 Siemens \ Glucose oxidase (oxygen electrode)	mmHg	32.1	24.4-39.8	
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	32.9	25-40.8	
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	36.3	27.6-45	
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	40	30.4-49.6	
Radiometer AQT Series Dedicated \ Electrode	mmHg	43	32.7-53.3	
Radiometer ABL Series Dedicated \ Electrode	mmHg	41	31.2-50.8	
<b>(pH)酸碱度</b>				
Roche cobas b 121/123/221 Dedicated \ Electrode	No Units	7.5	5.7-9.3	
Siemens RAPIDPoint 400/500 Siemens \ Electrode	No Units	7.49	5.69-9.29	
IL Gem Premier 3000 series Dedicated \ Electrode	No Units	7.53	5.72-9.34	
Radiometer AQT Series Dedicated \ Electrode	No Units	7.45	5.66-9.24	
Radiometer ABL Series Dedicated \ Electrode	No Units	7.56	5.74-9.38	
<b>(PO2)氧分压</b>				
Radiometer AQT Series Dedicated \ Electrode	mmHg	95.5	72.6-118	
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	97.2	73.9-121	
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	mmHg	123	93.5-153	
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	109	82.8-135	
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	124	94.3-154	
Radiometer ABL Series Dedicated \ Electrode	mmHg	120	91.2-149	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。