

<b>REF</b> 510-2	6x3mL	<b>EXP</b> 2026-01-08	<b>LOT</b> 30A8915	更新日期： 2022-03 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A8915				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(AFP)甲胎蛋白</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	80.7	62.5-98.9	
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	73.5	57.1-89.9	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	87.9	69.9-106	
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	69.8	54.6-85	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	75.4	60.3-90.5	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	72.8	55.4-90.2	
<b>(CA 125)糖类抗原 125</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	263	219-307	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	243	197-289	
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	112	90-134	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	174	139-209	
<b>(CA 15-3)糖类抗原 15-3</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	109	87-131	
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	108	86-130	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	96.9	78.9-115	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	88.4	70.4-106	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	59.8	47.8-71.8	
<b>(CA 19-9)糖类抗原 19-9</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	268	210-326	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	145	113-177	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	121	99-143	
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	94.2	78.2-110	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	81.8	65.4-98.2	
<b>(CA 50)糖类抗原 50</b>				
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/mL	72.7	58.2-87.2	
<b>(CA 72-4)糖类抗原 72-4</b>				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	28.9	22.7-35.1	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	13	10.4-15.6	
<b>(CA242)糖类抗原 242</b>				
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	U/mL	23.7	19-28.4	
<b>(CEA)癌胚抗原</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	35.8	28.2-43.4	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.7	17.3-26.1	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	37.6	31.2-44	
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	38.6	30.2-47	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	26.2	21-31.4	

<b>REF</b> 510-2	6x3mL	<b>EXP</b> 2026-01-08	<b>LOT</b> 30A8915	更新日期： 2022-03 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A8915				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(Cyfra 21-1)细胞角蛋白19片段</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	13.2	10.8-15.6	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	15.4	12.3-18.5	
<b>(Ferritin)铁蛋白</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	µg/L	349	291-407	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	266	212-320	
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	µg/L	313	250-376	
<b>(F-PSA)游离前列腺特异性抗原</b>				
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.26	4.04-6.48	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.35	3.48-5.22	
<b>(PSA)前列腺特异性抗原</b>				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	15.9	12.7-19.1	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	21.3	20.9-21.7	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11.8	9.6-14	
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15.4	12.2-18.6	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	17.9	14.3-21.5	
<b>(SCC)鳞状细胞癌抗原</b>				
Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.9	10.5-15.3	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.73	7.78-11.7	
<b>(TPA)组织多肽抗原</b>				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/L	6040	4670-7400	
<b>(TPS)组织多肽特异抗原</b>				
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/L	2210	1740-2670	
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>				
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/L	156	126-186	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/L	163	130-196	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。